

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN042420\
 Data File : BN010588.D
 Acq On : 24 Apr 2020 12:17
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 24 12:49:44 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 24 12:47:59 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.294	0.290	1.4	102	0.00
3	n-Nitrosodimethylamine	0.191	0.206	-7.9	131	0.00
4 S	2-Fluorophenol	0.876	0.957	-9.2	112	0.00
5 S	Phenol-d6	1.145	1.195	-4.4	107	0.00
6	bis(2-Chloroethyl)ether	0.992	1.062	-7.1	107	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
8 S	Nitrobenzene-d5	0.280	0.284	-1.4	110	0.00
9	Naphthalene	0.993	1.026	-3.3	108	0.00
10	Hexachlorobutadiene	0.239	0.244	-2.1	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.658	0.760	-15.5	121	0.00
12	2-Methylnaphthalene	0.714	0.734	-2.8	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.189	0.161	14.8	93	0.00
15 S	2-Fluorobiphenyl	1.456	1.489	-2.3	106	0.00
16	Acenaphthylene	1.728	1.647	4.7	102	0.00
17	Acenaphthene	1.110	1.132	-2.0	107	0.00
18	Fluorene	1.454	1.445	0.6	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
20	4-Bromophenyl-phenylether	0.222	0.218	1.8	104	0.00
21	Hexachlorobenzene	0.254	0.270	-6.3	108	0.00
22	Pentachlorophenol	0.100	0.080	20.0	97	0.00
23	Phenanthrene	1.080	1.106	-2.4	105	0.00
24	Anthracene	0.988	0.931	5.8	99	0.00
25 SURR	Fluoranthene-d10	1.213	1.300	-7.2	110	0.00
26	Fluoranthene	1.345	1.299	3.4	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
28	Pvrene	1.346	1.477	-9.7	96	0.00
29 S	Terphenyl-d14	0.915	1.000	-9.3	97	0.00
30	Benzo(a)anthracene	1.305	1.284	1.6	87	0.00
31	Chrysene	1.300	1.305	-0.4	89	0.00
32	Bis(2-ethylhexyl)phthalate	0.670	0.583	13.0	80	0.00
33	Indeno(1,2,3-cd)pyrene	1.149	1.286	-11.9	97	0.00
34 I	Perylene-d12	1.000	1.000	0.0	91	0.00
35	Benzo(b)fluoranthene	1.317	1.337	-1.5	92	0.00
36	Benzo(k)fluoranthene	1.320	1.345	-1.9	93	0.00
37 C	Benzo(a)pyrene	1.166	1.145	1.8	90	0.00
38	Dibenzo(a,h)anthracene	1.066	1.111	-4.2	94	0.00
39	Benzo(a,h,i)perylene	1.073	1.169	-8.9	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN042420\
Data File : BN010588.D
Acq On : 24 Apr 2020 12:17
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 24 12:49:44 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 24 12:47:59 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			