

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN042420\
 Data File : BN010597.D
 Acq On : 24 Apr 2020 20:13
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 24 20:43:36 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 24 12:47:59 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
2	1,4-Dioxane	0.294	0.296	-0.7	82	0.00
3	n-Nitrosodimethylamine	0.191	0.193	-1.0	97	0.00
4 S	2-Fluorophenol	0.876	0.931	-6.3	86	0.00
5 S	Phenol-d6	1.145	1.234	-7.8	87	0.00
6	bis(2-Chloroethyl)ether	0.992	1.041	-4.9	83	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
8 S	Nitrobenzene-d5	0.280	0.283	-1.1	90	0.00
9	Naphthalene	0.993	1.023	-3.0	89	0.00
10	Hexachlorobutadiene	0.239	0.241	-0.8	85	0.00
11 SURR	2-Methylnaphthalene-d10	0.658	0.757	-15.0	99	0.00
12	2-Methylnaphthalene	0.714	0.731	-2.4	88	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.189	0.162	14.3	79	0.00
15 S	2-Fluorobiphenyl	1.456	1.448	0.5	86	0.00
16	Acenaphthylene	1.728	1.625	6.0	84	0.00
17	Acenaphthene	1.110	1.139	-2.6	91	0.00
18	Fluorene	1.454	1.450	0.3	88	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
20	4-Bromophenyl-phenylether	0.222	0.212	4.5	87	0.00
21	Hexachlorobenzene	0.254	0.264	-3.9	91	0.00
22	Pentachlorophenol	0.100	0.084	16.0	88	0.00
23	Phenanthrene	1.080	1.088	-0.7	89	0.00
24	Anthracene	0.988	0.908	8.1	84	0.00
25 SURR	Fluoranthene-d10	1.213	1.273	-4.9	94	0.00
26	Fluoranthene	1.345	1.269	5.7	84	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	72	0.00
28	Pvrene	1.346	1.518	-12.8	82	0.00
29 S	Terphenyl-d14	0.915	1.026	-12.1	82	0.00
30	Benzo(a)anthracene	1.305	1.260	3.4	70	0.00
31	Chrysene	1.300	1.280	1.5	72	0.00
32	Bis(2-ethylhexyl)phthalate	0.670	0.620	7.5	70	0.00
33	Indeno(1,2,3-cd)pyrene	1.149	1.289	-12.2	80	0.00
34 I	Perylene-d12	1.000	1.000	0.0	74	0.00
35	Benzo(b)fluoranthene	1.317	1.294	1.7	73	0.00
36	Benzo(k)fluoranthene	1.320	1.365	-3.4	77	0.00
37 C	Benzo(a)pyrene	1.166	1.134	2.7	73	0.00
38	Dibenzo(a,h)anthracene	1.066	1.136	-6.6	79	0.00
39	Benzo(a,h,i)perylene	1.073	1.165	-8.6	80	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN042420\
Data File : BN010597.D
Acq On : 24 Apr 2020 20:13
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 24 20:43:36 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 24 12:47:59 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			