

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042624\
 Data File : BN031243.D
 Acq On : 26 Apr 2024 13:09
 Operator : MA/JU
 Sample : P2249-04
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E08M8

Quant Time: Apr 26 13:39:26 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.622	152	253689	20.000	ng/ul	-0.02
20) Naphthalene-d8	10.404	136	1164484	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.269	164	778923	20.000	ng/ul	-0.01
64) Phenanthrene-d10	17.022	188	1568883	20.000	ng/ul	-0.01
79) Chrysene-d12	21.257	240	1318980	20.000	ng/ul	-0.01
88) Perylene-d12	24.163	264	1233459	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.140	96	20924	3.853	ng/uL	0.00
4) Pyridine-d5	3.546	84	234683	15.007	ng/ul	0.00
7) Phenol-d5	6.799	99	448683	20.966	ng/ul	-0.01
9) Bis-(2-Chloroethyl)eth...	6.969	67	276619	22.217	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.158	132	378373	22.708	ng/ul	-0.01
15) 4-Methylphenol-d8	8.334	113	380376	21.135	ng/ul	0.00
21) Nitrobenzene-d5	8.781	128	203293	24.298	ng/ul	-0.01
24) 2-Nitrophenol-d4	9.493	143	218650	25.141	ng/ul	-0.02
28) 2,4-Dichlorophenol-d3	10.028	165	378785	22.086	ng/ul	-0.01
31) 4-Chloroaniline-d4	10.546	131	541935	20.220	ng/ul	-0.02
46) Dimethylphthalate-d6	13.687	166	1360962	25.108	ng/ul	-0.01
49) Acenaphthylene-d8	13.963	160	1509879	24.582	ng/ul	-0.01
54) 4-Nitrophenol-d4	14.487	143	223772	19.780	ng/ul	0.00
60) Fluorene-d10	15.269	176	1145672	24.280	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.392	200	170638	21.558	ng/ul	-0.01
73) Anthracene-d10	17.122	188	1771775	26.055	ng/ul	-0.01
81) Pyrene-d10	19.422	212	1983378	30.601	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.957	264	1558505	26.645	ng/ul	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042624\
Data File : BN031243.D
Acq On : 26 Apr 2024 13:09
Operator : MA/JU
Sample : P2249-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
E08M8

Quant Time: Apr 26 13:39:26 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 19 22:56:25 2024
Response via : Initial Calibration

