

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042624\
 Data File : BN031240.D
 Acq On : 26 Apr 2024 11:11
 Operator : MA/JU
 Sample : P2249-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E0812

Quant Time: Apr 26 12:51:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.622	152	244138	20.000	ng/ul	-0.02
20) Naphthalene-d8	10.398	136	1093847	20.000	ng/ul	-0.02
38) Acenaphthene-d10	14.269	164	726146	20.000	ng/ul	-0.01
64) Phenanthrene-d10	17.022	188	1525627	20.000	ng/ul	-0.01
79) Chrysene-d12	21.257	240	1238738	20.000	ng/ul	-0.01
88) Perylene-d12	24.151	264	1192745	20.000	ng/ul	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.140	96	19142	3.663	ng/uL	0.00
4) Pyridine-d5	3.546	84	163556	10.868	ng/ul	0.00
7) Phenol-d5	6.799	99	400310	19.438	ng/ul	-0.01
9) Bis-(2-Chloroethyl)eth...	6.963	67	254935	21.276	ng/ul	-0.02
11) 2-Chlorophenol-d4	7.157	132	329384	20.541	ng/ul	-0.01
15) 4-Methylphenol-d8	8.328	113	339322	19.592	ng/ul	-0.01
21) Nitrobenzene-d5	8.775	128	182310	23.197	ng/ul	-0.02
24) 2-Nitrophenol-d4	9.493	143	182421	22.330	ng/ul	-0.02
28) 2,4-Dichlorophenol-d3	10.028	165	328362	20.382	ng/ul	-0.01
31) 4-Chloroaniline-d4	10.545	131	466629	18.535	ng/ul	-0.02
46) Dimethylphthalate-d6	13.686	166	1218641	24.117	ng/ul	-0.01
49) Acenaphthylene-d8	13.963	160	1438106	25.115	ng/ul	-0.01
54) 4-Nitrophenol-d4	14.486	143	180096	17.076	ng/ul	0.00
60) Fluorene-d10	15.269	176	1091203	24.806	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.392	200	136665	17.755	ng/ul	-0.01
73) Anthracene-d10	17.122	188	1699059	25.694	ng/ul	-0.01
81) Pyrene-d10	19.421	212	1903940	31.278	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.951	264	1499879	26.518	ng/ul	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042624\
Data File : BN031240.D
Acq On : 26 Apr 2024 11:11
Operator : MA/JU
Sample : P2249-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
E0812

Quant Time: Apr 26 12:51:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 19 22:56:25 2024
Response via : Initial Calibration

