

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042624\
 Data File : BN031246.D
 Acq On : 26 Apr 2024 15:07
 Operator : MA/JU
 Sample : P2249-07
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E08N1

Quant Time: Apr 26 15:37:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.622	152	269958	20.000	ng/ul	-0.02
20) Naphthalene-d8	10.404	136	1207436	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.269	164	798682	20.000	ng/ul	-0.01
64) Phenanthrene-d10	17.022	188	1644696	20.000	ng/ul	-0.01
79) Chrysene-d12	21.257	240	1336012	20.000	ng/ul	-0.01
88) Perylene-d12	24.157	264	1224930	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.140	96	22067	3.819	ng/uL	0.00
4) Pyridine-d5	3.546	84	270385	16.248	ng/ul	0.00
7) Phenol-d5	6.799	99	473261	20.782	ng/ul	-0.01
9) Bis-(2-Chloroethyl)eth...	6.969	67	291643	22.012	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.158	132	391921	22.103	ng/ul	-0.01
15) 4-Methylphenol-d8	8.334	113	404336	21.112	ng/ul	0.00
21) Nitrobenzene-d5	8.781	128	214979	24.780	ng/ul	-0.01
24) 2-Nitrophenol-d4	9.493	143	236117	26.183	ng/ul	-0.02
28) 2,4-Dichlorophenol-d3	10.028	165	400764	22.536	ng/ul	-0.01
31) 4-Chloroaniline-d4	10.552	131	602935	21.696	ng/ul	-0.01
46) Dimethylphthalate-d6	13.687	166	1428788	25.708	ng/ul	-0.01
49) Acenaphthylene-d8	13.963	160	1613711	25.622	ng/ul	-0.01
54) 4-Nitrophenol-d4	14.492	143	243637	21.003	ng/ul	0.00
60) Fluorene-d10	15.269	176	1197736	24.755	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.392	200	180222	21.719	ng/ul	-0.01
73) Anthracene-d10	17.122	188	1883659	26.424	ng/ul	-0.01
81) Pyrene-d10	19.427	212	2070272	31.535	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.957	264	1610611	27.728	ng/ul	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042624\
Data File : BN031246.D
Acq On : 26 Apr 2024 15:07
Operator : MA/JU
Sample : P2249-07
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
E08N1

Quant Time: Apr 26 15:37:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 19 22:56:25 2024
Response via : Initial Calibration

