

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN042820\
 Data File : BN010618.D
 Acq On : 28 Apr 2020 09:29
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 28 10:20:01 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 28 10:15:56 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.294	0.300	-2.0	119	0.00
3	n-Nitrosodimethylamine	0.191	0.231	-20.9	167#	0.00
4 S	2-Fluorophenol	0.876	0.951	-8.6	126	0.00
5 S	Phenol-d6	1.145	1.238	-8.1	126	0.00
6	bis(2-Chloroethyl)ether	0.992	1.042	-5.0	120	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	122	0.00
8 S	Nitrobenzene-d5	0.280	0.284	-1.4	129	0.00
9	Naphthalene	0.993	1.015	-2.2	126	0.00
10	Hexachlorobutadiene	0.239	0.245	-2.5	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.658	0.764	-16.1	143	0.00
12	2-Methylnaphthalene	0.714	0.737	-3.2	127	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	125	0.00
14 S	2,4,6-Tribromophenol	0.189	0.172	9.0	120	0.00
15 S	2-Fluorobiphenyl	1.456	1.448	0.5	124	0.00
16	Acenaphthylene	1.728	1.681	2.7	125	0.00
17	Acenaphthene	1.110	1.116	-0.5	127	0.00
18	Fluorene	1.454	1.454	0.0	127	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	126	0.00
20	4-Bromophenyl-phenylether	0.222	0.211	5.0	124	0.00
21	Hexachlorobenzene	0.254	0.263	-3.5	129	0.00
22	Pentachlorophenol	0.100	0.078	22.0	117	0.00
23	Phenanthrene	1.080	1.083	-0.3	127	0.00
24	Anthracene	0.988	0.960	2.8	126	0.00
25 SURR	Fluoranthene-d10	1.213	1.302	-7.3	136	0.00
26	Fluoranthene	1.345	1.261	6.2	119	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
28	Pvrene	1.346	1.569	-16.6	117	0.00
29 S	Terphenyl-d14	0.915	1.074	-17.4	120	0.00
30	Benzo(a)anthracene	1.305	1.304	0.1	101	0.00
31	Chrysene	1.300	1.316	-1.2	103	0.00
32	Bis(2-ethylhexyl)phthalate	0.670	0.807	-20.4	128	0.00
33	Indeno(1,2,3-cd)pyrene	1.149	1.321	-15.0	115	0.00
34 I	Perylene-d12	1.000	1.000	0.0	109	0.00
35	Benzo(b)fluoranthene	1.317	1.259	4.4	104	0.00
36	Benzo(k)fluoranthene	1.320	1.320	0.0	109	0.00
37 C	Benzo(a)pyrene	1.166	1.160	0.5	109	0.00
38	Dibenzo(a,h)anthracene	1.066	1.120	-5.1	114	0.00
39	Benzo(a,h,i)perylene	1.073	1.119	-4.3	113	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN042820\
Data File : BN010618.D
Acq On : 28 Apr 2020 09:29
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 28 10:20:01 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 28 10:15:56 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			