

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN042820\
 Data File : BN010618.D
 Acq On : 28 Apr 2020 09:29
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 28 10:20:01 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 28 10:15:56 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	116	0.00
2	1,4-Dioxane	0.400	0.407	-1.7	119	0.00
3	n-Nitrosodimethylamine	0.400	0.483	-20.7	167	0.00
4 S	2-Fluorophenol	0.400	0.434	-8.5	126	0.00
5 S	Phenol-d6	0.400	0.433	-8.2	126	0.00
6	bis(2-Chloroethyl)ether	0.400	0.420	-5.0	120	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	122	0.00
8 S	Nitrobenzene-d5	0.400	0.406	-1.5	129	0.00
9	Naphthalene	0.400	0.409	-2.2	126	0.00
10	Hexachlorobutadiene	0.400	0.409	-2.2	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.465	-16.3	143	0.00
12	2-Methylnaphthalene	0.400	0.413	-3.2	127	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	125	0.00
14 S	2,4,6-Tribromophenol	0.400	0.365	8.8	120	0.00
15 S	2-Fluorobiphenyl	0.400	0.398	0.5	124	0.00
16	Acenaphthylene	0.400	0.389	2.8	125	0.00
17	Acenaphthene	0.400	0.402	-0.5	127	0.00
18	Fluorene	0.400	0.400	0.0	127	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	126	0.00
20	4-Bromophenyl-phenylether	0.400	0.380	5.0	124	0.00
21	Hexachlorobenzene	0.400	0.414	-3.5	129	0.00
22	Pentachlorophenol	0.400	0.313	21.8	117	0.00
23	Phenanthrene	0.400	0.401	-0.3	127	0.00
24	Anthracene	0.400	0.389	2.8	126	0.00
25 SURR	Fluoranthene-d10	0.400	0.429	-7.2	136	0.00
26	Fluoranthene	0.400	0.375	6.3	119	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	100	0.00
28	Pvrene	0.400	0.466	-16.5	117	0.00
29 S	Terphenyl-d14	0.400	0.469	-17.2	120	0.00
30	Benzo(a)anthracene	0.400	0.400	0.0	101	0.00
31	Chrysene	0.400	0.405	-1.3	103	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.482	-20.5	128	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.460	-15.0	115	0.00
34 I	Perylene-d12	0.400	0.400	0.0	109	0.00
35	Benzo(b)fluoranthene	0.400	0.382	4.5	104	0.00
36	Benzo(k)fluoranthene	0.400	0.400	0.0	109	0.00
37 C	Benzo(a)pyrene	0.400	0.398	0.5	109	0.00
38	Dibenzo(a,h)anthracene	0.400	0.420	-5.0	114	0.00
39	Benzo(a,h,i)perylene	0.400	0.417	-4.2	113	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN042820\
Data File : BN010618.D
Acq On : 28 Apr 2020 09:29
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 28 10:20:01 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 28 10:15:56 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				