

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN042820\
 Data File : BN010640.D
 Acq On : 29 Apr 2020 01:28
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 29 02:04:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 28 11:13:37 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.294	0.296	-0.7	103	0.00
3	n-Nitrosodimethylamine	0.191	0.206	-7.9	130	0.00
4 S	2-Fluorophenol	0.876	0.925	-5.6	107	0.00
5 S	Phenol-d6	1.145	1.244	-8.6	110	0.00
6	bis(2-Chloroethyl)ether	0.992	1.024	-3.2	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
8 S	Nitrobenzene-d5	0.280	0.285	-1.8	115	0.00
9	Naphthalene	0.993	1.019	-2.6	112	0.00
10	Hexachlorobutadiene	0.239	0.238	0.4	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.658	0.760	-15.5	126	0.00
12	2-Methylnaphthalene	0.714	0.738	-3.4	113	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
14 S	2,4,6-Tribromophenol	0.189	0.163	13.8	100	0.00
15 S	2-Fluorobiphenyl	1.456	1.435	1.4	108	0.00
16	Acenaphthylene	1.728	1.646	4.7	108	0.00
17	Acenaphthene	1.110	1.140	-2.7	115	0.00
18	Fluorene	1.454	1.460	-0.4	112	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
20	4-Bromophenyl-phenylether	0.222	0.214	3.6	110	0.00
21	Hexachlorobenzene	0.254	0.263	-3.5	114	0.00
22	Pentachlorophenol	0.100	0.082	18.0	108	0.00
23	Phenanthrene	1.080	1.096	-1.5	113	0.00
24	Anthracene	0.988	0.938	5.1	108	0.00
25 SURR	Fluoranthene-d10	1.213	1.297	-6.9	119	0.00
26	Fluoranthene	1.345	1.284	4.5	106	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
28	Pvrene	1.346	1.543	-14.6	105	0.00
29 S	Terphenyl-d14	0.915	1.052	-15.0	107	0.00
30	Benzo(a)anthracene	1.305	1.292	1.0	91	-0.01
31	Chrysene	1.300	1.289	0.8	91	0.00
32	Bis(2-ethylhexyl)phthalate	0.670	0.659	1.6	94	0.00
33	Indeno(1,2,3-cd)pyrene	1.149	1.332	-15.9	105	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	96	-0.02
35	Benzo(b)fluoranthene	1.317	1.303	1.1	96	0.00
36	Benzo(k)fluoranthene	1.320	1.328	-0.6	97	0.00
37 C	Benzo(a)pyrene	1.166	1.151	1.3	96	-0.02
38	Dibenzo(a,h)anthracene	1.066	1.150	-7.9	104	-0.01
39	Benzo(a,h,i)perylene	1.073	1.158	-7.9	103	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN042820\
Data File : BN010640.D
Acq On : 29 Apr 2020 01:28
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 29 02:04:41 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 28 11:13:37 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			