

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN043019\
 Data File : BN005388.D
 Acq On : 01 May 2019 00:51
 Operator : JU/SJ
 Sample : K2596-09
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BFH66

Quant Time: May 01 07:33:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN041919MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 25 03:19:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	284421	20.00	ng/ul	0.00
18) Naphthalene-d8	10.35	136	1273580	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.22	164	837918	20.00	ng/ul	-0.01
61) Phenanthrene-d10	16.98	188	2081357	20.00	ng/ul	0.00
77) Chrysene-d12	21.20	240	2224945	20.00	ng/ul	-0.02
85) Perylene-d12	23.40	264	2317462	20.00	ng/ul	-0.04

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.22	96	43759	7.43	ng/uL	0.00
5) Phenol-d5	6.78	99	902688	40.41	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.94	67	505308	36.26	ng/ul	0.00
9) 2-Chlorophenol-d4	7.13	132	730218	42.71	ng/ul	0.00
13) 4-Methylphenol-d8	8.29	113	742431	42.52	ng/ul	0.00
19) Nitrobenzene-d5	8.73	128	370123	47.22	ng/ul	0.00
22) 2-Nitrophenol-d4	9.45	143	431775	54.33	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.98	165	791389	43.45	ng/ul	0.00
29) 4-Chloroaniline-d4	10.49	131	1045031	55.63	ng/ul	0.00
43) Dimethylphthalate-d6	13.65	166	2553172	41.98	ng/ul	0.00
46) Acenaphthylene-d8	13.92	160	3160335	42.30	ng/ul	0.00
51) 4-Nitrophenol-d4	14.43	143	443561	46.18	ng/ul	0.00
57) Fluorene-d10	15.23	176	2227352	43.63	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.36	200	396400	42.35	ng/ul	0.00
70) Anthracene-d10	17.08	188	3597030	41.23	ng/ul	0.00
78) Pyrene-d10	19.39	212	4233250	37.19	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.26	264	4682931	46.24	ng/ul	-0.04


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
44) Dimethylphthalate	13.69	163	254375	4.197	ng/ul#	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN043019\
Data File : BN005388.D
Acq On : 01 May 2019 00:51
Operator : JU/SJ
Sample : K2596-09
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BFH66

Quant Time: May 01 07:33:56 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN041919MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Apr 25 03:19:34 2019
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 4.197 ng/ul

RT: 13.69 min Scan# 1819

Delta R.T. -0.01 min

Lab File: BN005388.D

Acq: 01 May 2019 00:51

Instrument :

BNA_N

ClientSampleId :

BFH66

Tot Ion:163 Resp: 254375

Ion Ratio Lower Upper

163 100

194 5.1 3.4 5.0#

164 10.6 8.1 12.1

