


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN043019\
 Data File : BN005366.D
 Acq On : 30 Apr 2019 08:29
 Operator : JU/SJ
 Sample : DFTPP59
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP59

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN041919MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Apr 25 03:19:34 2019



AutoFind: Scans 2457, 2458, 2459; Background Corrected with Scan 2452

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.9	213632	PASS
68	69	0.00	2	1.9	4211	PASS
69	198	0.00	100	32.0	221824	PASS
70	69	0.00	2	0.4	825	PASS
127	198	10	80	40.5	280469	PASS
197	198	0.00	2	0.2	1324	PASS
198	198	100	100	100.0	692138	PASS
199	198	5	9	6.7	46666	PASS
275	198	10	60	26.5	183341	PASS
365	198	1	100	2.9	20313	PASS
441	198	0.01	100	12.1	83986	PASS
442	198	50	100	78.6	543829	PASS
443	442	15	24	19.3	104848	PASS

m/z	Abundance
39.90	2910.0
40.90	318.0
44.00	10909.0
55.10	359.0
73.10	518.0
95.90	485.0
132.90	420.0
165.00	478.0
167.10	491.0
190.90	871.0
192.00	679.0
193.10	593.0
207.00	3902.0
208.00	1070.0
209.00	772.0
281.00	588.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

m/z	Abundance
37.00	972.0
38.00	2183.0
39.00	12585.0
40.00	3534.0
41.10	468.0
44.00	11444.0
45.10	394.0
48.90	1955.0
50.00	45536.0
51.00	178688.0
52.00	9780.0
53.10	430.0
54.90	1139.0
56.00	6512.0
57.00	13142.0
58.00	742.0
61.00	2506.0
62.00	2488.0
63.00	7945.0
64.00	1114.0
65.00	4499.0
67.00	444.0
68.00	3878.0
69.00	183552.0
69.90	989.0
73.00	1963.0
74.00	17584.0
75.00	29136.0
76.00	9971.0
77.00	206528.0
78.00	15052.0
78.90	12859.0
80.00	11107.0
81.00	14760.0
82.00	3671.0
83.00	3654.0
83.80	525.0
84.90	3791.0
86.00	4512.0
86.90	1841.0
87.90	1056.0
89.00	502.0
91.00	2893.0
92.00	4341.0
93.00	22552.0
93.90	1404.0
95.10	695.0
95.90	1370.0
97.20	650.0
98.00	17056.0
99.00	15205.0
100.00	1404.0
101.00	8510.0
102.90	3405.0
103.90	7197.0
105.00	4714.0
106.10	1243.0
107.00	63864.0
108.00	9948.0
110.00	118088.0
111.00	18312.0
111.90	1944.0
113.00	589.0
114.00	312.0
115.00	858.0
116.00	3860.0
117.00	43136.0
118.00	3532.0
118.80	467.0
120.00	662.0
122.00	5053.0
123.00	5879.0
124.00	3359.0
124.90	2772.0
127.00	220736.0
128.00	17968.0
129.00	83688.0
130.00	7930.0
130.90	1619.0
132.00	749.0
132.80	734.0
133.90	2203.0
135.00	7314.0
136.00	2834.0
137.00	3571.0
137.80	604.0
138.90	340.0
139.80	901.0
140.20	982.0
141.00	10989.0
142.00	3151.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

143.00	2096.0
143.90	757.0
144.80	403.0
146.00	1758.0
147.00	5088.0
148.00	11591.0
148.90	2542.0
150.00	883.0
151.00	1541.0
151.60	945.0
152.20	791.0
152.90	3420.0
154.00	2578.0
155.10	6676.0
156.00	8913.0
157.00	1497.0
158.00	2061.0
159.00	1749.0
160.00	3797.0
161.00	5980.0
162.00	1690.0
162.90	812.0
164.00	587.0
165.00	3436.0
166.00	3570.0
167.00	22032.0
168.00	11701.0
169.00	2079.0
170.10	530.0
170.90	1004.0
172.00	1948.0
173.00	2303.0
174.00	5329.0
175.00	7758.0
175.90	2461.0
177.00	4331.0
178.20	1356.0
179.00	17016.0
180.00	11614.0
181.00	5153.0
181.90	1108.0
182.80	779.0
183.80	1086.0
185.00	8758.0
186.00	63632.0
187.00	17984.0
188.10	2677.0
189.00	4313.0
189.90	991.0
190.90	2649.0
191.90	5925.0
193.00	6803.0
194.00	1479.0
195.00	639.0
196.00	13518.0
197.10	3974.0
197.90	497920.0
198.90	34728.0
200.00	2411.0
201.40	2910.0
203.00	3317.0
204.00	16992.0
205.00	28384.0
206.00	124328.0
207.00	21560.0
208.00	5416.0
209.00	2040.0
210.10	2055.0
211.00	5200.0
213.00	503.0
214.90	1597.0
215.90	2624.0
216.90	31464.0
217.90	4165.0
219.00	547.0
219.80	369.0
221.00	20176.0
223.00	6871.0
224.00	65600.0
225.00	17848.0
226.00	1813.0
227.00	26648.0
228.00	4354.0
229.00	5643.0
229.90	1314.0
231.00	2520.0
232.10	532.0
233.00	534.0
234.00	2047.0
234.90	2366.0
236.00	1328.0
237.00	2225.0
238.10	340.0
239.00	1232.0
240.10	692.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

240.90	1657.0
242.00	3355.0
243.00	4187.0
244.00	49544.0
245.00	6415.0
246.00	9605.0
246.90	2061.0
248.00	614.0
249.00	1938.0
250.10	498.0
250.90	622.0
252.00	607.0
253.00	983.0
255.00	263360.0
256.00	38024.0
257.00	3237.0
258.00	13830.0
259.10	2094.0
260.00	361.0
260.90	372.0
263.90	334.0
265.00	6154.0
265.90	815.0
270.90	672.0
272.00	836.0
273.00	8356.0
274.00	19984.0
275.00	115272.0
276.00	15767.0
277.00	9811.0
278.00	1794.0
278.90	329.0
281.00	602.0
282.00	608.0
283.00	1102.0
284.00	745.0
285.00	1692.0
286.00	357.0
288.90	354.0
290.00	395.0
292.10	409.0
292.90	2256.0
294.00	412.0
296.00	27080.0
296.90	3624.0
298.00	382.0
301.00	462.0
302.00	407.0
303.00	4065.0
304.10	1184.0
308.10	410.0
310.10	346.0
313.90	1643.0
315.00	3686.0
316.00	2057.0
317.00	369.0
321.10	690.0
323.00	8654.0
324.00	1621.0
326.90	1661.0
328.00	735.0
332.00	965.0
332.90	666.0
334.00	5623.0
335.00	1573.0
340.90	1056.0
346.00	1678.0
352.00	2812.0
353.00	2150.0
354.00	2752.0
354.90	886.0
365.00	11724.0
365.90	2226.0
370.00	366.0
371.00	478.0
372.00	4762.0
373.00	1083.0
383.00	1567.0
383.90	482.0
389.90	366.0
391.00	635.0
400.80	428.0
402.00	1735.0
403.00	2242.0
404.00	1276.0
421.00	2405.0
422.10	2438.0
423.00	16928.0
424.00	4611.0
424.80	418.0
439.20	391.0
441.00	42856.0
442.00	280064.0
443.00	54456.0
444.00	5344.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

m/z	Abundance
37.10	1130.0
38.10	3017.0
39.00	16776.0
40.00	3768.0
41.20	596.0
44.00	9786.0
44.90	484.0
49.10	1482.0
50.00	65072.0
51.00	253504.0
52.00	12523.0
53.00	392.0
55.00	1385.0
56.00	8037.0
57.00	18112.0
58.00	1018.0
59.00	381.0
61.00	3698.0
62.00	4440.0
63.00	9937.0
63.90	1391.0
65.00	5452.0
66.00	323.0
68.10	4817.0
69.00	268160.0
70.10	1496.0
73.10	2499.0
74.00	27040.0
75.00	41480.0
76.10	14175.0
77.00	294016.0
78.00	20456.0
79.00	19384.0
80.00	14617.0
81.00	21824.0
82.00	6152.0
83.00	5585.0
84.00	636.0
85.00	5528.0
86.00	6588.0
87.00	3357.0
88.00	937.0
89.00	402.0
91.00	5531.0
91.90	4903.0
93.00	33288.0
94.00	2433.0
95.00	958.0
96.00	2057.0
97.20	753.0
98.00	24912.0
99.00	20768.0
100.10	1862.0
101.00	12220.0
101.90	697.0
103.00	4126.0
103.90	8065.0
105.00	7590.0
106.00	2785.0
107.00	96816.0
108.00	15364.0
109.00	2995.0
110.00	176704.0
111.00	26936.0
112.00	3809.0
113.00	993.0
113.70	403.0
114.90	784.0
116.00	6383.0
117.00	68608.0
118.00	4786.0
119.00	695.0
119.90	1226.0
120.90	363.0
121.90	5756.0
123.00	10441.0
123.90	5156.0
124.90	3401.0
127.00	336960.0
128.00	25216.0
129.00	127952.0
130.00	11498.0
131.00	2596.0
132.10	1073.0
132.80	1031.0
134.00	3782.0
135.00	9198.0
135.90	4186.0
137.00	5227.0
137.90	1349.0
138.80	961.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

138.90	950.0
140.00	1653.0
141.00	16448.0
142.00	5813.0
142.90	4350.0
144.10	1417.0
145.00	953.0
146.00	3068.0
147.00	8227.0
148.00	18888.0
149.00	4276.0
150.00	1173.0
151.10	2533.0
151.80	1891.0
153.00	6258.0
153.90	5232.0
155.00	9644.0
156.00	14363.0
157.10	2934.0
158.00	3384.0
158.90	1824.0
160.00	5127.0
161.10	8030.0
162.00	2543.0
163.00	722.0
163.90	1010.0
165.00	6957.0
166.00	6099.0
167.00	31984.0
168.00	17184.0
169.00	3117.0
169.90	1422.0
171.00	1593.0
171.90	3858.0
173.00	4338.0
174.00	9057.0
175.00	14462.0
176.00	3996.0
176.90	6675.0
177.90	3213.0
179.00	28104.0
180.00	17080.0
180.90	8785.0
182.00	1168.0
182.90	985.0
184.00	2585.0
185.00	12674.0
186.00	105264.0
187.00	30264.0
188.00	3507.0
189.00	7794.0
189.90	1889.0
191.00	3605.0
192.00	8732.0
193.00	10770.0
194.00	2115.0
194.90	1733.0
196.00	21848.0
198.00	813952.0
199.00	55120.0
200.00	4160.0
201.50	3285.0
202.90	5369.0
204.00	26616.0
205.00	49216.0
206.00	207296.0
207.00	30768.0
208.00	7751.0
208.90	2455.0
210.20	2817.0
210.90	6908.0
213.00	521.0
215.00	2484.0
216.00	4525.0
217.00	52488.0
217.90	6979.0
219.10	989.0
221.00	34288.0
223.00	12365.0
224.00	111792.0
225.00	29328.0
226.20	3199.0
227.00	44808.0
227.90	6532.0
229.00	10015.0
230.00	1347.0
231.00	4412.0
232.00	685.0
232.90	684.0
234.00	2747.0
234.90	3470.0
236.00	2744.0
237.00	3813.0
238.00	648.0
239.00	1278.0

Instrument :

BNA_N

ClientSampleId :

DFTPP59

239.90	1184.0
240.90	2073.0
242.00	5764.0
243.10	6638.0
244.00	90784.0
245.00	11791.0
246.00	16736.0
247.00	3298.0
247.90	1212.0
249.00	3337.0
250.00	932.0
251.00	988.0
251.90	1180.0
253.00	1883.0
255.00	475584.0
256.00	67784.0
257.00	4985.0
258.00	26168.0
259.00	3781.0
260.00	842.0
260.90	595.0
263.90	746.0
265.00	8876.0
266.00	1628.0
270.00	552.0
270.90	859.0
272.00	1454.0
273.00	13548.0
274.00	36856.0
275.00	218304.0
276.00	29176.0
277.00	16400.0
278.00	2251.0
279.00	980.0
281.00	797.0
282.10	370.0
282.90	1643.0
284.00	1332.0
285.00	3063.0
286.00	673.0
288.90	883.0
290.00	687.0
291.00	544.0
291.90	1006.0
293.00	3404.0
294.00	1138.0
295.10	1598.0
296.00	51920.0
297.00	7665.0
298.00	676.0
301.00	1090.0
301.90	1073.0
303.00	6545.0
304.10	1826.0
305.00	365.0
308.00	568.0
308.90	532.0
310.10	900.0
314.00	2741.0
315.00	5894.0
316.10	3053.0
316.90	890.0
320.20	433.0
321.00	1671.0
322.20	1092.0
323.00	17336.0
324.00	3796.0
325.20	411.0
327.00	3384.0
328.00	1927.0
332.00	1133.0
333.00	1772.0
334.00	11736.0
335.00	3449.0
336.00	535.0
341.00	1569.0
342.00	774.0
346.00	3726.0
346.90	388.0
347.10	391.0
350.80	350.0
352.00	5440.0
353.00	4070.0
354.00	5283.0
355.10	1103.0
359.10	579.0
365.00	23616.0
366.00	3358.0
367.00	405.0
370.00	525.0
371.00	1448.0
372.00	10403.0
373.10	2735.0
377.00	336.0
383.10	2029.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

384.00	880.0
390.10	1265.0
390.90	845.0
392.10	589.0
401.10	735.0
402.00	3914.0
403.00	5545.0
404.00	1904.0
405.10	344.0
421.10	5071.0
422.10	5014.0
423.00	36320.0
424.00	8733.0
425.00	1065.0
439.00	387.0
439.50	420.0
441.10	97408.0
442.00	627520.0
443.00	117752.0
444.00	11095.0
445.00	1111.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

m/z	Abundance
37.10	815.0
38.00	2788.0
39.00	14863.0
40.00	3196.0
41.00	769.0
43.10	310.0
44.00	10815.0
44.90	674.0
49.00	1587.0
50.00	49632.0
51.00	208704.0
52.00	10130.0
53.00	489.0
55.00	1039.0
56.00	5878.0
57.00	15296.0
58.00	810.0
59.00	434.0
61.00	2920.0
61.90	3650.0
63.00	8434.0
64.10	1434.0
65.00	5089.0
67.10	351.0
68.10	3938.0
69.00	213760.0
70.10	980.0
73.00	2396.0
74.00	21256.0
75.00	34472.0
76.10	11776.0
77.00	250816.0
78.00	17160.0
79.00	16880.0
80.00	12081.0
81.00	17464.0
81.90	4130.0
83.00	4075.0
83.90	486.0
84.90	4586.0
86.00	4621.0
87.00	2690.0
88.00	801.0
89.00	718.0
91.00	4660.0
92.00	4415.0
93.00	29496.0
94.00	1800.0
95.90	1555.0
96.20	1438.0
97.20	585.0
98.00	20776.0
99.00	16008.0
99.90	1927.0
101.00	9483.0
102.10	513.0
102.90	3580.0
104.00	7018.0
105.00	5850.0
106.10	2366.0
107.00	82280.0
108.00	12501.0
109.10	2434.0
110.00	148928.0
111.00	23056.0
111.90	2806.0
113.10	1207.0
115.00	775.0
116.00	3789.0
117.00	60184.0
118.00	4120.0
118.80	720.0
120.00	792.0
121.00	528.0
122.00	4687.0
122.90	8173.0
123.90	3748.0
124.90	3192.0
127.00	283712.0
128.00	21512.0
129.00	109408.0
130.00	8865.0
131.00	1712.0
132.00	841.0
132.80	673.0
133.90	3102.0
135.00	8160.0
136.00	3849.0
137.00	4874.0
137.70	1098.0
138.00	1070.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

138.90	712.0
140.00	1402.0
141.00	14692.0
142.10	5082.0
143.00	3564.0
143.90	693.0
145.00	1048.0
146.10	3344.0
147.00	7127.0
148.00	16007.0
149.00	3492.0
150.00	857.0
150.90	1626.0
151.80	1635.0
152.90	4783.0
154.00	4460.0
155.00	7942.0
156.00	13090.0
157.10	2804.0
157.90	2289.0
159.00	1971.0
160.00	5355.0
161.00	6734.0
161.90	2353.0
163.00	1031.0
163.90	428.0
165.00	6390.0
166.00	4312.0
167.00	29448.0
168.00	14618.0
169.00	2801.0
170.10	965.0
170.90	1666.0
172.00	4065.0
173.10	3255.0
174.00	6931.0
175.00	12941.0
176.00	4196.0
177.00	5461.0
178.00	2425.0
179.00	23904.0
180.00	14036.0
181.00	7869.0
182.00	1279.0
182.90	1118.0
184.10	1833.0
185.00	11505.0
186.00	100256.0
187.00	28408.0
188.00	2716.0
188.90	5791.0
189.90	1481.0
191.00	3315.0
192.00	8964.0
193.00	9381.0
194.00	2385.0
195.00	1336.0
196.00	20320.0
197.90	764544.0
198.90	50152.0
200.00	3988.0
201.30	3003.0
203.00	5510.0
204.00	23664.0
205.00	47488.0
206.00	190720.0
207.00	31552.0
208.00	6515.0
208.90	2106.0
210.20	3357.0
211.10	8590.0
212.00	1049.0
213.00	618.0
213.80	499.0
215.00	2119.0
216.00	4071.0
217.00	49408.0
217.90	6733.0
219.00	691.0
221.00	35352.0
223.00	12600.0
224.00	112872.0
225.00	28552.0
226.00	2754.0
227.00	43176.0
228.00	5748.0
229.00	8919.0
229.90	1210.0
231.00	4605.0
232.10	581.0
233.00	668.0
233.90	3295.0
235.00	2975.0
236.00	1355.0
237.00	3165.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

238.00	562.0
238.90	1448.0
239.90	1304.0
240.90	2307.0
242.00	4888.0
243.00	6491.0
244.00	83528.0
245.00	10532.0
246.00	16188.0
246.90	3285.0
247.80	543.0
249.00	2924.0
249.90	721.0
251.10	512.0
252.10	1092.0
253.00	1940.0
255.00	452096.0
256.00	66888.0
257.00	4508.0
258.00	25488.0
259.00	3358.0
260.00	559.0
260.80	737.0
262.90	363.0
263.70	410.0
264.20	524.0
264.90	9429.0
265.90	1993.0
269.90	393.0
271.00	892.0
272.10	1462.0
273.00	12198.0
274.00	35304.0
275.00	216448.0
276.00	28552.0
277.00	15202.0
277.90	2783.0
279.10	591.0
281.00	485.0
282.00	632.0
283.10	1686.0
284.00	1265.0
285.00	3939.0
286.00	589.0
289.00	633.0
290.10	768.0
290.90	487.0
291.90	803.0
293.00	3580.0
293.90	912.0
295.10	1815.0
296.00	56232.0
297.00	7330.0
298.00	519.0
300.90	770.0
302.10	954.0
303.10	6523.0
303.90	2143.0
307.90	1053.0
309.00	555.0
310.00	1079.0
312.90	508.0
313.90	2561.0
315.00	5857.0
316.00	3879.0
317.00	567.0
320.00	372.0
320.90	1776.0
322.00	1024.0
323.00	19864.0
324.00	2785.0
324.90	444.0
327.00	3031.0
328.00	1670.0
329.10	454.0
331.90	1134.0
333.00	1705.0
334.00	12815.0
335.10	3279.0
336.00	337.0
339.80	364.0
341.00	2191.0
341.90	649.0
346.00	4373.0
346.90	996.0
351.00	390.0
352.00	6350.0
353.10	4679.0
354.00	6775.0
355.10	1275.0
359.10	679.0
365.00	25600.0
366.00	3745.0
366.90	499.0
370.10	839.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59

371.00	2050.0
372.00	12497.0
373.10	2830.0
373.80	325.0
376.90	401.0
383.00	2772.0
384.00	804.0
385.00	336.0
389.90	1254.0
391.10	1217.0
391.90	719.0
400.80	621.0
402.00	4755.0
403.00	6939.0
403.90	3142.0
421.00	5421.0
422.00	5844.0
423.00	39224.0
424.10	8366.0
425.10	1021.0
438.30	313.0
439.00	527.0
441.00	111696.0
442.00	723904.0
443.00	142336.0
444.10	13683.0
444.90	760.0

Instrument :
BNA_N
ClientSampleId :
DFTPP59