


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050120\
 Data File : BN010682.D
 Acq On : 01 May 2020 10:57
 Operator : CG/JU
 Sample : DFTPP74
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP74

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN040920MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri May 01 12:05:34 2020



AutoFind: Scans 2488, 2489, 2490; Background Corrected with Scan 2482

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	21.0	139997	PASS
68	69	0.00	2	2.0	4883	PASS
69	198	0.00	100	37.4	249301	PASS
70	69	0.00	2	0.5	1190	PASS
127	198	10	80	43.5	289877	PASS
197	198	0.00	2	0.2	1626	PASS
198	198	100	100	100.0	666709	PASS
199	198	5	9	6.6	44144	PASS
275	198	10	60	27.6	183722	PASS
365	198	1	100	4.2	27984	PASS
441	198	0.01	100	12.5	83664	PASS
442	198	50	100	83.2	554474	PASS
443	442	15	24	17.8	98845	PASS

Scan 2482 (17.386 min): BN010682.D
DFTPP74

m/z	Abundance
40.00	469.0
44.00	2109.0
73.10	1115.0
77.10	491.0
83.90	425.0
95.90	449.0
132.90	670.0
146.90	319.0
191.00	880.0
193.00	421.0
207.00	4193.0
208.10	927.0
209.00	682.0
221.10	394.0
281.00	1128.0
282.00	409.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

m/z	Abundance
36.90	976.0
38.10	2114.0
39.00	12244.0
40.00	816.0
40.90	487.0
44.00	1754.0
45.00	391.0
49.00	1764.0
50.00	38904.0
51.00	145792.0
52.00	6717.0
53.10	377.0
55.00	1056.0
56.00	5334.0
57.00	10336.0
58.00	485.0
61.00	2877.0
62.00	3075.0
63.00	9787.0
63.90	1703.0
65.10	3649.0
67.00	526.0
68.00	4973.0
69.00	243136.0
70.00	1302.0
73.10	3344.0
74.00	27344.0
75.00	43672.0
76.00	14881.0
77.00	283264.0
78.00	20752.0
79.00	20832.0
80.00	14226.0
81.00	20216.0
82.00	4851.0
83.00	3931.0
83.90	884.0
85.00	3752.0
86.00	6981.0
86.90	3156.0
88.00	1004.0
89.00	890.0
91.00	3927.0
91.90	5367.0
93.00	32632.0
94.10	2288.0
95.00	565.0
96.00	1364.0
96.90	936.0
98.00	27448.0
99.00	22496.0
100.00	1841.0
101.00	10573.0
102.00	729.0
102.90	3175.0
103.90	7742.0
104.90	6692.0
105.90	2022.0
107.00	83184.0
108.00	12172.0
109.00	2540.0
110.00	150976.0
111.00	24624.0
112.00	3960.0
113.00	789.0
113.90	423.0
115.00	869.0
116.00	4611.0
117.00	78528.0
118.00	6222.0
119.00	1083.0
120.00	1264.0
121.00	503.0
121.90	6163.0
123.00	9570.0
124.00	4517.0
125.00	2785.0
127.00	279232.0
128.00	22384.0
129.00	116528.0
130.00	10740.0
131.00	2093.0
131.80	557.0
133.00	786.0
133.90	3167.0
135.00	8600.0
136.00	3889.0
136.90	6352.0
137.80	1675.0
139.00	454.0
140.10	1540.0
141.00	14625.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

142.10	5183.0
142.90	3715.0
144.00	695.0
144.90	1201.0
146.00	2813.0
146.90	8131.0
148.00	17336.0
148.90	3988.0
149.90	1750.0
151.00	1901.0
151.80	1182.0
153.10	5061.0
154.00	3796.0
155.00	8606.0
156.00	13750.0
156.90	1970.0
157.90	2336.0
159.00	2029.0
159.90	4815.0
160.90	5887.0
161.90	2077.0
163.10	644.0
163.90	785.0
165.00	4495.0
166.10	3712.0
167.00	32960.0
168.00	21672.0
169.00	2926.0
170.00	2091.0
170.90	1003.0
171.90	3014.0
173.00	3861.0
174.00	6589.0
175.00	12769.0
176.00	3264.0
177.00	5669.0
178.00	2134.0
179.00	23032.0
180.00	16864.0
181.00	6759.0
182.00	1432.0
182.80	504.0
184.00	1569.0
185.00	11129.0
186.00	84616.0
187.00	26024.0
188.00	3352.0
188.90	4933.0
190.00	935.0
191.00	3457.0
192.00	8535.0
193.00	8103.0
194.00	2013.0
195.10	1774.0
196.00	17304.0
197.10	4880.0
197.90	602240.0
198.90	42456.0
200.00	3280.0
201.40	2511.0
203.00	4209.0
204.00	23584.0
205.00	39752.0
206.00	161792.0
207.10	28776.0
208.00	7862.0
209.00	2864.0
210.10	2631.0
211.00	6028.0
213.00	491.0
215.00	1566.0
216.00	4131.0
217.00	45912.0
218.00	6619.0
218.90	834.0
221.00	26416.0
221.90	5820.0
223.00	9605.0
224.00	85488.0
225.00	24528.0
226.10	4343.0
227.00	39104.0
228.00	6826.0
229.00	9059.0
230.10	1224.0
231.10	3335.0
231.90	397.0
233.00	692.0
234.00	2786.0
235.00	2718.0
236.00	2445.0
237.00	4831.0
238.00	805.0
239.00	1941.0
240.00	1646.0
241.00	1850.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

242.00	5040.0
243.10	6097.0
244.00	63984.0
245.00	8973.0
246.00	13279.0
246.90	3301.0
248.00	666.0
249.00	3297.0
249.90	513.0
251.00	701.0
252.00	890.0
253.00	1133.0
255.00	332032.0
256.00	51328.0
257.00	5201.0
258.00	20680.0
259.00	3633.0
260.10	615.0
260.90	899.0
263.90	778.0
265.00	8110.0
266.00	1201.0
266.90	818.0
270.00	767.0
270.90	969.0
272.00	1076.0
273.00	10331.0
274.00	28416.0
275.00	147584.0
276.00	19200.0
277.00	12732.0
277.90	3070.0
278.90	475.0
281.00	1804.0
282.00	739.0
282.90	1427.0
284.00	1036.0
285.10	2199.0
288.80	626.0
290.00	483.0
291.80	683.0
293.00	3043.0
294.10	908.0
296.00	45712.0
297.00	6242.0
298.10	611.0
301.00	828.0
302.00	855.0
303.00	5331.0
303.90	1780.0
308.10	547.0
309.00	612.0
309.80	548.0
313.10	423.0
314.10	2498.0
315.00	4883.0
316.00	2523.0
316.90	447.0
320.90	1984.0
322.00	801.0
323.00	10619.0
324.00	2998.0
325.90	305.0
327.00	3505.0
328.00	929.0
329.00	552.0
332.00	1010.0
333.00	1404.0
334.10	8889.0
335.00	1949.0
336.00	320.0
341.00	1967.0
342.20	380.0
342.70	306.0
346.00	2519.0
347.00	1144.0
351.00	384.0
352.00	3894.0
353.00	3100.0
354.00	3694.0
355.00	1394.0
358.90	625.0
365.00	19576.0
366.00	2824.0
366.90	441.0
369.90	502.0
371.00	777.0
372.00	6039.0
373.10	1538.0
373.90	421.0
382.90	2118.0
384.20	505.0
390.00	878.0
391.10	1103.0
392.00	619.0
400.90	580.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

401.90	2444.0
402.90	3936.0
403.90	1629.0
405.00	580.0
421.00	2822.0
422.00	3988.0
423.00	22752.0
424.00	8381.0
425.10	1160.0
441.00	56144.0
442.00	332928.0
443.00	69720.0
444.00	6424.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

m/z	Abundance
36.90	1002.0
38.00	2839.0
39.00	12471.0
40.10	1095.0
41.00	621.0
42.80	349.0
43.90	1551.0
44.90	314.0
47.80	318.0
49.00	1211.0
50.00	45288.0
51.00	166848.0
52.00	8991.0
55.00	1064.0
56.00	7575.0
57.00	12772.0
58.00	620.0
61.00	3499.0
62.00	3761.0
63.00	10804.0
63.90	1025.0
65.00	5101.0
66.10	693.0
66.80	407.0
68.00	6442.0
69.00	310464.0
70.00	1429.0
73.00	3932.0
74.00	34528.0
75.00	53080.0
76.00	18392.0
77.00	335040.0
78.00	22288.0
79.00	25072.0
80.00	16616.0
80.90	24472.0
82.00	4974.0
83.00	5578.0
83.90	1115.0
84.90	4547.0
85.90	8385.0
87.00	3106.0
88.00	841.0
89.00	704.0
91.00	5769.0
92.00	7142.0
93.00	42272.0
93.90	3787.0
95.10	629.0
96.00	1839.0
97.10	907.0
98.00	33696.0
99.00	28040.0
99.90	2134.0
101.00	13265.0
102.00	1041.0
103.00	4369.0
104.00	8466.0
105.00	8973.0
106.00	2653.0
107.00	108168.0
108.00	15584.0
110.00	186240.0
111.00	28648.0
111.90	3619.0
112.90	927.0
114.90	656.0
116.10	5929.0
117.00	99016.0
118.00	8047.0
119.00	877.0
120.00	1164.0
120.80	421.0
122.00	7108.0
123.00	11061.0
124.00	5342.0
125.00	4804.0
127.00	344256.0
128.00	27048.0
129.00	142656.0
130.00	12993.0
131.00	2089.0
132.10	1171.0
133.00	991.0
134.00	4943.0
135.00	10728.0
135.90	4586.0
137.00	7591.0
138.00	1555.0
138.90	1033.0
140.00	1270.0
141.00	19424.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

142.00	5954.0
143.00	4585.0
143.90	1676.0
145.10	1060.0
146.00	4113.0
147.10	10552.0
148.00	22496.0
149.00	6344.0
149.90	1385.0
151.00	2322.0
152.00	1104.0
153.00	6854.0
154.00	4843.0
155.00	11493.0
156.00	16112.0
157.00	3254.0
158.00	4458.0
159.00	2955.0
160.00	6169.0
161.00	8391.0
162.10	2804.0
162.80	802.0
164.00	1257.0
165.00	7473.0
166.00	5579.0
167.00	40776.0
168.00	27376.0
169.00	3783.0
170.00	1356.0
171.10	1697.0
172.00	5062.0
172.90	4828.0
174.00	8694.0
175.00	16792.0
176.00	4792.0
177.00	7969.0
178.00	2895.0
179.00	30880.0
180.00	24368.0
181.00	10693.0
182.10	1681.0
182.60	547.0
183.00	490.0
183.90	2116.0
185.00	17032.0
186.00	116360.0
187.00	33688.0
188.00	3834.0
189.00	8581.0
189.90	878.0
190.90	4248.0
192.00	11263.0
193.00	11653.0
193.90	2439.0
195.10	1778.0
196.00	21656.0
198.00	796096.0
198.90	52528.0
199.90	4243.0
201.50	3421.0
203.00	5777.0
204.00	28880.0
205.10	54928.0
206.00	220416.0
207.10	33344.0
208.00	8682.0
208.90	3356.0
210.30	4691.0
211.10	10204.0
212.00	1624.0
213.00	701.0
213.80	429.0
215.00	2723.0
216.20	4677.0
217.00	63064.0
218.00	6632.0
219.00	695.0
221.00	29608.0
221.90	7888.0
223.00	15306.0
224.00	126296.0
225.00	32240.0
226.00	4378.0
227.00	56248.0
228.00	8868.0
229.00	11803.0
230.00	1306.0
231.00	4562.0
232.10	798.0
233.00	1031.0
234.00	3464.0
234.90	4472.0
236.00	3142.0
237.00	6096.0
238.10	599.0
239.00	2814.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

239.90	1696.0
240.90	3665.0
242.00	6252.0
243.10	7840.0
244.00	97952.0
245.00	11317.0
246.00	21160.0
247.00	4180.0
247.90	1337.0
249.00	3655.0
249.90	945.0
251.00	788.0
252.10	906.0
253.00	2204.0
255.00	501120.0
256.00	67624.0
257.00	6362.0
258.00	29448.0
259.00	4575.0
259.90	949.0
261.00	971.0
263.00	478.0
263.70	431.0
264.20	409.0
265.00	13517.0
266.00	2047.0
267.00	577.0
270.00	677.0
270.90	1279.0
272.10	1498.0
273.00	14592.0
274.00	38640.0
275.00	219456.0
276.00	26848.0
277.00	20960.0
277.90	3423.0
279.00	812.0
281.00	1606.0
282.00	805.0
283.00	2497.0
283.90	1883.0
285.10	3813.0
285.90	603.0
288.90	771.0
290.00	1000.0
290.80	693.0
292.00	551.0
292.90	4289.0
294.00	1015.0
295.00	1524.0
296.00	66160.0
297.00	8086.0
297.90	887.0
299.00	337.0
300.90	846.0
301.90	1166.0
303.00	7069.0
304.10	1651.0
304.90	546.0
307.70	696.0
308.90	511.0
309.80	392.0
310.10	450.0
312.00	380.0
313.10	704.0
314.00	2821.0
315.00	7734.0
316.10	3544.0
317.10	713.0
320.00	337.0
321.00	2588.0
322.20	1235.0
323.00	22592.0
324.00	3726.0
325.00	397.0
326.00	317.0
326.90	4551.0
327.90	2368.0
329.00	540.0
332.00	1496.0
333.00	2349.0
334.00	13556.0
335.00	4124.0
335.90	315.0
339.00	332.0
339.80	473.0
341.00	2499.0
341.90	883.0
342.90	434.0
345.90	4736.0
347.00	1466.0
352.00	6148.0
353.10	4395.0
354.00	8165.0
354.90	1720.0
359.20	460.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

365.00	34480.0
365.90	3929.0
366.90	470.0
370.00	519.0
371.00	1787.0
371.90	10425.0
373.00	2788.0
373.90	326.0
376.70	351.0
382.90	3672.0
384.00	1015.0
385.00	389.0
390.00	1877.0
391.00	1254.0
392.00	697.0
401.10	1201.0
401.90	5067.0
403.00	6295.0
404.00	2780.0
405.00	715.0
420.90	5420.0
421.90	4748.0
423.00	45576.0
424.00	10666.0
425.00	1949.0
441.10	97992.0
442.00	641920.0
442.90	116264.0
444.00	10064.0
445.00	616.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

m/z	Abundance
37.00	774.0
38.00	1831.0
39.00	9676.0
39.90	1020.0
41.00	399.0
43.10	336.0
43.90	1868.0
49.10	1510.0
50.00	26904.0
51.00	107352.0
52.00	5243.0
54.90	1026.0
55.90	3980.0
57.00	9029.0
57.90	468.0
60.90	1564.0
62.00	2776.0
63.00	7141.0
64.00	1144.0
65.00	2834.0
67.00	325.0
68.00	3236.0
69.00	194304.0
70.00	839.0
72.90	2899.0
74.00	22152.0
75.00	32384.0
76.00	11337.0
77.00	217280.0
78.00	14572.0
79.00	16776.0
79.90	11724.0
81.00	16424.0
82.00	4108.0
83.00	3893.0
84.10	590.0
85.00	2686.0
85.90	5000.0
86.90	1824.0
87.80	1198.0
90.90	3860.0
92.00	4274.0
93.00	28160.0
94.00	1768.0
95.00	564.0
95.90	1148.0
97.00	755.0
98.00	21408.0
99.00	17904.0
99.80	1421.0
101.00	8148.0
101.90	629.0
102.90	3090.0
103.90	5898.0
105.00	5642.0
105.90	2170.0
107.00	72840.0
108.00	10428.0
110.00	125384.0
111.00	19384.0
111.90	2676.0
113.10	1395.0
115.20	332.0
116.00	3579.0
117.00	64888.0
118.00	5441.0
119.10	1030.0
120.10	996.0
121.00	526.0
122.00	4238.0
123.00	8296.0
123.90	3966.0
124.90	3390.0
127.00	246144.0
128.00	18600.0
129.00	100776.0
130.00	8142.0
131.00	2049.0
132.00	1236.0
132.90	606.0
133.90	2925.0
135.00	9613.0
136.00	3619.0
137.00	4104.0
138.20	976.0
139.20	488.0
139.90	1590.0
141.00	13891.0
142.00	5009.0
143.00	2866.0
143.90	772.0
145.20	1041.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

146.00	2363.0
147.00	7777.0
148.00	17592.0
149.00	4025.0
149.90	1333.0
151.00	2333.0
151.90	1033.0
153.00	4955.0
154.00	3179.0
155.00	8570.0
156.00	11171.0
156.90	1833.0
158.10	2250.0
159.00	2369.0
159.90	4093.0
161.00	6072.0
162.00	1627.0
163.00	691.0
163.90	1021.0
165.00	6171.0
166.00	4989.0
167.00	30984.0
168.00	16808.0
168.90	2799.0
169.90	1329.0
171.00	1393.0
172.00	1900.0
173.00	3515.0
174.00	6295.0
175.10	12532.0
176.00	2649.0
177.00	6432.0
178.10	2424.0
179.00	23296.0
180.00	15607.0
181.00	7666.0
181.80	620.0
182.10	754.0
183.00	1103.0
184.00	1978.0
185.00	11453.0
186.00	90064.0
187.00	23024.0
188.10	2128.0
189.00	5718.0
189.90	1083.0
190.90	2649.0
192.00	7661.0
193.00	10022.0
194.10	2141.0
195.10	1036.0
196.00	18224.0
198.00	601792.0
199.00	37448.0
200.00	3267.0
201.60	2491.0
203.00	5644.0
204.00	23064.0
205.00	41488.0
206.00	175680.0
207.00	25528.0
208.00	7159.0
209.00	2949.0
210.00	2653.0
211.00	7946.0
211.90	1488.0
213.20	529.0
214.90	2157.0
216.00	3589.0
217.00	45776.0
218.00	6426.0
219.10	582.0
221.00	24832.0
221.80	6454.0
223.00	10661.0
224.00	92848.0
225.00	25384.0
226.10	2880.0
227.00	46240.0
228.00	6685.0
229.00	7904.0
229.90	1127.0
231.10	3863.0
232.00	1016.0
233.00	627.0
234.00	2307.0
234.90	3283.0
236.00	1773.0
237.00	5176.0
238.00	792.0
239.00	1786.0
240.10	2342.0
241.00	1783.0
242.00	5035.0
243.00	5639.0
244.00	74920.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

245.00	7466.0
246.00	15678.0
247.00	3758.0
247.80	659.0
248.90	2644.0
249.90	579.0
251.00	935.0
251.90	841.0
253.10	1882.0
255.00	414784.0
256.00	60456.0
257.00	4373.0
258.00	23280.0
259.00	3706.0
260.00	888.0
260.90	741.0
263.00	331.0
264.00	565.0
265.00	9548.0
266.00	1624.0
267.00	421.0
269.90	734.0
270.90	1067.0
272.10	1052.0
273.00	13034.0
274.00	30072.0
275.00	184128.0
276.00	24200.0
277.00	16488.0
277.90	2963.0
279.20	1005.0
281.10	1239.0
282.10	554.0
283.00	2340.0
283.90	1635.0
285.00	3545.0
286.10	486.0
289.00	521.0
290.10	342.0
290.80	379.0
291.90	1345.0
293.00	3027.0
294.00	1206.0
296.00	58024.0
297.00	7230.0
297.90	808.0
301.00	624.0
302.00	1183.0
303.00	6875.0
304.00	2458.0
305.00	374.0
306.80	413.0
307.90	655.0
308.80	600.0
309.90	864.0
312.90	691.0
314.00	3248.0
315.00	6709.0
316.00	3439.0
316.90	501.0
320.90	1844.0
321.90	962.0
323.00	16768.0
324.00	3908.0
326.00	315.0
327.00	3876.0
328.10	1580.0
328.90	383.0
332.00	1704.0
333.00	1880.0
334.00	13040.0
335.00	3820.0
336.10	404.0
340.00	458.0
340.90	2980.0
341.90	817.0
345.90	4405.0
346.90	891.0
350.90	423.0
352.00	5836.0
353.00	4069.0
354.00	6303.0
355.00	1099.0
356.00	344.0
365.00	29896.0
365.90	4358.0
369.90	824.0
370.90	1526.0
372.00	10050.0
373.00	2527.0
374.10	544.0
383.00	3187.0
383.90	1323.0
385.10	373.0
390.00	1624.0
391.10	618.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74

392.00	671.0
401.10	561.0
402.00	4614.0
403.10	5045.0
404.00	2138.0
404.80	526.0
421.00	5680.0
421.90	6380.0
423.00	33096.0
424.10	11218.0
425.10	1174.0
439.30	312.0
441.10	96856.0
442.00	688576.0
443.10	110552.0
444.00	11478.0
445.00	629.0

Instrument :
BNA_N
ClientSampleId :
DFTPP74