





#12
 2,2'-oxybis(1-Chloropropane)
 Concen: 15.626 ng/ul
 RT: 8.12 min Scan# 873
 Delta R.T. 0.00 min
 Lab File: BN005416.D
 Acq: 02 May 2019 15:34

Instrument :
 BNA_N
 ClientSampled :

Tgt Ion: 45 Resp: 452241
 Ion Ratio Lower Upper
 45 100
 77 14.0 9.8 14.6
 79 11.5 7.4 11.0#



#14
 Acetophenone
 Concen: 20.617 ng/ul
 RT: 8.40 min Scan# 921
 Delta R.T. 0.00 min
 Lab File: BN005416.D
 Acq: 02 May 2019 15:34

Tgt Ion: 105 Resp: 539816
 Ion Ratio Lower Upper
 105 100
 77 75.8 63.4 95.0
 51 26.9 24.2 36.4

