

#87

Benzo(b)fluoranthene

Concen: 19.285 ng/ul

RT: 22.80 min Scan# 3369

Delta R.T. 0.04 min

Lab File: BN005416.D

Acq: 02 May 2019 15:34

Instrument :

BNA_N

ClientSampleId :


Tgt Ion:252 Resp: 1898326

Ion Ratio Lower Upper

252 100

253 22.4 17.7 26.5




125 8.4 8.4 12.6



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 20.192 ng/ul

RT: 22.80 min Scan# 3369

Delta R.T. 0.00 min

Lab File: BN005416.D

Acq: 02 May 2019 15:34


Tgt Ion:252 Resp: 1898326

Ion Ratio Lower Upper

252 100

253 22.4 17.4 26.0

125 8.4 7.5 11.3



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E

