





#34
 2-Methylnaphthalene
 Concen: 36.582 ng/ul
 RT: 12.02 min Scan# 1536
 Delta R.T. -0.00 min
 Lab File: BN005430.D
 Acq: 02 May 2019 23:38

Instrument :
 BNA_N
 ClientSampleId :

Tgt Ion:142 Resp: 1429210
 Ion Ratio Lower Upper
 142 100
 141 88.6 72.0 108.0



#40
 1,1'-Biphenyl
 Concen: 1.847 ng/ul
 RT: 13.05 min Scan# 1711
 Delta R.T. -0.00 min
 Lab File: BN005430.D
 Acq: 02 May 2019 23:38

Tgt Ion:154 Resp: 98399
 Ion Ratio Lower Upper
 154 100
 153 38.4 32.6 49.0
 76 10.4 10.4 15.6#

