

#6

Phenol

Concen: 19.879 ng/ul

RT: 6.80 min Scan# 648

Delta R.T. -0.00 min

Lab File: BN005431.D

Acq: 03 May 2019 00:12

Instrument :

BNA_N

ClientSampleId :


Tgt Ion: 94 Resp: 456290

Ion Ratio Lower Upper




94 100

65 29.8 24.2 36.4

66 42.4 33.8 50.6



Abundance Ion 94.00 (93.70 to 94.70): BNC
Ion 65.00 (64.70 to 65.70): BNC
Ion 66.00 (65.70 to 66.70): BNC



#8

Bis(2-Chloroethyl)ether

Concen: 19.064 ng/ul

RT: 7.03 min Scan# 687

Delta R.T. -0.00 min

Lab File: BN005431.D

Acq: 03 May 2019 00:12


Tgt Ion: 93 Resp: 350370

Ion Ratio Lower Upper

93 100

63 74.5 67.1 100.7

95 32.8 26.9 40.3



Abundance Ion 93.00 (92.70 to 93.70): BNC
Ion 63.00 (62.70 to 63.70): BNC
Ion 95.00 (94.70 to 95.70): BNC

