

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN050219\
 Data File : BN005446.D
 Acq On : 03 May 2019 09:23
 Operator : JU/SJ
 Sample : PB119318BL
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK18

Quant Time: May 03 19:09:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN041919MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 03 03:02:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	258681	20.00	ng/ul	0.00
18) Naphthalene-d8	10.35	136	1111160	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.22	164	708954	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.98	188	1592009	20.00	ng/ul	0.00
77) Chrysene-d12	21.20	240	1316345	20.00	ng/ul	-0.01
85) Perylene-d12	23.41	264	1318549	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.22	96	31405	5.86	ng/uL	0.00
5) Phenol-d5	6.78	99	627348	30.88	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.94	67	361709	28.54	ng/ul	0.00
9) 2-Chlorophenol-d4	7.13	132	515182	33.13	ng/ul	0.00
13) 4-Methylphenol-d8	8.29	113	520513	32.77	ng/ul	0.00
19) Nitrobenzene-d5	8.73	128	275680	40.31	ng/ul	0.00
22) 2-Nitrophenol-d4	9.45	143	313683	45.24	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.98	165	541600	34.08	ng/ul	0.00
29) 4-Chloroaniline-d4	10.49	131	756208	46.14	ng/ul	0.00
43) Dimethylphthalate-d6	13.64	166	1960088	38.09	ng/ul	0.00
46) Acenaphthylene-d8	13.92	160	2294352	36.30	ng/ul	0.00
51) 4-Nitrophenol-d4	14.44	143	284378	35.00	ng/ul	0.00
57) Fluorene-d10	15.22	176	1632587	37.80	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.36	200	254751	35.58	ng/ul	0.00
70) Anthracene-d10	17.08	188	2603266	39.01	ng/ul	0.00
78) Pyrene-d10	19.39	212	2769218	41.12	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.27	264	2334736	40.52	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN050219\
Data File : BN005446.D
Acq On : 03 May 2019 09:23
Operator : JU/SJ
Sample : PB119318BL
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK18

Quant Time: May 03 19:09:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN041919MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 03 03:02:31 2019
Response via : Initial Calibration

