

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN050319\
 Data File : BN005465.D
 Acq On : 03 May 2019 22:47
 Operator : JU/SJ
 Sample : K2602-11
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BFH70

Quant Time: May 04 00:58:49 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN041919MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 03 03:02:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	254777	20.00	ng/ul	0.00
18) Naphthalene-d8	10.35	136	1159664	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.22	164	755316	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.97	188	1768923	20.00	ng/ul	0.00
77) Chrysene-d12	21.20	240	1575822	20.00	ng/ul	-0.01
85) Perylene-d12	23.40	264	1452692	20.00	ng/ul	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.22	96	16838	3.19	ng/uL	0.00
5) Phenol-d5	6.77	99	382501	19.12	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.94	67	243579	19.51	ng/ul	0.00
9) 2-Chlorophenol-d4	7.13	132	325761	21.27	ng/ul	0.00
13) 4-Methylphenol-d8	8.29	113	327535	20.94	ng/ul	0.00
19) Nitrobenzene-d5	8.73	128	182427	25.56	ng/ul	0.00
22) 2-Nitrophenol-d4	9.45	143	149860	20.71	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.98	165	324179	19.54	ng/ul	0.00
29) 4-Chloroaniline-d4	10.49	131	371827	21.74	ng/ul	0.00
43) Dimethylphthalate-d6	13.64	166	1109202	20.23	ng/ul	0.00
46) Acenaphthylene-d8	13.91	160	1519236	22.56	ng/ul	0.00
51) 4-Nitrophenol-d4	14.46	143	50631	5.85	ng/ul	0.02
57) Fluorene-d10	15.22	176	1062953	23.10	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.36	200	12114	1.52	ng/ul	0.00
70) Anthracene-d10	17.07	188	1665072	22.45	ng/ul	0.00
78) Pyrene-d10	19.39	212	1880966	23.33	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.26	264	1454313	22.91	ng/ul	-0.02


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
44) Dimethylphthalate	13.69	163	455267	8.333	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN050319\
Data File : BN005465.D
Acq On : 03 May 2019 22:47
Operator : JU/SJ
Sample : K2602-11
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BFH70

Quant Time: May 04 00:58:49 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN041919MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 03 03:02:31 2019
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 8.333 ng/ul

RT: 13.69 min Scan# 1819

Delta R.T. -0.01 min

Lab File: BN005465.D

Acq: 03 May 2019 22:47

Instrument :

BNA_N

ClientSampleId :

BFH70

Tgt Ion:163 Resp: 455267

Ion Ratio Lower Upper

163 100

194 3.9 3.4 5.0

164 10.1 8.1 12.1

