



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN050324\
Data File : BN031383.D
Acq On : 03 May 2024 09:29
Operator : MA/JU
Sample : DFTPP327
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP327

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN050324.MA.M
Title : SVOA CALIBRATION
Last Update : Fri May 03 14:17:04 2024



AutoFind: Scans 2457, 2458, 2459; Background Corrected with Scan 2451

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	32.2	91048	PASS
68	69	0.00	2	1.8	1842	PASS
69	198	0.00	100	35.8	101309	PASS
70	69	0.00	2	0.6	585	PASS
127	198	10	80	46.1	130360	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	283051	PASS
199	198	5	9	6.7	19003	PASS
275	198	10	60	29.5	83467	PASS
365	198	1	100	4.7	13273	PASS
441	198	0.01	100	15.2	43093	PASS
442	198	50	100	93.3	264192	PASS
443	442	15	24	19.3	50904	PASS

m/z	Abundance
39.90	381.0
41.00	237.0
44.00	2031.0
44.90	168.0
54.90	166.0
72.90	569.0
77.00	187.0
190.80	201.0
207.00	1114.0
208.00	254.0
209.10	207.0
221.10	193.0
281.00	296.0

Instrument :

BNA_N

ClientSampleId :

DFTPP327

m/z	Abundance
37.00	770.0
38.00	1860.0
39.00	8786.0
40.00	806.0
41.00	375.0
43.00	267.0
43.90	1440.0
44.90	204.0
49.00	1055.0
50.00	32376.0
51.00	115264.0
52.00	6212.0
52.90	261.0
55.00	871.0
56.00	3628.0
57.00	7451.0
58.00	351.0
58.90	157.0
61.00	1731.0
62.10	1977.0
63.00	5625.0
64.00	840.0
65.00	2152.0
66.10	271.0
67.00	339.0
68.00	2736.0
69.00	127808.0
70.00	746.0
73.00	1388.0
74.00	14842.0
75.00	21736.0
76.00	7141.0
77.00	135424.0
78.00	9295.0
79.00	11093.0
80.00	8046.0
81.00	11362.0
82.00	2772.0
83.00	2150.0
83.90	258.0
85.00	1826.0
85.90	3356.0
87.00	1397.0
88.10	583.0
89.00	221.0
91.00	2473.0
92.00	2667.0
93.00	20408.0
94.00	1185.0
95.00	251.0
96.00	659.0
96.90	508.0
98.00	15715.0
99.00	11460.0
100.00	850.0
101.00	6326.0
101.90	385.0
102.90	2325.0
104.00	3874.0
105.00	4002.0
106.10	1198.0
107.00	48928.0
108.00	7230.0
109.10	1495.0
110.00	82848.0
111.00	12816.0
111.90	1529.0
112.90	493.0
115.00	395.0
116.00	3035.0
117.00	45808.0
117.90	3474.0
118.90	439.0
120.00	627.0
121.90	3688.0
123.00	5376.0
124.00	2151.0
124.90	1963.0
127.00	148544.0
128.00	12958.0
129.00	66048.0
129.90	5977.0
130.90	1165.0
132.00	637.0
133.10	500.0
133.90	2194.0
135.00	5579.0
136.00	2104.0
137.00	2874.0
139.00	275.0
140.20	923.0
141.00	9173.0

Instrument :
BNA_N
ClientSampleId :
DFTPP327

142.00	3023.0
143.00	1998.0
144.00	519.0
145.00	446.0
145.90	1935.0
147.00	4991.0
148.00	11487.0
149.00	2103.0
150.00	854.0
151.00	1471.0
151.80	711.0
153.00	2898.0
154.00	2103.0
155.00	4851.0
156.00	7053.0
157.00	1275.0
157.90	1741.0
158.90	1445.0
159.90	3179.0
161.00	3851.0
161.90	1162.0
163.10	387.0
164.10	374.0
164.90	3475.0
166.00	2230.0
167.00	18400.0
168.00	9922.0
169.00	1723.0
170.10	768.0
171.00	734.0
171.90	1508.0
172.90	2155.0
174.00	4011.0
174.90	7170.0
176.00	1709.0
177.00	3570.0
177.90	1142.0
179.00	14272.0
180.00	8650.0
181.00	3965.0
181.90	622.0
182.90	400.0
184.00	1166.0
185.00	6523.0
186.00	47976.0
187.00	14527.0
188.00	1543.0
188.90	3598.0
190.00	557.0
190.90	1488.0
192.00	4248.0
192.90	4752.0
194.00	1121.0
196.00	8558.0
197.90	311104.0
198.90	22592.0
199.90	1729.0
201.40	1256.0
203.00	2769.0
204.00	12605.0
205.00	21928.0
206.00	89776.0
207.00	13230.0
208.00	3144.0
209.10	1395.0
210.00	1581.0
210.90	3428.0
212.00	664.0
212.90	343.0
215.00	1308.0
215.90	1715.0
216.90	23584.0
217.90	3640.0
219.00	207.0
221.00	11274.0
221.90	2796.0
223.00	4796.0
224.00	48280.0
225.00	12156.0
227.00	23616.0
227.90	3765.0
228.90	4091.0
229.90	520.0
231.00	1606.0
231.90	330.0
232.90	389.0
233.90	1318.0
235.00	1522.0
236.00	1016.0
236.90	1912.0
238.10	271.0
238.90	903.0
239.80	595.0
240.90	1197.0
242.00	2494.0
243.00	2843.0

Instrument :
BNA_N
ClientSampleId :
DFTPP327

244.00	34352.0
245.00	4628.0
245.90	9029.0
246.90	1601.0
247.90	298.0
248.90	1582.0
250.00	246.0
251.00	332.0
252.10	318.0
253.00	946.0
255.00	196544.0
256.00	30152.0
257.00	2496.0
257.90	11759.0
258.90	1881.0
259.90	280.0
261.00	295.0
262.90	236.0
263.80	316.0
264.90	5057.0
266.00	730.0
267.00	192.0
269.70	201.0
270.00	183.0
270.90	601.0
272.20	663.0
273.00	5648.0
274.00	15912.0
275.00	86224.0
276.00	11752.0
276.90	8650.0
277.90	1551.0
278.90	431.0
280.80	309.0
281.10	258.0
282.10	258.0
283.00	1011.0
284.00	611.0
284.90	1445.0
285.90	302.0
288.90	410.0
289.90	257.0
290.90	215.0
292.10	518.0
293.00	1700.0
293.90	387.0
294.90	674.0
295.90	26944.0
296.90	3886.0
297.90	282.0
300.90	281.0
302.00	436.0
303.00	3313.0
304.00	996.0
305.00	211.0
307.80	381.0
309.10	230.0
310.10	381.0
312.80	204.0
314.00	1149.0
314.90	2960.0
316.00	1493.0
316.90	285.0
320.90	1042.0
322.10	370.0
323.00	8483.0
324.00	1538.0
325.00	221.0
326.00	208.0
326.90	1705.0
327.90	734.0
328.90	224.0
332.10	556.0
332.90	956.0
334.00	5914.0
335.10	1378.0
335.80	209.0
341.00	883.0
341.90	300.0
345.90	1729.0
346.90	294.0
351.00	206.0
351.90	2734.0
352.90	1719.0
354.00	2914.0
354.90	498.0
364.90	13115.0
365.90	1989.0
369.70	291.0
370.90	601.0
372.00	4768.0
373.00	1107.0
373.90	190.0
382.90	1112.0
384.00	349.0
389.70	315.0

Instrument :
BNA_N
ClientSampleId :
DFTPP327

390.10	410.0
390.90	603.0
391.90	282.0
401.00	232.0
401.90	2344.0
403.00	2548.0
403.90	823.0
404.80	262.0
414.90	162.0
420.90	2166.0
422.00	2439.0
422.90	15566.0
424.00	3730.0
425.00	418.0
441.00	37560.0
442.00	218496.0
443.00	48560.0
443.90	4628.0
444.90	279.0

Instrument :
BNA_N
ClientSampleId :
DFTPP327

m/z	Abundance
37.00	589.0
38.00	1605.0
39.00	7897.0
40.00	875.0
41.00	344.0
43.20	213.0
44.00	1676.0
45.00	346.0
49.00	1119.0
50.00	28568.0
51.00	104072.0
52.00	5286.0
53.00	340.0
55.00	765.0
56.00	3478.0
57.00	6955.0
57.90	263.0
60.00	252.0
61.00	1774.0
62.00	1655.0
63.00	4771.0
64.00	645.0
65.00	1843.0
66.00	239.0
66.90	224.0
68.10	2791.0
69.00	117120.0
69.90	724.0
73.00	1681.0
74.00	12918.0
75.00	20112.0
76.00	6848.0
77.00	128192.0
78.00	8238.0
79.00	11221.0
80.00	7440.0
81.00	10084.0
82.00	2767.0
83.00	2075.0
83.90	403.0
85.00	1886.0
85.90	3491.0
86.90	1411.0
88.00	570.0
88.90	301.0
90.90	2495.0
92.00	2757.0
93.00	18312.0
94.00	1520.0
94.90	353.0
96.00	832.0
98.00	15074.0
99.00	10881.0
99.90	914.0
101.00	4939.0
102.00	437.0
103.00	2438.0
104.00	4296.0
105.00	3587.0
106.10	1211.0
107.00	46656.0
108.00	7139.0
110.00	81848.0
111.00	11529.0
111.90	1677.0
113.00	569.0
114.90	291.0
115.90	3167.0
117.00	45560.0
118.00	3843.0
119.00	307.0
119.90	676.0
121.20	250.0
122.00	3759.0
123.00	5010.0
123.90	2324.0
125.00	2094.0
127.00	156032.0
128.00	12426.0
129.00	66192.0
130.00	5769.0
130.90	1306.0
131.80	569.0
133.00	419.0
133.90	2369.0
135.00	5750.0
136.00	2149.0
137.00	3005.0
138.00	658.0
138.90	473.0
139.90	868.0
141.00	9142.0

Instrument :
BNA_N
ClientSampleId :
DFTPP327

142.00	3063.0
142.90	1796.0
144.00	713.0
144.90	416.0
146.00	1806.0
147.00	4494.0
147.90	10376.0
149.00	2072.0
149.90	538.0
150.90	940.0
151.80	595.0
152.90	3267.0
154.00	2490.0
155.00	5547.0
156.00	8512.0
157.00	1324.0
157.90	1815.0
158.90	1170.0
159.90	3129.0
161.00	4217.0
161.90	1301.0
162.90	352.0
163.90	274.0
164.90	3716.0
165.90	2867.0
167.00	18648.0
168.00	9100.0
169.00	1334.0
170.00	737.0
170.90	1052.0
171.90	1627.0
173.00	1929.0
174.00	4000.0
175.00	6833.0
175.90	1987.0
177.00	3492.0
178.00	1060.0
178.90	13797.0
180.00	9105.0
181.00	4879.0
182.00	843.0
183.00	402.0
184.00	1357.0
185.00	6169.0
186.00	49840.0
187.00	14871.0
187.90	1761.0
189.00	3906.0
189.90	700.0
191.00	1898.0
192.00	4394.0
193.00	5726.0
194.00	1243.0
194.90	709.0
196.00	9357.0
197.90	336064.0
198.90	22192.0
199.90	1841.0
201.20	1072.0
201.70	1160.0
203.00	3060.0
204.00	13303.0
205.00	23664.0
206.00	98336.0
207.00	14436.0
208.00	3559.0
209.00	1227.0
210.20	1694.0
211.00	3894.0
211.80	740.0
212.90	309.0
214.90	1538.0
216.00	2219.0
216.90	25968.0
218.00	3037.0
218.90	320.0
221.00	10002.0
223.00	5924.0
224.00	54456.0
225.00	12610.0
226.00	1675.0
227.00	25632.0
227.90	3514.0
229.00	5206.0
230.00	597.0
231.00	1767.0
231.90	310.0
232.90	386.0
233.90	1567.0
234.90	1887.0
236.00	1085.0
237.00	1968.0
237.80	309.0
238.80	1035.0
239.90	1014.0
241.00	1414.0

Instrument :
BNA_N
ClientSampleId :
DFTPP327

242.00	2795.0
243.00	3391.0
244.00	40720.0
245.00	5799.0
246.00	9181.0
246.80	1897.0
247.90	486.0
249.00	1323.0
249.80	327.0
250.90	433.0
251.90	292.0
253.10	1010.0
255.00	231936.0
256.00	31456.0
257.00	2507.0
258.00	13999.0
259.00	2045.0
260.00	450.0
260.90	388.0
265.00	6398.0
266.00	1105.0
266.90	180.0
270.00	305.0
270.90	815.0
271.90	565.0
272.90	6049.0
274.00	18928.0
275.00	101112.0
275.90	12573.0
276.90	9648.0
278.00	1735.0
278.90	469.0
281.00	514.0
281.80	334.0
282.90	1107.0
283.90	775.0
285.00	1395.0
285.80	357.0
288.80	328.0
290.00	252.0
291.10	260.0
292.00	503.0
292.90	1699.0
293.80	596.0
296.00	33120.0
297.00	4217.0
298.10	284.0
300.90	300.0
302.00	570.0
303.00	3836.0
304.00	1134.0
307.90	511.0
308.70	273.0
309.80	368.0
310.90	176.0
312.90	235.0
313.90	1371.0
315.00	3702.0
315.90	1835.0
316.90	429.0
321.00	988.0
321.90	535.0
323.00	10722.0
324.00	1791.0
326.00	297.0
326.80	2181.0
328.00	1107.0
328.90	206.0
332.00	963.0
332.90	1075.0
334.00	6918.0
335.00	1898.0
336.00	230.0
339.00	337.0
339.90	202.0
341.00	1205.0
341.80	387.0
345.90	2302.0
346.80	446.0
350.90	212.0
352.00	3648.0
352.90	2126.0
353.90	3897.0
355.00	666.0
359.00	285.0
364.90	15524.0
366.00	1868.0
366.90	182.0
369.90	289.0
371.00	1054.0
372.00	5376.0
373.00	1636.0
373.90	220.0
383.00	1459.0
384.10	425.0
389.90	658.0

Instrument :
BNA_N
ClientSampleId :
DFTPP327

391.00	586.0
391.90	395.0
400.90	330.0
401.80	2159.0
402.90	3946.0
403.90	1342.0
404.90	317.0
420.90	3109.0
421.90	3475.0
423.00	21552.0
423.90	5106.0
425.00	508.0
438.90	164.0
439.70	362.0
441.00	52832.0
442.00	319360.0
442.90	58512.0
444.00	5711.0
444.90	298.0

Instrument :
BNA_N
ClientSampleId :
DFTPP327

m/z	Abundance
36.90	174.0
38.10	802.0
39.10	3896.0
40.00	678.0
41.00	320.0
43.00	411.0
44.00	1890.0
44.90	317.0
50.00	14142.0
51.00	53808.0
52.00	2249.0
54.90	684.0
56.00	1708.0
57.00	3515.0
61.00	950.0
62.00	787.0
63.00	2648.0
63.90	269.0
64.90	1021.0
66.90	154.0
69.00	59000.0
70.00	284.0
72.90	1003.0
74.00	7057.0
75.00	10672.0
76.10	3642.0
77.00	65576.0
78.00	4416.0
79.00	5432.0
79.90	3928.0
81.00	5922.0
82.00	1468.0
83.00	862.0
83.90	202.0
84.90	941.0
85.90	1442.0
86.90	729.0
87.90	245.0
88.90	185.0
90.90	1446.0
91.90	1567.0
93.00	10187.0
93.90	628.0
94.90	295.0
95.90	586.0
98.00	8058.0
99.00	6225.0
100.00	520.0
101.00	3395.0
102.00	221.0
102.90	1010.0
103.90	2140.0
105.00	1906.0
105.90	702.0
107.00	23632.0
108.00	3637.0
109.10	841.0
110.00	42560.0
110.90	6765.0
112.00	856.0
112.80	345.0
115.10	208.0
116.00	1536.0
117.00	24368.0
117.90	1732.0
118.90	262.0
119.80	362.0
121.90	1940.0
122.90	2718.0
123.90	1288.0
124.90	1151.0
127.00	86504.0
128.00	6574.0
129.00	35912.0
129.90	3157.0
130.90	859.0
132.00	282.0
132.90	338.0
134.00	1163.0
134.90	2899.0
135.90	1227.0
136.90	1592.0
137.80	272.0
138.80	222.0
139.90	574.0
141.00	4596.0
141.90	1285.0
143.00	1009.0
143.90	365.0
144.90	253.0
145.90	820.0
147.00	2702.0

Instrument :
BNA_N
ClientSampleId :
DFTPP327

148.00	6177.0
148.90	1251.0
149.80	459.0
151.10	586.0
151.80	332.0
153.00	1701.0
154.00	1290.0
155.00	2546.0
156.00	4561.0
157.00	696.0
158.00	863.0
158.90	771.0
159.90	1887.0
160.90	2368.0
161.90	507.0
162.90	201.0
164.10	275.0
164.90	2099.0
166.00	1627.0
167.00	11045.0
168.00	5424.0
169.00	787.0
170.00	386.0
171.00	585.0
171.80	904.0
173.00	1127.0
174.00	2163.0
175.00	4018.0
176.00	1130.0
176.90	2047.0
177.90	746.0
179.00	7662.0
180.00	5876.0
180.90	2048.0
182.00	584.0
183.10	218.0
184.00	720.0
184.90	3770.0
186.00	28872.0
187.00	9103.0
188.10	649.0
189.00	2034.0
189.90	316.0
191.00	1135.0
192.00	2763.0
193.00	2804.0
194.00	620.0
194.90	374.0
196.00	5881.0
197.90	201984.0
199.00	12224.0
199.90	996.0
201.50	614.0
203.00	1653.0
204.00	8355.0
205.00	13635.0
206.00	57536.0
207.00	7813.0
208.00	2304.0
208.90	901.0
210.10	1098.0
211.10	2330.0
211.60	538.0
212.90	200.0
214.90	601.0
216.00	1265.0
216.90	15738.0
218.00	1942.0
219.00	254.0
221.00	6360.0
221.90	2300.0
223.00	3644.0
224.00	33216.0
225.00	8011.0
226.90	15033.0
227.90	1884.0
229.00	2732.0
229.90	499.0
230.90	892.0
232.80	341.0
234.00	892.0
235.00	1058.0
235.80	652.0
236.90	1230.0
237.80	256.0
238.90	515.0
239.90	472.0
240.90	763.0
242.00	1773.0
243.00	1990.0
244.00	23560.0
245.00	2723.0
245.90	6404.0
246.90	1074.0
247.80	288.0
248.90	1156.0

Instrument :
BNA_N
ClientSampleId :
DFTPP327

249.90	277.0
250.90	323.0
251.90	285.0
253.00	476.0
255.00	134016.0
256.00	20688.0
257.00	1449.0
258.00	8783.0
259.00	1401.0
260.00	227.0
261.00	234.0
264.00	226.0
264.90	3810.0
266.00	579.0
266.80	260.0
269.80	201.0
271.00	303.0
271.90	384.0
272.90	4404.0
274.00	12089.0
275.00	63064.0
276.00	8129.0
276.90	6003.0
277.90	878.0
278.80	225.0
280.90	479.0
281.80	280.0
282.90	837.0
284.00	629.0
284.90	882.0
286.00	231.0
288.70	299.0
289.90	209.0
290.90	186.0
291.90	236.0
292.90	1274.0
293.90	348.0
296.00	20560.0
296.90	2807.0
298.00	240.0
301.00	442.0
301.90	524.0
303.00	2299.0
303.90	630.0
307.90	288.0
308.90	220.0
309.90	236.0
312.90	172.0
314.00	1023.0
315.00	2396.0
315.90	1190.0
320.80	610.0
322.10	353.0
323.00	6509.0
323.90	1252.0
324.90	207.0
325.90	174.0
327.00	1478.0
328.00	646.0
331.90	574.0
333.00	619.0
334.00	4537.0
334.90	1009.0
336.00	156.0
341.00	868.0
345.90	1470.0
346.80	236.0
352.00	2264.0
353.00	1459.0
354.00	2544.0
354.90	529.0
359.10	158.0
364.90	11179.0
365.90	1533.0
369.90	175.0
370.90	748.0
372.00	4117.0
373.00	1219.0
382.90	1210.0
383.90	418.0
389.90	475.0
390.90	283.0
392.00	369.0
400.90	446.0
402.00	2010.0
402.90	2391.0
403.90	864.0
404.90	168.0
415.00	159.0
421.00	2246.0
421.90	2354.0
423.00	14601.0
423.90	3454.0
424.90	321.0
440.90	38888.0
442.00	254720.0

Instrument :
BNA_N
ClientSampleId :
DFTPP327

442.9045640.0

444.004234.0

445.00203.0

Instrument :
BNA_N
ClientSampleId :
DFTPP327