

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN050719\
 Data File : BN005507.D
 Acq On : 07 May 2019 03:57
 Operator : JU/SJ
 Sample : K2403-08
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A5125

Manual Integrations
 APPROVED

mohammad
 5/8/2019 12:01:15 AM

Quant Time: May 07 07:57:28 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN050719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 07 07:04:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	2186	0.40	ng/ul	-0.02
2) Naphthalene-d8	10.35	136	8826	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.22	164	5699	0.40	ng/ul	-0.02
10) Phenanthrene-d10	16.98	188	13633m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.21	240	13610m	0.40	ng/ul	-0.01
20) Perylene-d12	23.41	264	11899m	0.40	ng/ul	-0.03
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.96	152	3886	0.28	ng/ul	-0.03
14) Fluoranthene-d10	19.02	212	11826m	0.28	ng/ul	-0.01
Target Compounds					Ovalue	
3) Naphthalene	10.40	128	2299	0.098	ng/ul	96
5) 2-Methylnaphthalene	12.04	142	1831	0.115	ng/ul	99
7) Acenaphthylene	13.94	152	2886	0.101	ng/ul	98
9) Fluorene	15.28	166	610	0.026	ng/ul#	94
12) Phenanthrene	17.02	178	9675	0.260	ng/ul	100
13) Anthracene	17.11	178	1481m	0.040	ng/ul	
15) Fluoranthene	19.06	202	11698m	0.228	ng/ul	
17) Pyrene	19.42	202	13698m	0.273	ng/ul	
18) Benzo(a)anthracene	21.20	228	6214m	0.132	ng/ul	
19) Chrysene	21.25	228	9193	0.160	ng/ul	95
21) Benzo(b)fluoranthene	22.77	252	9654m	0.270	ng/ul	
22) Benzo(k)fluoranthene	22.80	252	3256m	0.073	ng/ul	
23) Benzo(a)pyrene	23.32	252	6061m	0.152	ng/ul	
24) Indeno(1,2,3-cd)pyrene	25.61	276	3831m	0.086	ng/ul	
25) Dibenzo(a,h)anthracene	25.62	278	943m	0.026	ng/ul	
26) Benzo(a,h,i)perylene	26.27	276	2814m	0.074	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN050719\
Data File : BN005507.D
Acq On : 07 May 2019 03:57
Operator : JU/SJ
Sample : K2403-08
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A5125

Manual Integrations
APPROVED

mohammad
5/8/2019 12:01:15 AM

Quant Time: May 07 07:57:28 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN050719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 07 07:04:57 2019
Response via : Initial Calibration

