

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN050719\
 Data File : BN005515.D
 Acq On : 07 May 2019 13:19
 Operator : JU/SJ
 Sample : K2403-06
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A5117

Quant Time: May 07 13:52:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN050719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 07 07:04:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	1614	0.40	ng/ul	-0.01
2) Naphthalene-d8	10.36	136	6211	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.23	164	3839	0.40	ng/ul	-0.02
10) Phenanthrene-d10	16.99	188	10345	0.40	ng/ul	0.00
16) Chrysene-d12	21.24	240	10488	0.40	ng/ul	0.02
20) Perylene-d12	23.48	264	11240	0.40	ng/ul	0.04
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.97	152	3160	0.32	ng/ul	-0.02
14) Fluoranthene-d10	19.03	212	11912	0.37	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	10.41	128	10555	0.642	ng/ul	99
5) 2-Methylnaphthalene	12.04	142	7308	0.651	ng/ul	99
9) Fluorene	15.30	166	11913	0.746	ng/ul	100
11) Pentachlorophenol	16.66	266	5256	1.961	ng/ul	93
12) Phenanthrene	17.04	178	24651	0.875	ng/ul	100
15) Fluoranthene	19.07	202	26881	0.689	ng/ul	98
18) Benzo(a)anthracene	21.23	228	34280	0.948	ng/ul	100
22) Benzo(k)fluoranthene	22.86	252	31878	0.753	ng/ul	98
24) Indeno(1,2,3-cd)pyrene	25.67	276	24191	0.573	ng/ul#	89
26) Benzo(g,h,i)perylene	26.33	276	39543	1.098	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN050719\
Data File : BN005515.D
Acq On : 07 May 2019 13:19
Operator : JU/SJ
Sample : K2403-06
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A5117

Quant Time: May 07 13:52:58 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN050719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 07 07:04:57 2019
Response via : Initial Calibration

