


Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN050824\  
Data File : BN031431.D  
Acq On : 08 May 2024 08:09  
Operator : MA/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN050724.M  
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
Last Update : Tue May 07 17:14:43 2024






AutoFind: Scans 2452, 2453, 2454; Background Corrected with Scan 2445

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 34.5         | 149037     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.9          | 2965       | PASS                |
| 69             | 198             | 0.00            | 100             | 36.1         | 155909     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 837        | PASS                |
| 127            | 198             | 10              | 80              | 47.2         | 204131     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.0          | 0          | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 432320     | PASS                |
| 199            | 198             | 5               | 9               | 6.4          | 27824      | PASS                |
| 275            | 198             | 10              | 60              | 26.3         | 113597     | PASS                |
| 365            | 198             | 1               | 100             | 3.9          | 16808      | PASS                |
| 441            | 198             | 0.01            | 100             | 11.8         | 51171      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 304981     | PASS                |
| 443            | 442             | 15              | 24              | 19.8         | 60248      | PASS                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN050824\  
Data File : BN031431.D  
Acq On : 08 May 2024 08:09  
Operator : MA/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: May 09 02:19:32 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu May 09 02:19:28 2024  
Response via : Initial Calibration



TIC: BN031431.D\data.ms

**(70) Pentachlorophenol (C)**

**16.616min ( 0.000) 20844.94 ng**




**response 555701**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 265.70 | 100.00 | 100.00 |
| 268.00 | 62.20  | 61.84  |
| 264.00 | 61.60  | 59.64  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN050824\  
Data File : BN031431.D  
Acq On : 08 May 2024 08:09  
Operator : MA/JU  
Sample : DFTPP  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP

Quant Time: May 09 02:19:32 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270E-Tune.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu May 09 02:19:28 2024  
Response via : Initial Calibration



TIC: BN031431.D\data.ms

(77) Benzidine

19.239min ( 0.000) 0.00 ng

response 2732751

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 184.00 | 100.00 | 100.00 |
| 185.00 | 15.50  | 13.66  |
| 183.00 | 13.20  | 11.81  |
| 0.00   | 0.00   | 0.00   |

### DDT Breakdown

| Date          | Instrument Name  | DFTPP Data File    |
|---------------|------------------|--------------------|
| 5/8/2024      | BNA_N            | <u>BN031431.D</u>  |
|               |                  |                    |
|               |                  |                    |
| Compound Name | Response         | Retention Time     |
|               |                  |                    |
| DDT           | 1669241          | 20.468             |
| DDD           | 19378            | 20.074             |
| DDE           | 384              | 19.516             |
|               |                  |                    |
|               |                  |                    |
| SUM(DDD+DDE)  | SUM(DDT+DDD+DDE) | % Breakdown Of DDT |
|               |                  |                    |
| 19762         | 1689003          | 1.17               |
|               |                  |                    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP