


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051019\
 Data File : BN005585.D
 Acq On : 10 May 2019 10:21
 Operator : JU/SJ
 Sample : DFTPP66
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP66

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN041919MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue May 07 05:36:07 2019



AutoFind: Scans 2455, 2456, 2457; Background Corrected with Scan 2450

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.9	62314	PASS
68	69	0.00	2	1.5	983	PASS
69	198	0.00	100	32.6	63646	PASS
70	69	0.00	2	0.8	504	PASS
127	198	10	80	41.7	81450	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	195306	PASS
199	198	5	9	7.0	13667	PASS
275	198	10	60	27.1	52877	PASS
365	198	1	100	3.4	6625	PASS
441	198	0.01	100	12.5	24463	PASS
442	198	50	100	79.2	154778	PASS
443	442	15	24	18.6	28866	PASS

m/z	Abundance
39.90	1182.0
41.00	504.0
43.00	449.0
44.00	4443.0
46.90	1245.0
48.00	431.0
49.00	5887.0
49.90	477.0
50.90	2315.0
54.90	628.0
56.90	448.0
67.00	329.0
69.10	338.0
73.10	343.0
82.80	393.0
83.90	4571.0
85.90	3170.0
87.90	668.0
206.90	1798.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

m/z	Abundance
35.80	313.0
38.10	545.0
39.00	3646.0
39.90	1280.0
41.10	646.0
43.00	455.0
44.00	4535.0
47.00	1173.0
47.90	680.0
48.90	6255.0
50.00	14200.0
51.00	55280.0
52.00	2940.0
53.00	308.0
54.90	661.0
56.00	1683.0
57.00	3915.0
61.00	720.0
62.00	580.0
63.00	2070.0
65.00	1018.0
67.00	353.0
67.90	1073.0
68.90	51536.0
70.10	451.0
73.00	860.0
74.00	5223.0
75.00	9051.0
76.10	3109.0
77.00	55688.0
78.00	4331.0
79.00	4364.0
79.90	3371.0
81.00	4546.0
81.90	1322.0
82.90	1294.0
83.90	5546.0
84.90	1270.0
86.00	4285.0
87.00	797.0
87.90	780.0
91.00	1307.0
91.90	782.0
93.00	6141.0
94.00	314.0
95.00	305.0
96.00	791.0
98.00	5343.0
98.90	4339.0
101.00	2365.0
102.90	846.0
103.90	1933.0
104.90	1294.0
106.00	609.0
107.00	19832.0
108.10	2464.0
109.00	917.0
110.00	31504.0
111.00	5486.0
112.00	686.0
116.00	1331.0
117.00	12555.0
117.90	643.0
119.00	442.0
121.90	959.0
122.90	1695.0
123.90	655.0
124.90	920.0
127.00	61752.0
128.00	4683.0
129.00	24792.0
130.00	2157.0
131.00	438.0
132.10	353.0
133.10	367.0
134.00	544.0
135.00	1808.0
135.90	611.0
136.90	942.0
137.90	302.0
139.70	342.0
140.90	3227.0
142.10	1008.0
142.80	783.0
144.10	330.0
146.00	657.0
147.00	2185.0
148.00	3609.0
149.00	739.0
151.00	344.0
152.00	336.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

153.00	1002.0
153.90	889.0
155.00	1830.0
155.90	2934.0
157.00	663.0
157.90	704.0
159.00	777.0
160.10	898.0
161.00	1349.0
162.00	420.0
164.90	1355.0
165.90	1120.0
167.00	6873.0
168.00	3013.0
168.80	610.0
170.80	321.0
171.90	715.0
172.90	799.0
174.00	1278.0
175.10	2628.0
175.90	772.0
176.90	947.0
178.10	578.0
178.90	5052.0
180.00	3662.0
181.00	1360.0
182.10	365.0
184.00	374.0
184.90	2448.0
186.00	17152.0
187.00	5049.0
188.00	735.0
188.90	993.0
190.90	690.0
191.90	1734.0
193.00	1892.0
193.90	431.0
195.90	4127.0
197.90	135296.0
198.90	10196.0
200.00	665.0
201.40	797.0
202.90	921.0
204.00	4493.0
205.00	7946.0
206.00	34032.0
207.00	6661.0
208.00	1648.0
209.00	696.0
210.10	739.0
210.90	1444.0
215.00	625.0
215.90	897.0
217.00	10530.0
217.90	1081.0
221.00	6024.0
221.70	1505.0
222.90	2213.0
224.00	20320.0
225.00	4776.0
225.90	671.0
227.00	7864.0
227.80	937.0
229.10	1868.0
231.10	771.0
233.80	397.0
235.00	448.0
235.90	375.0
236.90	750.0
240.80	429.0
242.00	1269.0
242.90	1200.0
244.00	14901.0
245.10	1494.0
245.90	3069.0
247.00	516.0
249.10	644.0
252.90	370.0
255.00	73848.0
256.00	10683.0
257.00	722.0
258.00	3838.0
258.80	1133.0
265.00	1677.0
265.90	385.0
273.00	2118.0
274.00	5723.0
275.00	33256.0
276.00	4521.0
276.90	2087.0
283.00	407.0
284.80	367.0
292.90	557.0
296.00	7699.0
297.10	1440.0

Instrument :

BNA_N

ClientSampleId :

DFTPP66

302.90	943.0
303.90	518.0
308.00	317.0
314.00	399.0
315.00	1057.0
316.00	686.0
323.00	3177.0
324.10	703.0
327.00	531.0
333.00	381.0
334.10	1726.0
335.10	518.0
341.10	306.0
346.00	336.0
352.10	1102.0
352.90	713.0
354.00	938.0
354.90	371.0
364.90	3667.0
365.90	517.0
372.00	1064.0
373.00	367.0
383.00	466.0
401.90	752.0
402.90	691.0
421.00	708.0
422.10	593.0
423.00	4688.0
424.00	1230.0
441.00	13261.0
442.00	80208.0
443.00	16936.0
444.00	1697.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

m/z	Abundance
37.00	470.0
38.00	1140.0
39.00	5231.0
39.90	1381.0
41.10	660.0
41.90	359.0
42.90	551.0
44.00	4330.0
47.00	1380.0
48.00	588.0
49.00	7282.0
50.00	19128.0
51.00	75648.0
52.00	3543.0
54.90	777.0
55.90	2406.0
57.00	5664.0
61.00	1174.0
62.00	909.0
62.90	3016.0
64.00	485.0
64.90	1514.0
68.10	1439.0
69.00	73816.0
69.90	477.0
73.00	841.0
74.00	6807.0
75.00	11482.0
76.10	3833.0
77.00	82096.0
78.00	5979.0
79.00	5999.0
80.00	4307.0
80.90	6180.0
82.00	1757.0
82.90	1723.0
83.90	5493.0
84.90	1691.0
86.00	5239.0
86.90	1102.0
87.80	756.0
91.10	1340.0
92.00	1510.0
93.00	9333.0
93.90	670.0
96.10	600.0
97.00	381.0
98.00	7103.0
99.00	5751.0
99.90	651.0
101.00	3312.0
102.90	1291.0
104.00	2405.0
105.00	2104.0
106.10	872.0
107.00	26176.0
108.00	3979.0
109.10	916.0
110.00	50720.0
111.00	7142.0
111.90	989.0
116.00	2126.0
117.00	19536.0
118.00	1368.0
119.90	397.0
122.00	1766.0
123.00	2494.0
124.00	1271.0
125.00	1443.0
125.80	388.0
127.00	91592.0
128.00	6685.0
129.00	36576.0
130.00	2702.0
131.00	612.0
134.00	1021.0
135.00	2989.0
135.90	1227.0
137.00	1718.0
137.70	605.0
138.90	407.0
140.00	494.0
141.00	4853.0
142.00	1940.0
143.00	1166.0
144.00	560.0
146.00	1074.0
147.00	2466.0
148.00	5185.0
149.00	1231.0
150.80	869.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

151.60	578.0
152.90	1813.0
154.00	801.0
155.10	3082.0
156.00	4210.0
157.20	926.0
157.90	765.0
158.90	802.0
159.90	1342.0
161.00	2469.0
161.80	1010.0
165.00	1838.0
165.90	1698.0
167.00	8944.0
168.00	4637.0
168.80	711.0
169.80	437.0
171.00	566.0
172.00	1010.0
173.10	1579.0
174.10	1890.0
175.00	3938.0
175.90	1242.0
176.90	2277.0
178.10	717.0
179.00	7584.0
180.00	5257.0
181.00	2437.0
182.00	428.0
182.80	307.0
183.80	827.0
185.00	4025.0
186.00	31152.0
187.00	8221.0
187.90	818.0
189.10	1776.0
190.80	1088.0
192.00	2237.0
193.00	3197.0
193.90	608.0
195.10	374.0
196.00	6184.0
197.90	225344.0
198.90	15638.0
199.90	1304.0
201.50	1131.0
203.00	1361.0
203.90	7147.0
205.00	13837.0
206.00	57272.0
207.00	9851.0
207.90	2229.0
209.00	776.0
210.10	921.0
211.10	1818.0
215.00	690.0
215.90	1092.0
217.00	14353.0
217.90	1846.0
221.00	8922.0
221.90	2140.0
223.00	3356.0
224.00	29592.0
225.00	8804.0
226.00	894.0
227.00	12303.0
228.00	2198.0
229.00	2857.0
229.90	476.0
231.00	963.0
233.80	858.0
234.90	806.0
236.00	490.0
236.90	904.0
238.90	441.0
240.00	507.0
241.00	707.0
241.90	2132.0
243.00	2108.0
244.00	25040.0
245.00	3511.0
246.00	4663.0
247.00	915.0
249.00	673.0
253.00	794.0
255.00	130432.0
256.00	18784.0
257.00	1507.0
258.00	6856.0
259.00	1178.0
265.00	2710.0
266.00	422.0
272.00	323.0
273.00	3757.0
274.00	11098.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

275.00	59040.0
276.00	9046.0
277.00	4768.0
278.00	869.0
284.10	387.0
285.00	870.0
292.90	1293.0
294.00	457.0
296.00	15927.0
297.00	2057.0
303.00	1942.0
304.00	519.0
308.00	322.0
314.10	1048.0
314.90	1734.0
316.00	1228.0
321.00	756.0
323.00	4735.0
324.00	1198.0
326.90	917.0
328.00	592.0
332.00	443.0
333.00	432.0
334.00	3306.0
335.00	896.0
341.10	657.0
345.90	975.0
352.00	1426.0
352.90	1028.0
354.00	2172.0
355.00	507.0
365.00	7326.0
366.00	1093.0
371.00	599.0
372.00	3139.0
372.90	714.0
382.90	886.0
390.00	642.0
402.00	846.0
402.90	1666.0
403.90	579.0
404.10	641.0
421.00	1150.0
422.10	1193.0
423.00	9981.0
424.00	1878.0
441.00	26312.0
442.00	172032.0
443.00	31312.0
444.00	3460.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

m/z	Abundance
36.90	543.0
38.00	1010.0
39.00	4626.0
40.00	1303.0
41.00	584.0
41.90	371.0
43.10	461.0
44.00	4200.0
46.90	1385.0
47.90	699.0
49.00	7629.0
50.00	15461.0
51.00	62960.0
52.00	2816.0
55.10	751.0
55.90	2162.0
57.00	4871.0
60.90	742.0
62.00	1159.0
63.00	2643.0
64.00	391.0
65.00	1480.0
68.10	1512.0
69.00	66600.0
69.90	585.0
73.00	1160.0
74.00	6283.0
75.00	10279.0
76.00	3132.0
77.00	76272.0
78.00	5408.0
79.00	5853.0
80.00	4183.0
80.90	4819.0
82.00	1463.0
82.90	1461.0
83.90	5362.0
84.90	1151.0
86.00	4393.0
86.90	751.0
87.90	907.0
91.00	1092.0
92.00	1318.0
93.00	8501.0
94.00	587.0
96.10	668.0
97.20	373.0
98.00	6495.0
99.00	5859.0
100.10	403.0
100.90	3147.0
103.00	1080.0
104.00	2166.0
105.00	2068.0
107.00	25008.0
108.00	3208.0
110.00	45720.0
111.00	7264.0
111.90	759.0
112.90	338.0
116.00	1122.0
117.00	18272.0
118.00	1208.0
119.00	360.0
119.90	373.0
122.00	1582.0
122.90	2916.0
124.10	1703.0
125.10	944.0
126.10	764.0
127.00	91008.0
128.00	6281.0
129.00	33848.0
129.90	2451.0
130.90	563.0
131.90	506.0
134.00	1259.0
135.00	2526.0
135.90	995.0
137.10	1501.0
138.00	341.0
140.00	626.0
141.00	5039.0
142.00	1709.0
142.90	1142.0
145.90	980.0
147.00	2113.0
148.00	4955.0
149.00	1460.0
150.90	462.0
151.60	637.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

153.00	1248.0
154.10	1283.0
155.00	2327.0
155.90	3981.0
157.00	981.0
157.80	900.0
159.00	744.0
159.90	1316.0
160.90	2357.0
161.90	519.0
165.00	1728.0
165.90	1707.0
167.00	8842.0
167.90	4531.0
169.00	978.0
170.00	498.0
170.90	346.0
171.90	1187.0
173.10	1156.0
174.00	2292.0
174.90	4169.0
176.10	1124.0
177.00	1868.0
178.00	1111.0
179.00	6730.0
180.00	4750.0
180.90	2971.0
182.20	402.0
183.00	417.0
184.00	759.0
185.00	4579.0
186.00	28608.0
187.00	9345.0
187.90	800.0
189.00	1883.0
189.90	396.0
191.00	805.0
192.00	2039.0
193.00	2950.0
194.00	703.0
195.10	382.0
196.00	5333.0
197.90	225280.0
199.00	15167.0
200.00	1161.0
201.30	873.0
203.00	1286.0
204.00	7810.0
205.00	13481.0
206.00	60392.0
207.00	8868.0
208.00	2165.0
209.00	774.0
209.90	741.0
211.00	2849.0
211.70	537.0
215.00	820.0
216.10	1254.0
216.90	14487.0
218.00	2079.0
219.20	343.0
221.00	8648.0
223.00	3424.0
224.00	29488.0
225.00	7439.0
226.00	744.0
227.00	12108.0
228.10	1826.0
228.90	2784.0
230.10	351.0
231.00	1406.0
233.80	896.0
235.00	1052.0
235.90	535.0
237.00	1364.0
238.90	607.0
240.90	682.0
241.90	1932.0
243.00	1991.0
244.00	24712.0
245.00	3165.0
246.00	5440.0
247.00	1055.0
249.00	1015.0
249.80	317.0
253.10	572.0
255.00	135424.0
256.00	19912.0
256.90	1635.0
258.00	7112.0
259.00	1232.0
264.90	2958.0
266.10	715.0
271.00	464.0
272.00	448.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66

273.00	3966.0
274.00	11380.0
275.00	66336.0
276.00	7772.0
277.00	5374.0
278.10	656.0
278.80	327.0
283.00	620.0
283.90	475.0
285.00	1157.0
293.00	1006.0
295.00	945.0
296.00	16488.0
297.00	2294.0
300.90	404.0
303.10	2080.0
304.00	500.0
308.00	330.0
314.00	976.0
315.00	2104.0
315.90	1123.0
320.80	686.0
323.00	6119.0
324.10	1280.0
325.80	374.0
326.90	976.0
327.80	642.0
332.00	473.0
333.00	615.0
334.00	3090.0
335.10	836.0
341.00	874.0
346.00	1072.0
352.00	2023.0
353.10	1182.0
354.00	1868.0
355.00	366.0
365.00	8883.0
365.90	1340.0
370.90	345.0
372.00	3512.0
373.00	914.0
383.00	1209.0
389.90	797.0
390.90	527.0
401.40	334.0
401.90	1127.0
403.00	1885.0
404.10	807.0
420.90	1609.0
422.10	1832.0
423.00	11442.0
423.90	2255.0
441.00	33816.0
442.00	212096.0
443.00	38352.0
444.00	3928.0

Instrument :
BNA_N
ClientSampleId :
DFTPP66