


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051119\
 Data File : BN005621.D
 Acq On : 11 May 2019 10:49
 Operator : JU/SJ
 Sample : DFTPP67
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP67

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN041919MA.M
 Title : SVOA CALIBRATION
 Last Update : Sat May 11 23:17:30 2019



AutoFind: Scans 2455, 2456, 2457; Background Corrected with Scan 2450

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.2	187392	PASS
68	69	0.00	2	1.9	3689	PASS
69	198	0.00	100	32.6	195669	PASS
70	69	0.00	2	0.6	1096	PASS
127	198	10	80	40.6	243904	PASS
197	198	0.00	2	0.3	1553	PASS
198	198	100	100	100.0	600746	PASS
199	198	5	9	6.6	39496	PASS
275	198	10	60	26.6	159712	PASS
365	198	1	100	3.1	18451	PASS
441	198	0.01	100	12.7	76192	PASS
442	198	50	100	80.9	485994	PASS
443	442	15	24	18.7	90898	PASS

Scan 2450 (17.398 min): BN005621.D
DFTPP67

m/z	Abundance
39.90	1280.0
44.00	3736.0
207.00	1654.0
208.10	495.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

m/z	Abundance
37.00	604.0
38.00	1789.0
39.00	11828.0
39.90	1578.0
41.00	397.0
44.00	3802.0
45.00	463.0
49.10	1786.0
50.00	40184.0
51.00	158912.0
52.00	8500.0
52.90	414.0
55.00	964.0
56.00	4985.0
57.00	10642.0
57.90	584.0
61.00	2181.0
62.00	2201.0
63.00	6293.0
64.00	765.0
65.00	3585.0
68.00	3020.0
69.00	160512.0
70.00	861.0
72.90	1444.0
74.00	16297.0
75.00	25096.0
76.10	8294.0
77.00	178880.0
78.00	12130.0
79.00	12373.0
79.90	8711.0
81.00	12384.0
81.90	3208.0
83.00	2526.0
83.90	361.0
85.00	2438.0
86.00	4221.0
87.00	1546.0
88.00	613.0
91.00	3011.0
91.90	3509.0
93.00	20560.0
94.00	1387.0
94.90	613.0
95.90	1004.0
97.00	504.0
98.00	15338.0
99.00	12524.0
100.00	1013.0
101.00	6662.0
101.90	534.0
103.00	2863.0
104.00	4823.0
105.00	4523.0
106.10	1493.0
107.00	55320.0
108.00	8064.0
109.00	1504.0
110.00	100624.0
111.00	16680.0
112.00	2285.0
112.90	679.0
116.00	3608.0
117.00	39768.0
117.90	3566.0
120.00	457.0
121.30	382.0
121.90	3507.0
123.00	5802.0
124.00	2542.0
124.90	2053.0
127.00	184832.0
128.00	14679.0
129.00	72664.0
129.90	6180.0
130.90	1451.0
131.90	735.0
132.90	345.0
133.90	1682.0
135.00	6087.0
136.00	2358.0
137.00	2518.0
137.80	636.0
139.00	497.0
140.00	764.0
141.00	8668.0
142.00	3378.0
142.90	1884.0
144.00	491.0
145.00	646.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

146.00	1869.0
147.00	4201.0
147.90	10127.0
149.00	2820.0
149.90	660.0
151.10	1111.0
152.00	618.0
153.00	3105.0
154.00	2537.0
155.00	5562.0
156.00	8079.0
157.00	1933.0
157.90	2079.0
158.90	1458.0
160.00	3195.0
161.00	4482.0
161.80	1164.0
164.10	582.0
165.00	3969.0
166.20	2464.0
167.00	18112.0
168.00	8728.0
169.00	2300.0
170.00	873.0
170.90	950.0
171.90	1730.0
172.90	2034.0
174.00	3808.0
175.00	7478.0
176.10	1883.0
177.00	3654.0
178.00	1425.0
179.00	14193.0
180.00	10164.0
180.90	4717.0
182.00	875.0
182.80	508.0
183.90	1308.0
185.00	6878.0
186.00	56920.0
187.00	15615.0
187.90	2307.0
189.00	3784.0
189.70	560.0
191.00	2114.0
191.90	4939.0
193.00	4767.0
194.00	1319.0
194.90	758.0
196.00	10858.0
197.90	427072.0
198.90	29160.0
200.00	1898.0
201.50	1452.0
202.90	2889.0
204.00	14004.0
205.00	24352.0
206.00	105672.0
207.00	16280.0
208.00	3662.0
208.90	1223.0
211.00	4419.0
212.00	913.0
214.90	1505.0
216.00	2510.0
216.90	26216.0
218.00	3623.0
219.90	538.0
221.00	17248.0
221.70	5159.0
223.00	5944.0
224.00	58680.0
225.00	14973.0
226.00	1651.0
227.00	22656.0
227.90	3843.0
228.90	5837.0
229.90	805.0
231.00	2089.0
231.90	442.0
232.90	608.0
233.90	1452.0
235.00	1926.0
236.00	1201.0
237.00	1675.0
238.90	1174.0
240.10	670.0
241.00	1551.0
242.00	2908.0
243.10	2821.0
244.00	44096.0
245.00	6458.0
245.90	8134.0
247.00	1977.0
249.00	1356.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

250.90	426.0
252.00	362.0
252.80	863.0
253.20	797.0
255.00	233728.0
256.00	33560.0
257.00	2288.0
258.00	12145.0
258.90	2126.0
259.90	351.0
261.00	363.0
263.90	347.0
265.00	5073.0
266.00	982.0
271.00	772.0
272.00	610.0
273.00	6353.0
274.00	16392.0
275.00	102432.0
276.00	14316.0
277.00	8103.0
278.00	1828.0
279.00	426.0
282.10	407.0
283.00	1004.0
283.80	514.0
285.00	1638.0
288.90	320.0
290.00	346.0
291.00	414.0
291.90	393.0
293.10	1680.0
294.00	645.0
295.00	915.0
296.00	25136.0
296.90	3389.0
300.90	403.0
301.90	559.0
303.00	2508.0
304.10	919.0
307.90	316.0
309.90	456.0
314.00	1416.0
315.00	2817.0
315.90	1763.0
320.90	713.0
322.00	438.0
323.00	8585.0
324.00	2162.0
327.00	1618.0
327.90	814.0
331.80	720.0
332.90	743.0
333.10	754.0
334.00	5169.0
335.00	1277.0
335.90	316.0
340.90	898.0
342.00	339.0
346.00	1579.0
347.00	392.0
350.90	317.0
352.00	2243.0
353.00	1867.0
354.10	2938.0
355.00	943.0
364.90	10251.0
366.10	1696.0
369.80	355.0
370.90	935.0
372.00	3921.0
373.00	1213.0
382.90	1041.0
384.10	457.0
389.90	501.0
391.00	469.0
392.10	312.0
402.10	1589.0
403.00	2315.0
403.90	1011.0
421.00	2352.0
422.10	1768.0
423.00	15859.0
424.10	3356.0
424.90	349.0
439.20	330.0
441.00	38768.0
442.00	244544.0
443.00	47248.0
444.00	4357.0
445.00	476.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

m/z	Abundance
37.00	805.0
38.00	2375.0
39.00	15139.0
40.00	1537.0
41.00	805.0
44.00	3347.0
45.00	408.0
49.00	1289.0
50.00	54152.0
51.00	220736.0
52.00	11560.0
53.00	422.0
54.90	1131.0
56.00	6672.0
57.00	15207.0
58.00	664.0
59.00	311.0
61.00	2812.0
62.00	3204.0
63.00	9423.0
64.00	1228.0
65.00	4692.0
67.00	309.0
68.10	4610.0
69.00	233344.0
70.10	1157.0
73.10	1636.0
74.00	23880.0
75.00	36912.0
76.00	11628.0
77.00	252288.0
78.00	16472.0
79.00	17080.0
80.00	14279.0
81.00	18688.0
81.90	4909.0
82.90	4515.0
83.90	537.0
84.90	3250.0
86.00	5672.0
87.00	2326.0
87.90	834.0
89.00	495.0
90.90	4193.0
92.00	4583.0
93.00	30080.0
93.90	2108.0
95.10	634.0
96.10	1266.0
96.80	560.0
98.00	23984.0
99.00	17696.0
100.10	1563.0
101.00	10473.0
101.90	695.0
103.00	3911.0
104.00	6678.0
104.90	6709.0
105.90	1979.0
107.00	84592.0
108.00	12696.0
109.10	2586.0
110.00	153216.0
111.00	23568.0
111.90	2315.0
113.00	1122.0
116.00	4486.0
117.00	62624.0
117.90	5235.0
119.00	791.0
119.80	879.0
121.00	539.0
122.00	5812.0
123.00	8870.0
124.00	4239.0
125.00	3678.0
127.00	286848.0
128.00	22104.0
129.00	110360.0
130.00	8695.0
131.00	1904.0
132.10	800.0
132.80	563.0
133.90	3244.0
135.00	8403.0
136.00	3694.0
137.00	4867.0
138.00	1067.0
138.90	545.0
139.90	1722.0
141.00	14461.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

142.00	5214.0
143.00	3648.0
143.90	923.0
145.00	949.0
146.00	2356.0
147.00	7394.0
148.00	16440.0
149.00	3681.0
149.90	991.0
151.00	1932.0
152.00	1266.0
153.00	5175.0
154.00	3720.0
155.00	9022.0
156.00	12417.0
157.00	2774.0
158.00	2851.0
159.00	2240.0
160.00	5015.0
161.00	7717.0
161.90	2066.0
163.00	556.0
163.90	676.0
165.00	5868.0
166.00	5165.0
167.00	29720.0
168.00	13378.0
169.00	2311.0
170.00	1107.0
171.00	952.0
171.90	3151.0
173.00	3576.0
174.00	5901.0
175.00	11583.0
175.90	4324.0
177.00	6016.0
178.00	2333.0
179.00	22992.0
180.00	15561.0
181.00	8158.0
182.00	1371.0
182.70	521.0
183.90	1961.0
185.00	10467.0
186.00	93072.0
187.00	27328.0
188.00	2594.0
189.00	5753.0
190.00	816.0
191.00	3109.0
191.90	7046.0
193.00	7021.0
194.00	1887.0
194.90	974.0
196.00	18000.0
196.80	4659.0
197.90	705600.0
198.90	45120.0
200.00	3566.0
201.30	2733.0
203.00	4752.0
204.00	22136.0
205.00	42592.0
206.00	175488.0
207.00	24984.0
208.00	6197.0
209.00	2210.0
210.20	2845.0
211.00	6785.0
212.90	599.0
214.00	428.0
215.00	2063.0
216.00	4107.0
217.00	46672.0
218.00	6647.0
218.90	627.0
221.00	31360.0
221.80	7276.0
223.00	11130.0
224.00	97912.0
225.00	24696.0
226.00	2885.0
227.00	38256.0
228.00	6881.0
229.00	8157.0
229.80	969.0
231.00	3302.0
232.00	808.0
232.90	638.0
233.90	2860.0
235.00	3184.0
235.90	1876.0
236.90	2466.0
237.90	512.0
239.00	1795.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

239.90	958.0
240.90	2531.0
242.00	4695.0
243.00	5303.0
244.00	77352.0
245.00	11210.0
246.00	14162.0
246.90	3476.0
248.00	548.0
248.90	3061.0
249.80	480.0
251.00	956.0
252.00	906.0
253.10	1611.0
255.00	395264.0
256.00	60680.0
257.00	5202.0
258.00	21144.0
258.90	2983.0
259.90	848.0
261.00	673.0
263.90	644.0
265.00	7964.0
266.00	1427.0
270.00	464.0
270.90	975.0
271.90	1036.0
273.00	11219.0
274.00	31552.0
275.00	183872.0
276.00	23976.0
277.00	15286.0
278.00	2870.0
279.00	551.0
281.90	468.0
283.00	1508.0
284.00	894.0
285.00	2811.0
286.10	713.0
289.00	684.0
289.90	444.0
291.00	539.0
292.10	821.0
293.00	3639.0
294.00	683.0
296.00	47608.0
297.00	6531.0
298.00	485.0
300.80	347.0
301.20	430.0
302.00	1202.0
303.00	5128.0
304.10	1812.0
308.00	859.0
308.70	325.0
310.00	548.0
311.90	326.0
313.00	601.0
314.00	2558.0
315.00	5202.0
316.00	2529.0
317.10	662.0
321.00	2070.0
322.00	659.0
323.00	17256.0
324.10	2578.0
325.20	513.0
325.90	392.0
327.00	2759.0
327.80	1553.0
332.00	1145.0
333.00	1557.0
334.00	10593.0
335.00	2623.0
340.90	1841.0
342.00	591.0
346.00	3461.0
347.00	316.0
351.00	474.0
352.00	4973.0
353.00	3774.0
354.00	5468.0
355.00	1041.0
358.90	475.0
364.00	352.0
365.00	21112.0
366.00	3295.0
370.00	330.0
371.00	1470.0
372.10	8809.0
373.00	2041.0
373.90	360.0
382.90	2197.0
384.00	456.0
390.00	972.0

Instrument :

BNA_N

ClientSampleId :

DFTPP67

390.90	831.0
392.10	785.0
400.80	599.0
402.00	3304.0
403.00	4619.0
404.10	1726.0
415.10	322.0
421.00	4847.0
421.90	3382.0
423.00	33320.0
424.00	7065.0
424.90	921.0
438.30	446.0
439.20	525.0
441.00	85632.0
442.00	544192.0
443.00	103320.0
444.00	10756.0
445.00	615.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

m/z	Abundance
37.00	934.0
38.00	2023.0
39.00	13325.0
40.00	1354.0
40.90	559.0
44.00	3435.0
48.90	1385.0
50.00	46600.0
51.00	182528.0
52.00	8891.0
52.90	413.0
54.90	1140.0
56.00	5953.0
57.00	13444.0
57.90	440.0
61.00	2268.0
62.10	2713.0
63.00	7906.0
63.90	1439.0
65.00	3858.0
68.00	3439.0
69.00	193152.0
70.00	1272.0
73.10	2005.0
74.00	19128.0
75.00	31288.0
76.10	10437.0
77.00	213888.0
78.00	14877.0
79.00	15518.0
80.00	11911.0
81.00	16289.0
81.90	4179.0
83.00	3684.0
83.90	604.0
85.00	3084.0
86.00	4637.0
87.00	2089.0
88.00	959.0
91.00	4130.0
92.00	4180.0
93.00	26272.0
93.90	1569.0
94.90	413.0
96.00	1381.0
98.00	20496.0
99.00	16768.0
100.00	1211.0
101.00	9519.0
102.00	494.0
103.00	3287.0
104.00	6122.0
105.00	4990.0
106.00	1709.0
107.00	75616.0
108.00	11681.0
109.00	2352.0
110.00	133888.0
111.00	19120.0
111.90	1866.0
113.00	920.0
115.90	4037.0
117.00	53104.0
117.90	4137.0
118.90	792.0
119.90	996.0
121.90	5266.0
123.00	7613.0
123.90	3539.0
125.00	3721.0
127.00	260032.0
128.00	18504.0
129.00	101072.0
130.00	8828.0
131.00	1519.0
132.00	849.0
133.00	435.0
134.00	2941.0
135.00	8430.0
135.90	3594.0
137.00	4430.0
138.00	800.0
138.80	675.0
140.00	1627.0
141.00	13225.0
141.90	4327.0
143.00	2895.0
143.90	794.0
145.00	714.0
145.90	2319.0
147.00	6931.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

148.00	14982.0
149.00	3655.0
150.10	692.0
151.00	1839.0
151.80	1027.0
152.90	4797.0
154.00	3601.0
155.00	7895.0
156.00	12494.0
157.10	2318.0
157.90	2751.0
159.00	2254.0
160.00	4657.0
161.00	6193.0
162.00	2079.0
163.00	483.0
164.20	604.0
164.90	5720.0
166.00	4559.0
167.00	26360.0
168.00	11962.0
169.00	2413.0
169.80	963.0
170.90	1768.0
172.00	2955.0
173.00	3449.0
174.00	7003.0
175.00	12428.0
176.00	3351.0
176.90	5124.0
178.00	1852.0
179.00	21720.0
180.00	14529.0
181.00	6654.0
182.00	1400.0
183.10	655.0
184.00	1661.0
185.00	12024.0
186.00	83824.0
187.00	25456.0
188.10	2488.0
188.90	4367.0
190.00	840.0
191.00	2987.0
192.00	7196.0
193.00	7603.0
194.00	1793.0
194.90	942.0
196.00	16912.0
197.90	669568.0
198.90	44208.0
199.90	3280.0
201.70	3190.0
202.90	4225.0
204.00	23328.0
205.00	41880.0
206.00	172224.0
207.00	26048.0
208.00	6055.0
209.10	2240.0
210.00	2482.0
210.90	7053.0
212.00	1089.0
213.00	570.0
214.90	1970.0
216.10	3946.0
217.00	45576.0
218.00	6101.0
219.00	679.0
221.00	29160.0
221.80	7136.0
223.00	9972.0
224.00	98368.0
225.00	24936.0
226.10	2833.0
227.00	38312.0
228.00	5453.0
229.00	8141.0
230.00	1060.0
231.00	3596.0
232.00	632.0
233.00	946.0
234.10	2370.0
235.00	2877.0
236.00	2521.0
237.00	2966.0
237.90	403.0
239.00	1658.0
240.00	1393.0
241.00	2598.0
242.00	5504.0
243.10	5421.0
244.00	77128.0
245.00	10819.0
246.00	15326.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

247.00	2659.0
247.90	817.0
248.90	3019.0
249.90	456.0
251.00	895.0
252.00	822.0
253.10	1474.0
255.00	411392.0
256.00	57512.0
257.00	4271.0
258.00	22712.0
259.00	3655.0
260.00	512.0
261.00	598.0
263.00	306.0
264.00	728.0
265.00	8340.0
266.00	1546.0
267.10	312.0
269.80	457.0
271.00	935.0
271.90	1087.0
273.00	12087.0
274.00	33032.0
275.00	192832.0
276.00	24872.0
277.00	15003.0
278.00	3091.0
279.00	626.0
281.10	502.0
282.00	462.0
283.10	1662.0
283.90	1234.0
285.00	3501.0
285.90	603.0
289.00	641.0
290.00	470.0
291.00	455.0
292.00	761.0
292.90	3789.0
293.90	951.0
296.00	46352.0
297.00	6849.0
298.10	413.0
300.90	480.0
302.00	1228.0
303.10	5800.0
304.00	1472.0
307.90	795.0
309.10	418.0
310.00	849.0
313.00	790.0
314.00	2654.0
315.00	5895.0
316.00	2453.0
316.90	678.0
321.00	1690.0
322.20	1226.0
323.00	17048.0
324.00	2786.0
326.90	3457.0
328.10	1778.0
331.90	1538.0
333.00	1723.0
334.00	10934.0
335.00	3020.0
341.10	2067.0
341.70	525.0
346.00	4209.0
346.90	770.0
352.10	5408.0
353.00	3363.0
354.00	6268.0
355.00	1443.0
363.90	459.0
365.00	23992.0
366.00	2968.0
366.90	339.0
370.00	957.0
370.90	1350.0
372.00	10755.0
373.00	2605.0
373.90	430.0
377.00	315.0
383.00	2876.0
384.20	604.0
390.00	1623.0
391.00	862.0
392.10	903.0
401.00	627.0
402.00	3851.0
403.00	6757.0
404.00	2538.0
405.00	461.0
409.90	338.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67

421.00	5979.0
422.00	5441.0
423.00	39776.0
424.00	7168.0
425.10	593.0
438.10	402.0
438.80	329.0
441.00	104176.0
442.00	669248.0
443.00	122128.0
444.00	12125.0
444.90	540.0

Instrument :
BNA_N
ClientSampleId :
DFTPP67