

Data File : BN005675.D
Acq On : 14 May 2019 04:54
Operator : JU/SJ
Sample : K2692-13
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A5176

Quant Time: May 14 07:18:40 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN050719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 07:15:06 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	1301	0.40	ng/ul	-0.01
2) Naphthalene-d8	10.34	136	5622	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.22	164	3815	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.07	188	730020	0.40	ng/ul	0.08
16) Chrysene-d12	0.00	240	0	0.00	ng/ul	-21.22
20) Perylene-d12	23.28	264	754438	0.40	ng/ul	-0.14

System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.97	152	3097	0.35	ng/ul	0.00
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN005675.D
Acq On : 14 May 2019 04:54
Operator : JU/SJ
Sample : K2692-13
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A5176

Quant Time: May 14 07:18:40 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN050719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 07:15:06 2019
Response via : Initial Calibration

