

Data File : BN005676.D
Acq On : 14 May 2019 05:28
Operator : JU/SJ
Sample : K2692-14
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A5177

Quant Time: May 14 07:18:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN050719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 07:15:06 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	1213	0.40	ng/ul	-0.01
2) Naphthalene-d8	10.35	136	5159	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.22	164	3430	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.07	188	413054	0.40	ng/ul	0.08
16) Chrysene-d12	0.00	240	0	0.00	ng/ul	-21.22
20) Perylene-d12	23.30	264	453758	0.40	ng/ul	-0.13

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	11.98	152	1712	0.21	ng/ul	0.02
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN005676.D
Acq On : 14 May 2019 05:28
Operator : JU/SJ
Sample : K2692-14
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A5177

Quant Time: May 14 07:18:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN050719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 07:15:06 2019
Response via : Initial Calibration

