

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051319\
 Data File : BN005673.D
 Acq On : 14 May 2019 03:45
 Operator : JU/SJ
 Sample : K2692-08
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A5113

Manual Integrations
 APPROVED

mohammad
 5/16/2019 11:35:53 AM

Quant Time: May 14 07:32:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN050719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 14 07:15:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	1123	0.40	ng/ul	-0.02
2) Naphthalene-d8	10.34	136	4828	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.22	164	3337	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.98	188	7913m	0.40	ng/ul	-0.01
16) Chrysene-d12	21.21	240	8275m	0.40	ng/ul	0.00
20) Perylene-d12	23.42	264	7722m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.94	152	2393	0.32	ng/ul	-0.02
14) Fluoranthene-d10	19.02	212	9051m	0.36	ng/ul	-0.01
Target Compounds						Qvalue
3) Naphthalene	10.39	128	483636	37.855	ng/ul	96
5) 2-Methylnaphthalene	12.01	142	497435	57.019	ng/ul	98
7) Acenaphthylene	13.94	152	504274	30.013	ng/ul	97
8) Acenaphthene	14.28	153	336420	28.300	ng/ul	96
9) Fluorene	15.28	166	496593	35.757	ng/ul	98
12) Phenanthrene	17.02	178	735496m	34.113	ng/ul	
13) Anthracene	17.11	178	1129946m	52.537	ng/ul	
15) Fluoranthene	19.05	202	894366m	29.971	ng/ul	
17) Pyrene	19.42	202	803559	26.295	ng/ul	95
18) Benzo(a)anthracene	21.19	228	743861	26.073	ng/ul	100
19) Chrysene	21.25	228	618150	17.643	ng/ul	99
21) Benzo(b)fluoranthene	22.76	252	1278334	55.059	ng/ul	95
22) Benzo(k)fluoranthene	22.81	252	656859m	22.575	ng/ul	
23) Benzo(a)pyrene	23.32	252	621615m	23.960	ng/ul	
24) Indeno(1,2,3-cd)pyrene	25.61	276	718648m	24.761	ng/ul	
25) Dibenzo(a,h)anthracene	25.62	278	400982	16.715	ng/ul#	93
26) Benzo(g,h,i)perylene	26.28	276	936543m	37.843	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051319\
Data File : BN005673.D
Acq On : 14 May 2019 03:45
Operator : JU/SJ
Sample : K2692-08
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A5113

Manual Integrations
APPROVED

mohammad
5/16/2019 11:35:53 AM

Quant Time: May 14 07:32:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN050719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 07:15:06 2019
Response via : Initial Calibration

