

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051319\
 Data File : BN005668.D
 Acq On : 14 May 2019 00:52
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.474

Quant Time: May 14 05:28:39 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN050719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 14 05:24:26 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	60	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	68	0.00
3	Naphthalene	1.059	1.074	-1.4	71	0.00
4 SURR	2-Methylnaphthalene-d10	0.626	0.677	-8.1	75	0.00
5	2-Methylnaphthalene	0.723	0.787	-8.9	75	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	78	0.00
7	Acenaphthylene	2.014	2.098	-4.2	81	0.00
8 C	Acenaphthene	1.425	1.470	-3.2	81	0.00
9	Fluorene	1.665	1.744	-4.7	82	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	80	0.00
11	Pentachlorophenol	0.104	0.066	36.5#	62	0.00
12	Phenanthrene	1.090	1.151	-5.6	85	0.00
13	Anthracene	1.087	1.121	-3.1	84	0.00
14 SURR	Fluoranthene-d10	1.255	1.254	0.1	82	0.00
15 C	Fluoranthene	1.508	1.500	0.5	82	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	74	0.00
17	Pyrene	1.477	1.602	-8.5	81	0.00
18	Benzo(a)anthracene	1.379	1.498	-8.6	81	0.00
19	Chrysene	1.694	1.642	3.1	72	0.00
20 I	Perylene-d12	1.000	1.000	0.0	62	0.00
21	Benzo(b)fluoranthene	1.203	1.427	-18.6	76	0.00
22	Benzo(k)fluoranthene	1.507	1.551	-2.9	64	0.00
23 C	Benzo(a)pyrene	1.344	1.405	-4.5	66	0.00
24	Indeno(1,2,3-cd)pyrene	1.503	1.408	6.3	59	0.00
25	Dibenzo(a,h)anthracene	1.243	1.154	7.2	59	0.00
26	Benzo(g,h,i)perylene	1.282	1.149	10.4	57	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0