

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051319\
 Data File : BN005693.D
 Acq On : 14 May 2019 18:07
 Operator : JU/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.476

Quant Time: May 14 18:41:45 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN050719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 14 07:15:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	69	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	80	0.00
3	Naphthalene	1.059	1.081	-2.1	84	0.00
4 SURR	2-Methylnaphthalene-d10	0.626	0.696	-11.2	90	0.00
5	2-Methylnaphthalene	0.723	0.817	-13.0	91	-0.01
6 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
7	Acenaphthylene	2.014	2.101	-4.3	99	0.00
8 C	Acenaphthene	1.425	1.481	-3.9	99	0.00
9	Fluorene	1.665	1.747	-4.9	100	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
11	Pentachlorophenol	0.104	0.089	14.4	102	-0.01
12	Phenanthrene	1.090	1.166	-7.0	105	0.00
13	Anthracene	1.087	1.122	-3.2	102	0.00
14 SURR	Fluoranthene-d10	1.255	1.276	-1.7	101	0.00
15 C	Fluoranthene	1.508	1.534	-1.7	102	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	94	0.01
17	Pyrene	1.477	1.560	-5.6	100	0.00
18	Benzo(a)anthracene	1.379	1.543	-11.9	107	0.01
19	Chrysene	1.694	1.603	5.4	89	0.02
20 I	Perylene-d12	1.000	1.000	0.0	81	0.04
21	Benzo(b)fluoranthene	1.203	1.468	-22.0#	103	0.04
22	Benzo(k)fluoranthene	1.507	1.519	-0.8	82	0.04
23 C	Benzo(a)pyrene	1.344	1.410	-4.9	87	0.04
24	Indeno(1,2,3-cd)pyrene	1.503	1.419	5.6	78	0.04
25	Dibenzo(a,h)anthracene	1.243	1.184	4.7	79	0.04
26	Benzo(g,h,i)perylene	1.282	1.159	9.6	76	0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0