

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051320\
 Data File : BN010708.D
 Acq On : 13 May 2020 15:52
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN051320

Quant Time: May 13 16:25:51 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 13 14:18:48 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.345	0.365	-5.8	114	0.00
3	n-Nitrosodimethylamine	0.199	0.200	-0.5	110	0.00
4 S	2-Fluorophenol	0.783	0.786	-0.4	104	0.00
5 S	Phenol-d6	0.957	0.936	2.2	102	0.00
6	bis(2-Chloroethyl)ether	0.986	0.988	-0.2	99	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.278	0.270	2.9	100	0.00
9	Naphthalene	1.006	1.003	0.3	102	0.00
10	Hexachlorobutadiene	0.230	0.233	-1.3	101	-0.01
11 SURR	2-Methylnaphthalene-d10	0.652	0.643	1.4	100	0.00
12	2-Methylnaphthalene	0.709	0.696	1.8	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.145	0.133	8.3	104	0.00
15 S	2-Fluorobiphenyl	1.468	1.465	0.2	100	0.00
16	Acenaphthylene	1.634	1.533	6.2	101	0.00
17	Acenaphthene	1.104	1.110	-0.5	101	0.00
18	Fluorene	1.447	1.411	2.5	102	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
20	4-Bromophenyl-phenylether	0.221	0.215	2.7	102	0.00
21	Hexachlorobenzene	0.263	0.260	1.1	103	0.00
22	Pentachlorophenol	0.081	0.075	7.4	110	0.00
23	Phenanthrene	1.073	1.037	3.4	104	0.00
24	Anthracene	0.902	0.838	7.1	104	0.00
25 SURR	Fluoranthene-d10	1.085	1.039	4.2	108	0.00
26	Fluoranthene	1.219	1.167	4.3	107	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	115	0.01
28	Pyrene	1.434	1.379	3.8	108	0.00
29 S	Terphenyl-d14	0.949	0.933	1.7	112	0.00
30	Benzo(a)anthracene	1.167	1.155	1.0	123	0.00
31	Chrysene	1.321	1.301	1.5	112	0.00
32	Bis(2-ethylhexyl)phthalate	0.511	0.473	7.4	114	0.01
33	Indeno(1,2,3-cd)pyrene	1.175	1.334	-13.5	139	0.00
34 I	Perylene-d12	1.000	1.000	0.0	115	0.00
35	Benzo(b)fluoranthene	1.322	1.278	3.3	117	0.00
36	Benzo(k)fluoranthene	1.424	1.492	-4.8	124	0.00
37 C	Benzo(a)pyrene	1.144	1.110	3.0	115	0.00
38	Dibenzo(a,h)anthracene	0.978	1.174	-20.0	150#	0.00
39	Benzo(g,h,i)perylene	1.147	1.198	-4.4	126	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051320\
Data File : BN010708.D
Acq On : 13 May 2020 15:52
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN051320

Quant Time: May 13 16:25:51 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 13 14:18:48 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			