


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051320\
 Data File : BN010699.D
 Acq On : 13 May 2020 08:47
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051320.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed May 13 14:18:48 2020



AutoFind: Scans 2487, 2488, 2489; Background Corrected with Scan 2480

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	20.8	64709	PASS
68	69	0.00	2	1.2	1434	PASS
69	198	0.00	100	37.0	115347	PASS
70	69	0.00	2	0.6	694	PASS
127	198	10	80	45.1	140509	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	311552	PASS
199	198	5	9	6.9	21506	PASS
275	198	10	60	28.0	87213	PASS
365	198	1	100	4.2	13116	PASS
441	198	0.01	100	11.3	35272	PASS
442	198	50	100	76.9	239637	PASS
443	442	15	24	19.3	46197	PASS