

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051322\
 Data File : BN019765.D
 Acq On : 13 May 2022 10:42
 Operator : CG/JU
 Sample : PB144707BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB144707BL

Quant Time: May 13 23:58:59 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 13 02:41:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3604	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	10706	0.400	ng	0.00
13) Acenaphthene-d10	14.474	164	5825	0.400	ng	0.00
19) Phenanthrene-d10	17.215	188	13511	0.400	ng	0.00
29) Chrysene-d12	21.396	240	12766	0.400	ng	0.00
35) Perylene-d12	23.735	264	12385	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.442	112	3493	0.387	ng	0.00
5) Phenol-d6	7.017	99	3991	0.353	ng	0.00
8) Nitrobenzene-d5	9.003	82	2710	0.335	ng	0.00
11) 2-Methylnaphthalene-d10	12.231	152	6337	0.340	ng	0.00
14) 2,4,6-Tribromophenol	15.964	330	634	0.275	ng	0.00
15) 2-Fluorobiphenyl	13.105	172	8584	0.338	ng	0.00
27) Fluoranthene-d10	19.240	212	14868	0.387	ng	0.00
31) Terphenyl-d14	19.843	244	10195	0.339	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051322\
Data File : BN019765.D
Acq On : 13 May 2022 10:42
Operator : CG/JU
Sample : PB144707BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB144707BL

Quant Time: May 13 23:58:59 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 13 02:41:37 2022
Response via : Initial Calibration

