

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051322\
 Data File : BN019776.D
 Acq On : 13 May 2022 19:53
 Operator : CG/JU
 Sample : N2794-03MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW9-MW01D1-20220510MS

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 05/16/2022
 Supervised By : Jagrut Upadhyay 05/16/2022

Quant Time: May 13 23:59:57 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 13 02:41:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3791	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	11648	0.400	ng	0.00
13) Acenaphthene-d10	14.474	164	6899	0.400	ng	0.00
19) Phenanthrene-d10	17.215	188	15954	0.400	ng	0.00
29) Chrysene-d12	21.395	240	15698	0.400	ng	0.00
35) Perylene-d12	23.738	264	14429	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.442	112	1196	0.126	ng	0.00
5) Phenol-d6	7.017	99	935	0.079	ng	0.00
8) Nitrobenzene-d5	9.003	82	2221	0.252	ng	0.00
11) 2-Methylnaphthalene-d10	12.231	152	6229m	0.307	ng	0.00
14) 2,4,6-Tribromophenol	15.954	330	659	0.242	ng	-0.01
15) 2-Fluorobiphenyl	13.105	172	7920	0.263	ng	0.00
27) Fluoranthene-d10	19.244	212	16894	0.373	ng	0.00
31) Terphenyl-d14	19.843	244	12388	0.335	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.355	88	2479m	0.551	ng	
3) n-Nitrosodimethylamine	3.673	42	624	0.127	ng	# 89
6) bis(2-Chloroethyl)ether	7.277	93	2946	0.255	ng	98
9) Naphthalene	10.690	128	9366	0.265	ng	99
10) Hexachlorobutadiene	10.989	225	1482	0.230	ng	# 99
12) 2-Methylnaphthalene	12.302	142	6025	0.250	ng	99
16) Acenaphthylene	14.196	152	8289m	0.256	ng	
17) Acenaphthene	14.538	154	5895	0.246	ng	98
18) Fluorene	15.521	166	7610	0.251	ng	99
20) 4,6-Dinitro-2-methylph...	15.596	198	628	0.274	ng	# 83
21) 4-Bromophenyl-phenylether	16.405	248	2460	0.252	ng	98
22) Hexachlorobenzene	16.528	284	2702	0.250	ng	99
23) Atrazine	16.672	200	1847	0.290	ng	98
24) Pentachlorophenol	16.866	266	1775	0.426	ng	96
25) Phenanthrene	17.256	178	15467	0.280	ng	100
26) Anthracene	17.338	178	12699	0.272	ng	100
28) Fluoranthene	19.274	202	19199	0.324	ng	99
30) Pyrene	19.636	202	19906	0.299	ng	100
32) Benzo(a)anthracene	21.386	228	17583	0.307	ng	99
33) Chrysene	21.431	228	19944	0.311	ng	98
34) Bis(2-ethylhexyl)phtha...	21.324	149	9158	0.384	ng	98
36) Indeno(1,2,3-cd)pyrene	26.141	276	20089	0.297	ng	100
37) Benzo(b)fluoranthene	23.027	252	19256	0.306	ng	98
38) Benzo(k)fluoranthene	23.074	252	19882	0.297	ng	97
39) Benzo(a)pyrene	23.632	252	16781	0.339	ng	98
40) Dibenzo(a,h)anthracene	26.156	278	17034	0.313	ng	99
41) Benzo(g,h,i)perylene	26.872	276	17607	0.314	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051322\
 Data File : BN019776.D
 Acq On : 13 May 2022 19:53
 Operator : CG/JU
 Sample : N2794-03MS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW9-MW01D1-20220510MS

Manual Integrations APPROVED

Reviewed By : Christian Giraldo 05/16/2022
 Supervised By : Jagrut Upadhyay 05/16/2022

Quant Time: May 13 23:59:57 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 13 02:41:37 2022
 Response via : Initial Calibration

