

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051322\
 Data File : BN019781.D
 Acq On : 13 May 2022 22:52
 Operator : CG/JU
 Sample : N2794-06
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW9-MW01D3-F-20220510

Quant Time: May 14 00:00:23 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 13 02:41:37 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 05/16/2022
 Supervised By :Jagrut Upadhyay 05/16/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	4518	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	13997	0.400	ng	0.00
13) Acenaphthene-d10	14.474	164	8404	0.400	ng	0.00
19) Phenanthrene-d10	17.215	188	18777	0.400	ng	0.00
29) Chrysene-d12	21.395	240	19500	0.400	ng	0.00
35) Perylene-d12	23.738	264	18082	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.442	112	1638	0.145	ng	0.00
5) Phenol-d6	7.017	99	1211	0.086	ng	0.00
8) Nitrobenzene-d5	9.003	82	2713m	0.257	ng	0.00
11) 2-Methylnaphthalene-d10	12.231	152	6791	0.279	ng	0.00
14) 2,4,6-Tribromophenol	15.954	330	738	0.222	ng	-0.01
15) 2-Fluorobiphenyl	13.105	172	10416	0.284	ng	0.00
27) Fluoranthene-d10	19.240	212	19651	0.368	ng	0.00
31) Terphenyl-d14	19.839	244	15025	0.327	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.324	149	2908	0.098	ng	Qvalue # 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051322\
Data File : BN019781.D
Acq On : 13 May 2022 22:52
Operator : CG/JU
Sample : N2794-06
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW9-MW01D3-F-20220510

Quant Time: May 14 00:00:23 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 13 02:41:37 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 05/16/2022
Supervised By :Jagrut Upadhyay 05/16/2022

