

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051322\
 Data File : BN019758.D
 Acq On : 12 May 2022 20:13
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICBVN051322

Quant Time: May 13 02:45:50 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 13 02:41:37 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	98	0.00
2	1,4-Dioxane	0.400	0.395	1.3	97	0.00
3	n-Nitrosodimethylamine	0.400	0.368	8.0	92	0.00
4 S	2-Fluorophenol	0.400	0.395	1.3	97	0.00
5 S	Phenol-d6	0.400	0.383	4.3	96	0.00
6	bis(2-Chloroethyl)ether	0.400	0.388	3.0	95	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	97	0.01
8 S	Nitrobenzene-d5	0.400	0.370	7.5	96	0.01
9	Naphthalene	0.400	0.388	3.0	97	0.00
10	Hexachlorobutadiene	0.400	0.397	0.8	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.386	3.5	97	0.00
12	2-Methylnaphthalene	0.400	0.384	4.0	96	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	97	0.01
14 S	2,4,6-Tribromophenol	0.400	0.345	13.8	95	0.00
15 S	2-Fluorobiphenyl	0.400	0.390	2.5	97	0.01
16	Acenaphthylene	0.400	0.365	8.8	98	0.01
17	Acenaphthene	0.400	0.381	4.8	96	0.01
18	Fluorene	0.400	0.382	4.5	96	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	97	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.375	6.3	98	0.00
21	4-Bromophenyl-phenylether	0.400	0.374	6.5	95	0.00
22	Hexachlorobenzene	0.400	0.391	2.3	99	0.00
23	Atrazine	0.400	0.354	11.5	97	0.00
24	Pentachlorophenol	0.400	0.323	19.3	96	0.00
25	Phenanthrene	0.400	0.385	3.8	97	0.01
26	Anthracene	0.400	0.372	7.0	97	0.01
27 SURR	Fluoranthene-d10	0.400	0.383	4.3	98	0.00
28	Fluoranthene	0.400	0.382	4.5	98	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	95	0.00
30	Pyrene	0.400	0.401	-0.3	98	0.00
31 S	Terphenyl-d14	0.400	0.396	1.0	96	0.00
32	Benzo(a)anthracene	0.400	0.366	8.5	92	0.00
33	Chrysene	0.400	0.402	-0.5	97	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.371	7.3	99	0.00
35 I	Perylene-d12	0.400	0.400	0.0	95	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.356	11.0	90	0.00
37	Benzo(b)fluoranthene	0.400	0.382	4.5	94	0.00
38	Benzo(k)fluoranthene	0.400	0.385	3.8	95	0.00
39 C	Benzo(a)pyrene	0.400	0.367	8.3	93	0.00
40	Dibenzo(a,h)anthracene	0.400	0.356	11.0	91	0.00
41	Benzo(g,h,i)perylene	0.400	0.352	12.0	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN051322