

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN051322\
 Data File : BN019761.D
 Acq On : 12 May 2022 23:12
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 13 03:48:56 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 13 02:41:37 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	84	0.00
2	1,4-Dioxane	0.400	0.404	-1.0	86	0.00
3	n-Nitrosodimethylamine	0.400	0.364	9.0	78	0.00
4 S	2-Fluorophenol	0.400	0.375	6.3	79	0.00
5 S	Phenol-d6	0.400	0.370	7.5	80	0.00
6	bis(2-Chloroethyl)ether	0.400	0.390	2.5	82	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	85	0.01
8 S	Nitrobenzene-d5	0.400	0.355	11.3	81	0.01
9	Naphthalene	0.400	0.385	3.8	84	0.00
10	Hexachlorobutadiene	0.400	0.390	2.5	85	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.386	3.5	85	0.00
12	2-Methylnaphthalene	0.400	0.386	3.5	85	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	86	0.01
14 S	2,4,6-Tribromophenol	0.400	0.334	16.5	81	0.00
15 S	2-Fluorobiphenyl	0.400	0.388	3.0	85	0.01
16	Acenaphthylene	0.400	0.355	11.3	85	0.01
17	Acenaphthene	0.400	0.378	5.5	84	0.01
18	Fluorene	0.400	0.381	4.8	85	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	87	0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.349	12.8	81	0.00
21	4-Bromophenyl-phenylether	0.400	0.372	7.0	85	0.00
22	Hexachlorobenzene	0.400	0.385	3.8	87	0.00
23	Atrazine	0.400	0.331	17.3	81	0.00
24	Pentachlorophenol	0.400	0.303	24.3	81	0.00
25	Phenanthrene	0.400	0.381	4.8	86	0.01
26	Anthracene	0.400	0.361	9.8	84	0.00
27 SURR	Fluoranthene-d10	0.400	0.369	7.8	85	0.00
28	Fluoranthene	0.400	0.371	7.3	85	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	85	0.00
30	Pyrene	0.400	0.390	2.5	85	0.00
31 S	Terphenyl-d14	0.400	0.389	2.8	85	0.00
32	Benzo(a)anthracene	0.400	0.367	8.3	83	0.00
33	Chrysene	0.400	0.391	2.3	85	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.344	14.0	83	0.00
35 I	Perylene-d12	0.400	0.400	0.0	89	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.378	5.5	90	0.00
37	Benzo(b)fluoranthene	0.400	0.369	7.8	85	0.00
38	Benzo(k)fluoranthene	0.400	0.394	1.5	91	0.00
39 C	Benzo(a)pyrene	0.400	0.361	9.8	86	0.00
40	Dibenzo(a,h)anthracene	0.400	0.377	5.8	90	0.00
41	Benzo(g,h,i)perylene	0.400	0.366	8.5	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4