


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN051519\
 Data File : BN005694.D
 Acq On : 15 May 2019 08:27
 Operator : JU/SJ
 Sample : DFTPP69
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP69

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN050719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue May 14 07:15:06 2019



AutoFind: Scans 2455, 2456, 2457; Background Corrected with Scan 2448

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.4	116088	PASS
68	69	0.00	2	1.8	2152	PASS
69	198	0.00	100	31.9	117784	PASS
70	69	0.00	2	0.5	640	PASS
127	198	10	80	40.2	148528	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	369685	PASS
199	198	5	9	6.8	25277	PASS
275	198	10	60	28.1	104013	PASS
365	198	1	100	3.6	13248	PASS
441	198	0.01	100	13.5	49949	PASS
442	198	50	100	86.6	319978	PASS
443	442	15	24	19.2	61360	PASS

Scan 2448 (17.386 min): BN005694.D
DFTPP69

m/z	Abundance
39.90	1603.0
44.00	6016.0
207.00	1058.0

Instrument :
BNA_N
ClientSampleId :
DFTPP69

m/z	Abundance
37.10	306.0
38.00	1474.0
39.00	7881.0
40.00	1538.0
41.00	368.0
43.10	320.0
44.00	5601.0
44.90	428.0
49.00	1131.0
50.00	28824.0
51.00	115000.0
52.00	5535.0
52.90	403.0
54.90	736.0
55.90	3423.0
57.00	8301.0
57.90	441.0
61.00	1789.0
61.90	1996.0
63.00	4479.0
64.00	859.0
65.00	2478.0
68.00	2129.0
69.00	114512.0
70.00	607.0
72.90	1044.0
74.00	11872.0
75.00	17864.0
76.10	6035.0
77.00	121504.0
78.00	8322.0
79.00	8902.0
80.00	6691.0
81.00	9264.0
82.00	2372.0
82.90	2156.0
85.00	1731.0
85.90	2530.0
86.90	1196.0
88.10	303.0
91.00	2865.0
92.00	2589.0
93.00	15836.0
93.90	1042.0
95.00	365.0
96.00	888.0
97.00	591.0
98.00	11164.0
99.00	9372.0
99.90	984.0
101.00	5393.0
103.00	1679.0
104.00	3366.0
105.00	3304.0
106.10	1200.0
107.00	40536.0
108.00	6358.0
110.00	71808.0
111.00	11694.0
112.00	1340.0
113.00	317.0
116.00	2396.0
117.00	31512.0
118.00	2414.0
119.00	340.0
119.80	506.0
122.00	2831.0
122.90	3656.0
123.90	1617.0
125.00	1817.0
127.00	136448.0
128.00	11499.0
129.00	52248.0
130.00	5427.0
131.00	1015.0
132.10	448.0
133.00	564.0
134.00	1821.0
135.00	3989.0
136.00	1914.0
137.00	2158.0
138.10	549.0
138.80	418.0
140.00	869.0
141.00	7444.0
142.00	2873.0
143.00	1517.0
143.80	371.0
144.90	468.0
145.90	1460.0
146.90	2994.0
148.00	8231.0

Instrument :
BNA_N
ClientSampleId :
DFTPP69

148.90	2167.0
149.80	427.0
150.10	442.0
150.90	931.0
151.90	360.0
152.90	2857.0
154.00	1979.0
155.00	4013.0
156.00	5982.0
156.90	1088.0
158.00	1157.0
158.90	1144.0
159.90	2444.0
161.00	3105.0
162.00	1021.0
164.00	320.0
165.00	2758.0
166.00	2566.0
167.00	13723.0
168.00	6760.0
169.00	998.0
169.90	668.0
171.00	809.0
171.90	1484.0
173.00	1900.0
174.00	2998.0
175.00	6707.0
175.90	1362.0
177.00	2838.0
178.00	1117.0
179.00	10725.0
180.00	7647.0
181.00	4278.0
181.90	555.0
182.90	329.0
183.90	980.0
185.00	5772.0
186.00	45040.0
187.00	12872.0
188.00	1068.0
189.00	2868.0
189.90	634.0
190.90	1297.0
191.90	3714.0
193.00	4131.0
194.00	1091.0
196.00	8763.0
197.90	317888.0
199.00	22432.0
199.90	1810.0
200.90	552.0
201.60	1176.0
202.90	2362.0
204.00	10786.0
205.00	18424.0
206.00	84152.0
207.00	12398.0
208.00	2652.0
209.00	1200.0
210.10	1326.0
210.90	3195.0
211.80	731.0
215.00	754.0
216.00	1066.0
216.90	21464.0
218.00	2783.0
219.80	415.0
221.00	12342.0
223.00	4970.0
224.00	47624.0
225.00	11987.0
226.00	1586.0
227.00	18352.0
228.00	3193.0
229.00	4084.0
229.90	608.0
231.00	1655.0
231.90	408.0
233.00	434.0
233.90	1161.0
235.00	1563.0
236.00	875.0
236.90	1271.0
238.90	911.0
239.90	602.0
241.00	1109.0
242.10	2432.0
243.00	2515.0
244.00	35528.0
245.00	4724.0
246.00	6369.0
247.00	1604.0
248.90	1362.0
250.00	342.0
251.80	335.0
253.00	847.0

Instrument :
BNA_N
ClientSampleId :
DFTPP69

255.00	183808.0
256.00	27240.0
257.00	2501.0
258.00	11273.0
259.00	1561.0
260.00	494.0
265.00	3999.0
265.90	637.0
270.90	767.0
272.00	675.0
273.00	5868.0
274.00	14267.0
275.00	84384.0
276.00	10395.0
277.00	6682.0
278.00	893.0
279.10	404.0
280.90	316.0
283.00	884.0
284.10	569.0
284.90	1206.0
290.00	301.0
292.10	457.0
293.00	1715.0
293.90	339.0
295.00	829.0
296.00	21192.0
297.10	2966.0
297.90	331.0
301.00	468.0
301.90	582.0
303.00	2435.0
303.90	698.0
310.00	446.0
314.00	1129.0
315.00	2384.0
316.00	1498.0
321.10	407.0
322.00	411.0
323.00	6699.0
324.00	1310.0
326.90	1087.0
328.00	736.0
332.00	357.0
333.10	739.0
334.00	4728.0
335.10	1401.0
341.00	847.0
346.00	1834.0
347.10	322.0
352.00	2157.0
353.00	1419.0
354.00	2853.0
355.00	376.0
365.00	9632.0
365.90	1875.0
370.90	438.0
372.00	4025.0
373.00	910.0
383.00	1020.0
383.90	428.0
389.90	716.0
391.20	374.0
401.00	309.0
402.10	1574.0
403.00	2448.0
404.10	623.0
420.90	1774.0
422.10	1514.0
423.00	12306.0
424.00	2401.0
441.00	31952.0
442.00	213248.0
443.00	42472.0
444.00	4843.0

Instrument :
BNA_N
ClientSampleId :
DFTPP69

m/z	Abundance
36.90	727.0
38.10	1702.0
39.00	10124.0
40.00	1622.0
41.10	529.0
44.00	4869.0
44.90	421.0
49.20	958.0
50.00	35496.0
51.00	136320.0
52.00	6535.0
53.00	349.0
55.00	982.0
55.90	4548.0
57.00	8869.0
58.00	657.0
60.90	1837.0
62.00	2034.0
63.00	5151.0
64.10	882.0
65.00	2819.0
67.00	324.0
68.00	2496.0
69.00	143232.0
70.00	720.0
73.00	1313.0
74.00	14393.0
75.00	23016.0
76.00	7751.0
77.00	156544.0
78.00	9893.0
79.00	10994.0
80.00	8016.0
81.00	11309.0
82.00	3106.0
83.00	2617.0
85.00	2170.0
86.00	3148.0
87.10	1996.0
87.80	463.0
89.00	316.0
91.00	2497.0
92.00	3117.0
93.00	19256.0
93.90	1116.0
95.00	392.0
95.90	747.0
97.10	429.0
98.00	14268.0
99.00	12149.0
100.00	1106.0
101.00	6735.0
101.90	410.0
102.90	2207.0
104.00	4334.0
105.00	4180.0
106.10	1303.0
107.00	53616.0
108.00	8514.0
109.10	2004.0
110.00	91256.0
111.00	13646.0
112.00	1689.0
113.00	429.0
115.00	350.0
116.00	3687.0
117.00	38648.0
117.90	3348.0
118.90	529.0
120.10	636.0
121.90	3043.0
123.00	5227.0
124.00	2332.0
125.00	2426.0
127.00	178880.0
128.00	13961.0
129.00	71888.0
130.00	5929.0
131.00	927.0
131.90	751.0
133.90	2406.0
135.00	6240.0
136.00	2566.0
137.00	2861.0
137.80	542.0
138.20	541.0
139.00	358.0
139.90	1209.0
141.00	9910.0
141.90	2922.0
143.00	1983.0
143.90	681.0

Instrument :
BNA_N
ClientSampleId :
DFTPP69

144.90	346.0
146.00	2017.0
147.00	4498.0
148.00	9832.0
149.00	2653.0
150.00	679.0
150.90	1272.0
151.80	1019.0
153.00	3087.0
153.90	2658.0
155.00	5471.0
156.00	8931.0
157.00	1616.0
157.90	1811.0
158.90	1520.0
160.00	3099.0
161.00	4674.0
162.00	1216.0
162.70	328.0
163.10	372.0
164.00	503.0
164.90	3646.0
166.10	3547.0
167.00	19504.0
168.00	8950.0
169.00	1646.0
169.90	514.0
170.80	552.0
171.90	2023.0
173.00	2153.0
173.90	4616.0
175.00	8123.0
176.00	2562.0
177.00	4465.0
178.10	1340.0
179.00	14943.0
179.90	9543.0
181.00	4783.0
182.00	987.0
182.80	397.0
184.00	1267.0
185.00	8029.0
186.00	60552.0
187.00	17680.0
188.00	2006.0
189.00	3720.0
189.90	398.0
191.00	2163.0
192.00	4760.0
193.00	5463.0
194.00	1297.0
194.70	312.0
195.20	879.0
196.00	11390.0
197.90	439552.0
198.90	30848.0
200.00	2321.0
201.60	1571.0
203.00	3367.0
204.00	15406.0
205.00	27304.0
206.00	116464.0
207.00	16880.0
208.00	4387.0
208.90	1082.0
210.10	1755.0
211.00	4629.0
212.10	682.0
212.80	361.0
215.00	1383.0
216.00	2602.0
217.00	29096.0
218.00	4669.0
218.90	642.0
221.00	18424.0
221.80	4201.0
223.00	6743.0
224.00	65376.0
225.00	17664.0
226.10	2081.0
227.00	27400.0
228.00	4474.0
229.00	6165.0
230.00	1129.0
231.00	2321.0
232.00	421.0
232.90	513.0
234.00	1629.0
235.00	1514.0
235.90	1495.0
237.00	2326.0
239.10	1193.0
239.80	719.0
241.00	1531.0
242.00	4059.0
243.10	3530.0

Instrument :
BNA_N
ClientSampleId :
DFTPP69

244.00	51560.0
245.10	6130.0
246.00	10866.0
247.00	1801.0
248.10	471.0
249.00	1946.0
250.00	327.0
251.00	492.0
251.90	550.0
253.00	1198.0
255.00	273664.0
256.00	38544.0
257.00	3068.0
258.00	15153.0
259.00	2430.0
259.90	415.0
261.10	497.0
263.90	392.0
265.00	5870.0
266.00	857.0
269.90	307.0
271.00	1223.0
272.00	786.0
272.90	8118.0
274.00	20680.0
275.00	121984.0
276.00	14819.0
277.00	10214.0
278.00	1522.0
281.90	313.0
283.10	1241.0
283.90	690.0
285.00	1968.0
285.90	392.0
289.00	520.0
290.90	311.0
292.10	540.0
293.00	2384.0
293.80	582.0
294.90	870.0
296.00	32016.0
297.00	4943.0
300.90	350.0
301.90	646.0
303.00	4288.0
304.00	1223.0
307.90	584.0
309.90	350.0
312.80	386.0
314.00	1860.0
314.90	3392.0
316.00	2214.0
316.80	318.0
321.10	1042.0
322.10	493.0
323.00	11751.0
324.10	1568.0
326.90	1964.0
328.00	1094.0
332.00	1014.0
333.00	1095.0
334.00	7410.0
335.00	1840.0
336.00	360.0
341.00	1099.0
342.10	436.0
346.00	2155.0
347.00	537.0
352.00	3707.0
353.00	1831.0
354.00	3558.0
355.10	795.0
365.00	15196.0
365.90	2276.0
369.90	337.0
370.90	784.0
372.00	6356.0
373.10	1566.0
374.00	335.0
382.90	1586.0
384.00	507.0
390.00	610.0
391.00	911.0
392.00	445.0
400.90	357.0
402.00	2863.0
402.90	3795.0
404.00	1787.0
421.00	2917.0
422.10	2529.0
423.00	22856.0
424.00	4168.0
425.10	327.0
438.40	350.0
439.10	376.0
440.00	839.0

Instrument :
BNA_N
ClientSampleId :
DFTPP69

441.00	57544.0
442.00	375936.0
443.00	70984.0
444.10	7187.0
445.00	483.0

Instrument :
BNA_N
ClientSampleId :
DFTPP69

m/z	Abundance
37.00	429.0
38.00	1036.0
39.00	7191.0
39.90	1189.0
41.00	450.0
44.00	5702.0
44.90	319.0
48.90	664.0
50.00	23312.0
51.00	96944.0
52.00	4187.0
54.90	765.0
56.00	2443.0
57.00	5952.0
61.00	1475.0
62.00	1202.0
63.00	3480.0
64.10	725.0
65.00	1993.0
68.10	1831.0
69.00	95608.0
70.10	594.0
73.00	1035.0
74.00	10540.0
75.00	16728.0
76.00	4869.0
77.00	105440.0
78.00	7650.0
79.00	7788.0
79.90	5764.0
80.90	7680.0
81.90	1929.0
83.00	1773.0
84.10	419.0
85.00	1325.0
85.90	2030.0
87.00	1166.0
88.00	377.0
89.10	458.0
91.00	1859.0
92.00	2296.0
93.00	13518.0
94.00	1084.0
95.00	353.0
95.90	708.0
96.90	491.0
98.00	10754.0
99.00	8585.0
99.90	728.0
101.00	4626.0
103.00	1319.0
104.00	3128.0
105.00	2967.0
106.00	1019.0
107.00	35136.0
108.00	6293.0
108.90	1283.0
110.00	64704.0
111.00	10016.0
112.00	1391.0
113.00	501.0
114.90	393.0
115.90	2195.0
117.00	28968.0
118.00	1831.0
118.90	391.0
119.90	466.0
122.00	2713.0
122.90	3998.0
124.00	1732.0
125.00	1575.0
127.00	130256.0
128.00	9108.0
129.00	52248.0
129.90	4989.0
131.00	738.0
131.90	464.0
132.50	341.0
133.90	1897.0
135.00	4245.0
135.90	1626.0
137.00	2756.0
137.90	500.0
139.90	833.0
140.90	7054.0
142.00	1911.0
143.00	1520.0
144.10	430.0
144.90	306.0
146.00	1621.0
147.00	2864.0
148.00	7888.0

Instrument :
BNA_N
ClientSampleId :
DFTPP69

148.90	1967.0
149.90	330.0
151.00	925.0
151.70	574.0
152.00	440.0
152.90	2347.0
154.00	1911.0
155.00	3670.0
156.00	6025.0
157.10	1578.0
157.90	1296.0
158.90	1297.0
159.90	2169.0
161.00	3637.0
162.10	1011.0
163.00	381.0
163.80	518.0
164.90	4045.0
166.00	2096.0
167.00	13850.0
168.00	6996.0
169.00	1155.0
170.00	656.0
170.90	959.0
172.00	1287.0
173.00	1962.0
174.00	3001.0
175.00	6464.0
176.00	2004.0
176.90	3089.0
177.90	1120.0
179.00	11551.0
180.00	7920.0
181.00	3066.0
182.10	786.0
183.10	440.0
183.80	897.0
185.00	5404.0
186.00	47032.0
187.00	13453.0
188.00	1284.0
188.90	3241.0
190.00	362.0
191.00	1804.0
192.00	3710.0
193.00	4665.0
194.00	1152.0
195.10	492.0
196.00	8512.0
198.00	351616.0
199.00	22552.0
200.00	1615.0
201.40	1501.0
203.00	2601.0
204.00	12674.0
205.00	21408.0
206.00	91056.0
207.00	13325.0
208.00	3653.0
209.00	1256.0
210.10	2004.0
210.90	3266.0
214.90	1398.0
216.10	1968.0
216.90	23904.0
218.00	3227.0
219.00	430.0
220.20	782.0
221.00	13903.0
222.00	3055.0
223.00	5882.0
224.00	52984.0
225.00	12299.0
227.00	23264.0
228.00	3317.0
229.00	5455.0
229.90	664.0
230.90	2435.0
232.00	347.0
232.90	391.0
234.00	1201.0
234.90	1475.0
235.90	1071.0
237.00	1846.0
239.00	972.0
239.90	824.0
240.90	1191.0
242.00	2760.0
243.00	2616.0
244.00	38144.0
245.00	4762.0
246.00	8242.0
246.90	1863.0
249.00	1493.0
251.00	370.0
251.80	383.0

Instrument :
BNA_N
ClientSampleId :
DFTPP69

253.00	1065.0
255.00	229056.0
256.00	32912.0
257.00	2197.0
258.00	12542.0
259.00	1776.0
259.90	337.0
261.00	395.0
264.00	487.0
265.00	5626.0
265.90	686.0
270.10	339.0
271.00	1025.0
271.90	439.0
273.00	7477.0
274.00	19296.0
275.00	105672.0
276.00	14024.0
277.00	8655.0
278.00	1194.0
283.00	869.0
284.00	927.0
285.00	1761.0
288.90	342.0
289.80	321.0
290.90	446.0
292.00	517.0
292.90	1962.0
293.80	427.0
296.00	27256.0
297.00	3869.0
302.00	586.0
303.00	3563.0
304.00	1025.0
308.00	462.0
310.00	421.0
314.00	1405.0
315.00	3276.0
316.00	1721.0
317.00	409.0
321.10	1297.0
321.90	585.0
323.00	10655.0
324.10	2118.0
326.00	318.0
327.10	1349.0
328.00	1294.0
331.80	525.0
333.00	980.0
334.00	6633.0
335.00	1729.0
341.10	959.0
342.00	596.0
345.90	2145.0
347.10	470.0
351.90	3019.0
353.00	2065.0
354.00	3651.0
355.00	766.0
365.00	14918.0
366.00	2115.0
370.10	328.0
370.90	831.0
372.00	5726.0
373.00	1672.0
382.90	1219.0
384.00	388.0
389.90	722.0
390.90	398.0
391.80	427.0
401.10	387.0
402.00	2202.0
403.00	3269.0
404.00	1593.0
420.90	3179.0
422.00	3060.0
423.00	22216.0
424.10	4342.0
425.00	511.0
438.80	420.0
439.40	329.0
441.00	60352.0
442.00	370752.0
443.00	70624.0
444.00	6402.0
445.10	609.0

Instrument :
BNA_N
ClientSampleId :
DFTPP69