

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN052120\
 Data File : BN010731.D
 Acq On : 21 May 2020 16:21
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 21 16:56:27 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN051320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 21 16:54:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.345	0.349	-1.2	101	0.00
3	n-Nitrosodimethylamine	0.199	0.190	4.5	97	0.00
4 S	2-Fluorophenol	0.783	0.780	0.4	95	0.00
5 S	Phenol-d6	0.957	0.931	2.7	94	0.00
6	bis(2-Chloroethyl)ether	0.986	0.959	2.7	88	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.278	0.268	3.6	94	0.00
9	Naphthalene	1.006	0.989	1.7	95	0.00
10	Hexachlorobutadiene	0.230	0.225	2.2	92	0.00
11 SURR	2-Methylnaphthalene-d10	0.652	0.634	2.8	94	0.00
12	2-Methylnaphthalene	0.709	0.694	2.1	94	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.145	0.127	12.4	94	0.00
15 S	2-Fluorobiphenyl	1.468	1.455	0.9	94	0.00
16	Acenaphthylene	1.634	1.500	8.2	93	0.00
17	Acenaphthene	1.104	1.090	1.3	94	0.00
18	Fluorene	1.447	1.395	3.6	95	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.221	0.213	3.6	94	0.00
21	Hexachlorobenzene	0.263	0.264	-0.4	96	0.00
22	Pentachlorophenol	0.081	0.082	-1.2	112	0.00
23	Phenanthrene	1.073	1.044	2.7	97	0.00
24	Anthracene	0.902	0.834	7.5	96	0.00
25 SURR	Fluoranthene-d10	1.085	1.044	3.8	100	0.00
26	Fluoranthene	1.219	1.164	4.5	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
28	Pvrene	1.434	1.434	0.0	100	0.00
29 S	Terphenyl-d14	0.949	0.958	-0.9	102	0.00
30	Benzo(a)anthracene	1.167	1.111	4.8	105	0.00
31	Chrysene	1.321	1.338	-1.3	103	0.00
32	Bis(2-ethylhexyl)phthalate	0.511	0.482	5.7	103	0.00
33	Indeno(1,2,3-cd)pyrene	1.175	1.091	7.1	101	0.00
34 I	Perylene-d12	1.000	1.000	0.0	103	0.00
35	Benzo(b)fluoranthene	1.322	1.260	4.7	104	0.00
36	Benzo(k)fluoranthene	1.424	1.463	-2.7	109	0.00
37 C	Benzo(a)pyrene	1.144	1.110	3.0	103	0.00
38	Dibenzo(a,h)anthracene	0.978	0.901	7.9	103	0.00
39	Benzo(a,h,i)perylene	1.147	1.089	5.1	103	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN052120\
Data File : BN010731.D
Acq On : 21 May 2020 16:21
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: May 21 16:56:27 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN051320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 21 16:54:53 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			