

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN052120\
 Data File : BN010739.D
 Acq On : 21 May 2020 22:22
 Operator : CG/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: May 22 00:50:45 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN051320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 21 16:54:53 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
2	1,4-Dioxane	0.345	0.331	4.1	114	0.00
3	n-Nitrosodimethylamine	0.199	0.186	6.5	112	0.00
4 S	2-Fluorophenol	0.783	0.817	-4.3	118	0.00
5 S	Phenol-d6	0.957	0.997	-4.2	119	0.00
6	bis(2-Chloroethyl)ether	0.986	1.030	-4.5	113	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
8 S	Nitrobenzene-d5	0.278	0.276	0.7	119	0.00
9	Naphthalene	1.006	0.997	0.9	118	0.00
10	Hexachlorobutadiene	0.230	0.223	3.0	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.652	0.648	0.6	118	0.00
12	2-Methylnaphthalene	0.709	0.710	-0.1	119	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	121	-0.01
14 S	2,4,6-Tribromophenol	0.145	0.130	10.3	121	0.00
15 S	2-Fluorobiphenyl	1.468	1.456	0.8	118	0.00
16	Acenaphthylene	1.634	1.539	5.8	120	0.00
17	Acenaphthene	1.104	1.086	1.6	118	0.00
18	Fluorene	1.447	1.405	2.9	121	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	118	-0.01
20	4-Bromophenyl-phenylether	0.221	0.222	-0.5	120	-0.01
21	Hexachlorobenzene	0.263	0.264	-0.4	118	0.00
22	Pentachlorophenol	0.081	0.083	-2.5	139	0.00
23	Phenanthrene	1.073	1.072	0.1	122	0.00
24	Anthracene	0.902	0.845	6.3	119	0.00
25 SURR	Fluoranthene-d10	1.085	1.021	5.9	120	0.00
26	Fluoranthene	1.219	1.145	6.1	119	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	117	0.00
28	Pvrene	1.434	1.485	-3.6	118	0.00
29 S	Terphenyl-d14	0.949	1.005	-5.9	122	0.00
30	Benzo(a)anthracene	1.167	1.141	2.2	123	-0.01
31	Chrysene	1.321	1.305	1.2	114	0.00
32	Bis(2-ethylhexyl)phthalate	0.511	0.502	1.8	122	0.00
33	Indeno(1,2,3-cd)pyrene	1.175	1.080	8.1	114	0.00
34 I	Perylene-d12	1.000	1.000	0.0	120	0.00
35	Benzo(b)fluoranthene	1.322	1.226	7.3	117	0.00
36	Benzo(k)fluoranthene	1.424	1.426	-0.1	124	0.00
37 C	Benzo(a)pyrene	1.144	1.107	3.2	119	0.00
38	Dibenzo(a,h)anthracene	0.978	0.854	12.7	114	0.00
39	Benzo(a,h,i)perylene	1.147	1.006	12.3	110	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN052120\
Data File : BN010739.D
Acq On : 21 May 2020 22:22
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: May 22 00:50:45 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN051320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 21 16:54:53 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			