

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN052219\
 Data File : BN005868.D
 Acq On : 23 May 2019 01:48
 Operator : JU/SJ
 Sample : K2927-04
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A4295

Manual Integrations
 APPROVED

mohammad
 5/23/2019 3:31:15 PM

Quant Time: May 23 06:49:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN051819MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 22 04:04:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	64048	20.00	ng/ul	0.00
18) Naphthalene-d8	10.32	136	287760	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.20	164	203503	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.96	188	556681	20.00	ng/ul	0.00
77) Chrysene-d12	21.19	240	707809	20.00	ng/ul	0.00
85) Perylene-d12	23.40	264	842539	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.18	96	5327	4.66	ng/uL	0.00
5) Phenol-d5	6.76	99	93297	18.85	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.91	67	58233	22.55	ng/ul	0.00
9) 2-Chlorophenol-d4	7.10	132	81199	20.40	ng/ul	0.00
13) 4-Methylphenol-d8	8.29	113	63811	15.08	ng/ul	0.01
19) Nitrobenzene-d5	8.72	128	41119	19.68	ng/ul	0.00
22) 2-Nitrophenol-d4	9.43	143	44137	18.30	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.99	165	87584m	19.02	ng/ul	0.02
29) 4-Chloroaniline-d4	10.49	131	96505m	18.88	ng/ul	0.02
43) Dimethylphthalate-d6	13.62	166	378691	24.18	ng/ul	0.00
46) Acenaphthylene-d8	13.89	160	440488	23.15	ng/ul	0.00
51) 4-Nitrophenol-d4	14.49	143	32200m	14.60	ng/ul	0.03
57) Fluorene-d10	15.20	176	339425	25.24	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.38	200	28615	8.55	ng/ul	0.01
70) Anthracene-d10	17.06	188	586190	24.37	ng/ul	0.00
78) Pyrene-d10	19.37	212	789218	22.63	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.26	264	927788	24.55	ng/ul	0.00

Target Compounds

					Ovalue
44) Dimethylphthalate	13.67	163	127109	8.233 ng/ul#	98
76) Fluoranthene	19.04	202	62284	1.826 ng/ul#	88
79) Pyrene	19.41	202	62471	1.437 ng/ul#	83
87) Benzo(b)fluoranthene	22.75	252	57819	1.247 ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN052219\
Data File : BN005868.D
Acq On : 23 May 2019 01:48
Operator : JU/SJ
Sample : K2927-04
Misc :
ALS Vial : 17 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
A4295

Manual Integrations
APPROVED

mohammad
5/23/2019 3:31:15 PM

Quant Time: May 23 06:49:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN051819MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed May 22 04:04:33 2019
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 8.233 ng/ul

RT: 13.67 min Scan# 1816

Delta R.T. 0.01 min

Lab File: BN005868.D

Acq: 23 May 2019 01:48

Instrument :

BNA_N


ClientSampled :

A4295




Tot Ion:	163	Resp:	127109
Ion Ratio	Lower	Upper	
163	100		
194	5.2	3.4	5.0#
164	9.5	8.1	12.1

Manual Integrations
APPROVED

mohammad
5/23/2019 3:31:15 PM



Abundance Ion 163.00 (162.70 to 163.70): E
100000 Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E



#76

Fluoranthene

Concen: 1.826 ng/ul


RT: 19.04 min Scan# 2730

Delta R.T. 0.01 min



Lab File: BN005868.D


Acq: 23 May 2019 01:48

Tgt Ion:	202	Resp:	62284
Ion Ratio	Lower	Upper	
202	100		
101	4.1	8.6	12.8#
100	6.0	6.3	9.5#



Abundance Ion 202.00 (201.70 to 202.70): E
60000 Ion 101.00 (100.70 to 101.70): E
Ion 100.00 (99.70 to 100.70): B









#79
Pvrene
Concen: 1.437 ng/ul
RT: 19.41 min Scan# 2792
Delta R.T. 0.01 min
Lab File: BN005868.D
Acq: 23 May 2019 01:48

Instrument :
BNA_N
ClientSampleId :
A4295

Tot Ion	Ratio	Resp Lower	Upper
202	100		
101	6.2	10.3	15.5#
100	3.8	7.9	11.9#

Manual Integrations
APPROVED

mohammad
5/23/2019 3:31:15 PM



#87
Benzo(b)fluoranthene
Concen: 1.247 ng/ul
RT: 22.75 min Scan# 3360
Delta R.T. 0.01 min
Lab File: BN005868.D
Acq: 23 May 2019 01:48

Tgt Ion	Ratio	Resp Lower	Upper
252	100		
253	24.6	17.7	26.5
125	5.0	8.4	12.6#

