

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052421\
 Data File : BN014714.D
 Acq On : 24 May 2021 10:55
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 24 12:11:43 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN051121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 21 11:30:28 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	176	0.00
2	1,4-Dioxane	0.400	0.414	-3.5	212	0.00
3	n-Nitrosodimethylamine	0.400	0.306	23.5	116	-0.02
4 S	2-Fluorophenol	0.400	0.433	-8.2	207	0.00
5 S	Phenol-d6	0.400	0.454	-13.5	217	0.00
6	bis(2-Chloroethyl)ether	0.400	0.467	-16.8	224	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	173	0.00
8 S	Nitrobenzene-d5	0.400	0.444	-11.0	216	0.00
9	Naphthalene	0.400	0.421	-5.2	197	0.00
10	Hexachlorobutadiene	0.400	0.465	-16.3	216	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.428	-7.0	202	0.00
12	2-Methylnaphthalene	0.400	0.419	-4.7	198	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	170	0.00
14 S	2,4,6-Tribromophenol	0.400	0.453	-13.2	242	0.00
15 S	2-Fluorobiphenyl	0.400	0.436	-9.0	202	0.00
16	Acenaphthylene	0.400	0.417	-4.2	205	0.00
17	Acenaphthene	0.400	0.395	1.3	185	0.00
18	Fluorene	0.400	0.416	-4.0	195	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	159	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.389	2.8	169	0.00
21	4-Bromophenyl-phenylether	0.400	0.450	-12.5	200	0.00
22	Hexachlorobenzene	0.400	0.473	-18.2	212	0.00
23	Atrazine	0.400	0.393	1.8	188	0.00
24	Pentachlorophenol	0.400	0.339	15.3	171	0.00
25	Phenanthrene	0.400	0.408	-2.0	184	0.00
26	Anthracene	0.400	0.416	-4.0	191	0.00
27 SURR	Fluoranthene-d10	0.400	0.392	2.0	179	0.00
28	Fluoranthene	0.400	0.393	1.8	179	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	134	0.00
30	Pyrene	0.400	0.465	-16.3	175	0.00
31 S	Terphenyl-d14	0.400	0.471	-17.7	168	0.00
32	Benzo(a)anthracene	0.400	0.418	-4.5	159	0.00
33	Chrysene	0.400	0.438	-9.5	164	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.429	-7.2	170	0.00
35 I	Perylene-d12	0.400	0.400	0.0	125	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.386	3.5	136	0.00
37	Benzo(b)fluoranthene	0.400	0.428	-7.0	151	0.00
38	Benzo(k)fluoranthene	0.400	0.410	-2.5	143	0.00
39 C	Benzo(a)pyrene	0.400	0.440	-10.0	159	0.00
40	Dibenzo(a,h)anthracene	0.400	0.381	4.8	133	0.01
41	Benzo(g,h,i)perylene	0.400	0.403	-0.8	144	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4