

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052521\
 Data File : BN014731.D
 Acq On : 25 May 2021 12:18
 Operator : CG/JU
 Sample : M2493-15
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GB432

Quant Time: May 25 12:53:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN052521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 25 10:33:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.710	152	200636	20.00	ng/ul	0.00
20) Naphthalene-d8	10.499	136	823660	20.00	ng/ul	# 0.00
38) Acenaphthene-d10	14.363	164	622153	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.116	188	1535273	20.00	ng/ul	0.00
79) Chrysene-d12	21.316	240	2036226	20.00	ng/ul	0.00
88) Perylene-d12	23.586	264	2361666	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.228	96	20759	4.21	ng/uL	0.00
4) Pyridine-d5	0.000	84	0d	0.00	ng/ul	
7) Phenol-d5	6.893	99	124820	6.35	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.046	67	281882	28.85	ng/ul	0.00
11) 2-Chlorophenol-d4	7.252	132	296895	24.40	ng/ul	0.00
15) 4-Methylphenol-d8	8.416	113	219701	14.38	ng/ul	0.00
21) Nitrobenzene-d5	8.863	128	187040	30.15	ng/ul	0.00
24) 2-Nitrophenol-d4	9.581	143	187076	28.05	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.128	165	401711	26.42	ng/ul	0.00
31) 4-Chloroaniline-d4	10.646	131	28548	1.59	ng/ul	0.01
46) Dimethylphthalate-d6	13.775	166	1673004	33.85	ng/ul	0.00
49) Acenaphthylene-d8	14.051	160	1822327	31.88	ng/ul	0.00
54) 4-Nitrophenol-d4	14.581	143	40801	5.20	ng/ul	0.00
60) Fluorene-d10	15.363	176	1495286	35.11	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.475	200	267910	31.90	ng/ul	0.00
73) Anthracene-d10	17.216	188	2828989	39.45	ng/ul	0.00
81) Pyrene-d10	19.522	212	3616251	37.66	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.439	264	4728339	37.59	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052521\
Data File : BN014731.D
Acq On : 25 May 2021 12:18
Operator : CG/JU
Sample : M2493-15
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GB432

Quant Time: May 25 12:53:13 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN052521.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 25 10:33:04 2021
Response via : Initial Calibration

