

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052521\
 Data File : BN014747.D
 Acq On : 25 May 2021 22:25
 Operator : CG/JU
 Sample : M2493-07
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GB422

Manual Integrations
 APPROVED

mohammad
 5/26/2021 4:27:34 PM

Quant Time: May 26 01:06:01 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN052521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 25 10:33:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	225336	20.00	ng/ul	0.00
20) Naphthalene-d8	10.499	136	873120	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.363	164	686590	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.116	188	1654960	20.00	ng/ul	0.00
79) Chrysene-d12	21.310	240	2116472	20.00	ng/ul	# 0.00
88) Perylene-d12	23.574	264	2539825	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.228	96	22111	4.00	ng/uL	0.00
4) Pyridine-d5	0.000	84	0d	0.00	ng/ul	
7) Phenol-d5	6.887	99	156317	7.08	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.046	67	323104	29.45	ng/ul	0.00
11) 2-Chlorophenol-d4	7.252	132	354728	25.96	ng/ul	0.00
15) 4-Methylphenol-d8	8.416	113	267813	15.61	ng/ul	0.00
21) Nitrobenzene-d5	8.863	128	210033	31.93	ng/ul	0.00
24) 2-Nitrophenol-d4	9.587	143	227870	32.23	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.122	165	464926	28.84	ng/ul	0.00
31) 4-Chloroaniline-d4	10.634	131	11979m	0.63	ng/ul	0.00
46) Dimethylphthalate-d6	13.775	166	1860069	34.10	ng/ul	0.00
49) Acenaphthylene-d8	14.051	160	2050278	32.50	ng/ul	0.00
54) 4-Nitrophenol-d4	14.581	143	46931m	5.42	ng/ul	0.00
60) Fluorene-d10	15.357	176	1650865	35.12	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.475	200	294444	32.53	ng/ul	0.00
73) Anthracene-d10	17.216	188	2911585	37.66	ng/ul	0.00
81) Pyrene-d10	19.516	212	3658616	36.66	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.439	264	4931825	36.45	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052521\
Data File : BN014747.D
Acq On : 25 May 2021 22:25
Operator : CG/JU
Sample : M2493-07
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GB422

Manual Integrations
APPROVED

mohammad
5/26/2021 4:27:34 PM

Quant Time: May 26 01:06:01 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN052521.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 25 10:33:04 2021
Response via : Initial Calibration

