


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052521\
 Data File : BN014726.D
 Acq On : 25 May 2021 09:04
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP042

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN052521.M
 Title : SVOA CALIBRATION
 Last Update : Tue May 25 10:33:04 2021



AutoFind: Scans 2461, 2462, 2463; Background Corrected with Scan 2456

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	18.9	111720	PASS
68	69	0.00	2	1.0	1449	PASS
69	198	0.00	100	25.4	150189	PASS
70	69	0.00	2	0.6	921	PASS
127	198	10	80	27.1	160611	PASS
197	198	0.00	2	0.5	2821	PASS
198	198	100	100	100.0	592213	PASS
199	198	5	9	7.1	42005	PASS
275	198	10	60	26.6	157587	PASS
365	198	1	100	2.5	14741	PASS
441	198	0.01	100	9.7	57373	PASS
442	198	50	100	58.8	348373	PASS
443	442	15	24	19.1	66693	PASS

m/z	Abundance
39.80	325.0
43.90	479.0
49.00	915.0
57.10	713.0
73.20	1189.0
77.90	447.0
83.80	889.0
85.10	445.0
85.60	411.0
90.90	408.0
113.00	743.0
133.00	950.0
207.10	700.0
207.90	764.0
280.80	859.0

Instrument :
BNA_N
ClientSampleId :
DFTPP042

m/z	Abundance
37.00	944.0
37.80	1006.0
39.10	4319.0
39.90	652.0
40.90	821.0
42.80	329.0
43.90	649.0
45.30	380.0
48.90	1182.0
50.00	25256.0
51.10	91424.0
52.00	5277.0
54.70	540.0
55.80	1458.0
56.20	2615.0
56.90	6008.0
57.70	408.0
60.90	1299.0
62.00	2198.0
63.00	5549.0
63.90	595.0
65.00	1935.0
66.90	319.0
68.00	3333.0
69.00	111112.0
70.00	783.0
70.80	402.0
72.00	308.0
73.00	2403.0
74.00	15717.0
75.00	20880.0
75.90	7043.0
77.00	131584.0
78.00	9876.0
79.00	12104.0
79.90	8414.0
81.00	10228.0
81.90	2457.0
83.10	1552.0
83.80	1135.0
85.10	2252.0
86.00	4671.0
87.00	1152.0
88.80	479.0
91.00	3679.0
92.00	3215.0
93.00	18864.0
93.90	1090.0
95.00	840.0
96.10	510.0
96.90	405.0
97.90	12968.0
98.90	14632.0
100.00	2030.0
100.90	6337.0
101.90	1059.0
103.00	2757.0
104.00	3874.0
105.00	3791.0
105.80	721.0
107.00	44968.0
108.10	7558.0
109.10	1880.0
110.00	102936.0
111.00	13950.0
112.00	1946.0
113.00	475.0
114.80	383.0
116.00	3131.0
117.00	40816.0
118.10	4483.0
118.70	474.0
119.00	518.0
119.90	940.0
120.90	816.0
121.90	3291.0
123.00	3310.0
124.00	1327.0
124.90	2182.0
127.00	123304.0
128.00	10412.0
129.00	49504.0
129.90	3609.0
131.00	1239.0
132.10	786.0
132.90	562.0
133.70	1292.0
134.10	1179.0
135.00	6419.0
136.00	2062.0
137.00	2809.0
138.30	1491.0

Instrument :
BNA_N
ClientSampleId :
DFTPP042

139.10	675.0
139.90	468.0
140.90	7941.0
142.00	2209.0
143.00	2129.0
143.90	635.0
146.10	1382.0
146.90	4514.0
148.00	9817.0
149.20	2192.0
150.80	525.0
151.10	670.0
151.90	581.0
153.10	2078.0
153.90	2605.0
155.10	4292.0
156.00	6990.0
157.20	1670.0
157.90	699.0
159.00	1153.0
159.90	2521.0
160.90	3722.0
162.00	588.0
165.00	3109.0
166.10	2057.0
166.90	16312.0
168.00	10831.0
168.90	1018.0
171.00	865.0
172.00	1944.0
172.90	1385.0
174.00	4298.0
175.10	7590.0
175.90	1442.0
176.90	3560.0
177.80	1583.0
179.00	8901.0
179.90	8992.0
181.10	3355.0
182.00	972.0
183.10	747.0
183.60	910.0
184.90	5798.0
186.00	48984.0
187.10	13145.0
188.00	1100.0
188.80	5141.0
189.90	1229.0
191.10	2148.0
191.90	3992.0
193.00	4656.0
193.90	786.0
196.00	8438.0
197.00	2681.0
197.90	429120.0
199.00	26792.0
199.90	1417.0
201.00	1176.0
201.60	1517.0
203.10	3263.0
204.10	11727.0
205.00	18024.0
206.00	79120.0
207.00	13789.0
208.00	4329.0
208.90	705.0
209.70	906.0
210.40	799.0
211.00	3485.0
213.00	1013.0
214.90	1535.0
217.00	18608.0
218.00	2084.0
218.80	671.0
221.00	10823.0
222.00	3223.0
223.10	4656.0
224.00	40632.0
225.00	12507.0
226.10	2961.0
227.00	23904.0
228.00	2574.0
228.90	3535.0
230.10	504.0
230.80	413.0
231.30	1094.0
233.90	2260.0
235.00	1408.0
235.70	398.0
237.00	3876.0
238.10	507.0
239.00	964.0
240.00	830.0
240.70	974.0
242.00	3049.0
243.00	2215.0

Instrument :
BNA_N
ClientSampleId :
DFTPP042

244.00	40584.0
245.00	5221.0
246.00	7040.0
247.00	1171.0
249.00	1146.0
249.80	562.0
251.00	350.0
252.90	1023.0
253.70	1437.0
255.00	196224.0
256.00	32280.0
257.10	2860.0
258.00	8967.0
258.80	2374.0
259.90	330.0
262.00	378.0
263.80	572.0
264.10	698.0
265.00	4643.0
266.00	1480.0
270.00	480.0
272.10	1036.0
273.00	4244.0
274.00	17280.0
275.00	97400.0
276.00	13544.0
277.00	9862.0
277.80	1802.0
281.20	871.0
283.10	877.0
283.80	945.0
285.00	1512.0
285.80	342.0
291.10	339.0
292.90	1467.0
293.20	1488.0
293.70	583.0
294.20	528.0
294.90	613.0
296.00	22824.0
297.00	3625.0
297.70	312.0
303.10	2910.0
304.00	873.0
308.00	629.0
314.10	1313.0
315.00	1989.0
316.00	2080.0
321.00	818.0
322.00	652.0
323.00	5677.0
324.10	2353.0
326.90	1809.0
327.90	1758.0
331.70	347.0
333.90	6396.0
334.90	1587.0
340.70	834.0
341.20	740.0
345.90	1239.0
346.90	507.0
352.10	3079.0
352.80	1529.0
354.10	2356.0
363.10	475.0
365.10	8167.0
365.70	1123.0
370.30	474.0
371.00	416.0
372.10	3778.0
373.10	579.0
384.30	699.0
389.70	347.0
390.80	574.0
392.00	325.0
400.90	331.0
402.10	504.0
402.90	2964.0
403.90	1563.0
421.10	2258.0
421.90	1827.0
423.00	14174.0
424.00	5689.0
441.00	26704.0
442.10	173952.0
443.00	39352.0
443.80	3113.0
445.00	353.0

Instrument :
BNA_N
ClientSampleId :
DFTPP042

m/z	Abundance
37.80	1435.0
39.10	8360.0
40.00	980.0
41.00	456.0
43.90	314.0
45.20	363.0
49.10	2195.0
50.00	35736.0
51.00	124184.0
52.00	7877.0
52.80	334.0
55.00	798.0
56.00	3613.0
56.90	8650.0
58.10	616.0
61.10	1969.0
62.00	2746.0
63.00	8361.0
64.10	1218.0
65.00	2633.0
67.00	494.0
69.00	170752.0
69.90	870.0
73.10	2175.0
74.00	23248.0
75.00	28600.0
76.00	9519.0
77.00	201792.0
78.00	16680.0
79.00	18648.0
80.10	10241.0
81.00	19184.0
81.90	4167.0
83.00	4055.0
83.90	380.0
85.00	3686.0
85.90	4707.0
86.90	3079.0
88.20	1671.0
88.80	457.0
90.90	4032.0
92.00	4015.0
93.00	23712.0
93.80	2530.0
95.60	365.0
96.10	461.0
97.10	1339.0
98.00	23488.0
99.00	20032.0
100.00	1788.0
101.00	10815.0
103.00	2889.0
103.90	5878.0
105.00	6727.0
106.00	3652.0
107.00	77472.0
108.00	9330.0
109.00	3485.0
110.00	156096.0
111.00	19680.0
111.90	2518.0
115.80	3340.0
117.00	60320.0
117.90	5031.0
119.10	578.0
119.70	1216.0
120.10	1011.0
121.00	539.0
121.90	3499.0
122.90	4612.0
124.00	2765.0
125.00	3869.0
127.00	191808.0
127.90	14112.0
129.00	71656.0
130.00	7050.0
131.00	1581.0
131.60	309.0
132.10	853.0
132.70	578.0
133.90	3506.0
135.00	7901.0
135.90	4191.0
137.00	5901.0
138.30	953.0
139.00	356.0
140.00	1690.0
141.00	11339.0
141.90	4119.0
142.90	2240.0
143.90	578.0
144.90	1195.0

Instrument :
BNA_N
ClientSampleId :
DFTPP042

145.80	2060.0
146.90	4446.0
148.00	13511.0
149.00	2424.0
150.00	669.0
151.00	1376.0
152.00	808.0
152.90	4225.0
154.00	2401.0
155.00	6345.0
156.00	10035.0
156.90	1234.0
157.90	2725.0
158.90	2737.0
159.80	3292.0
160.90	5538.0
162.00	3259.0
163.10	414.0
163.60	766.0
164.10	552.0
164.90	5474.0
166.00	4285.0
167.00	25280.0
168.00	14605.0
169.00	2858.0
170.10	772.0
171.10	1495.0
171.80	2146.0
173.10	2491.0
173.90	5034.0
175.00	8068.0
175.90	2267.0
176.90	5698.0
178.00	1687.0
179.00	17536.0
180.00	14573.0
180.90	8978.0
181.90	1552.0
183.10	577.0
183.80	1084.0
184.20	1086.0
185.00	10344.0
186.00	74552.0
187.00	24640.0
188.00	2737.0
189.00	4878.0
189.80	1769.0
190.90	3639.0
192.00	5229.0
193.10	7612.0
194.00	2142.0
195.00	600.0
196.00	12430.0
197.10	5782.0
197.90	680192.0
199.00	46712.0
200.10	2424.0
201.00	976.0
201.50	1891.0
201.90	1172.0
202.90	3393.0
204.00	22024.0
205.00	37256.0
206.00	142144.0
207.00	20128.0
208.00	3556.0
209.00	1245.0
209.70	1025.0
211.00	6430.0
211.90	1276.0
212.90	689.0
214.10	481.0
215.00	3169.0
216.90	31952.0
218.10	3466.0
218.90	1157.0
219.70	396.0
220.90	25304.0
222.10	3950.0
223.20	8152.0
224.00	86072.0
225.10	18448.0
226.00	3176.0
227.00	34096.0
228.00	6077.0
229.00	5493.0
229.90	2068.0
230.90	2436.0
232.70	905.0
234.00	2817.0
234.70	1819.0
235.90	918.0
237.00	6034.0
238.00	794.0
238.90	1184.0
240.00	1464.0

Instrument :
BNA_N
ClientSampleId :
DFTPP042

241.00	2621.0
242.10	4160.0
244.00	67424.0
245.00	8114.0
245.90	14584.0
246.90	3401.0
248.00	697.0
249.00	3941.0
249.90	752.0
250.90	812.0
251.80	1126.0
252.90	1654.0
255.00	327168.0
256.00	52112.0
257.00	5939.0
258.00	16928.0
259.00	2482.0
260.00	787.0
261.00	994.0
263.80	381.0
265.10	5646.0
266.00	946.0
267.00	430.0
270.00	692.0
271.20	305.0
272.00	856.0
273.00	9619.0
274.00	34648.0
275.00	173824.0
276.00	23088.0
277.00	13667.0
278.00	3343.0
280.90	423.0
283.00	1548.0
283.90	1075.0
284.30	745.0
285.10	2962.0
286.10	700.0
288.80	625.0
290.00	997.0
291.10	554.0
291.80	688.0
293.20	2808.0
293.80	1096.0
294.20	1232.0
295.00	1702.0
296.00	38344.0
297.10	5057.0
300.90	435.0
302.00	974.0
303.00	5642.0
304.20	1179.0
304.80	527.0
309.00	952.0
310.20	659.0
311.10	449.0
313.00	622.0
314.00	2477.0
314.90	4875.0
315.90	3938.0
316.90	714.0
321.00	682.0
322.00	823.0
323.10	14848.0
323.90	3467.0
325.80	529.0
327.00	3218.0
327.90	1875.0
328.90	662.0
331.90	1148.0
333.10	2426.0
334.00	11290.0
335.10	3338.0
336.00	1167.0
339.30	507.0
341.00	2261.0
341.80	577.0
342.30	456.0
346.00	1355.0
347.00	1295.0
351.00	516.0
352.00	2033.0
353.00	2839.0
354.10	6392.0
354.90	1762.0
365.00	17944.0
365.90	2401.0
369.70	361.0
370.30	952.0
371.00	2003.0
372.00	9918.0
373.00	2723.0
382.80	2157.0
384.10	642.0
389.70	859.0
390.00	1079.0

Instrument :
BNA_N
ClientSampleId :
DFTPP042

390.90	875.0
401.00	599.0
402.00	2848.0
403.00	5190.0
403.90	1820.0
405.10	422.0
420.90	3681.0
422.10	2577.0
423.00	27896.0
424.00	8851.0
425.10	1477.0
425.80	443.0
439.80	684.0
441.00	61552.0
442.00	358464.0
443.10	70240.0
444.00	5784.0
445.00	515.0

Instrument :
BNA_N
ClientSampleId :
DFTPP042

m/z	Abundance
35.50	319.0
36.80	526.0
38.00	1331.0
39.00	6207.0
40.10	700.0
42.60	435.0
43.10	367.0
44.00	734.0
47.40	441.0
48.90	1530.0
50.00	31856.0
51.00	119552.0
52.00	3719.0
55.10	1235.0
55.90	3651.0
57.00	7861.0
59.10	367.0
61.00	1878.0
62.00	2473.0
63.10	5357.0
64.10	854.0
65.00	2459.0
67.00	609.0
68.20	4346.0
69.00	168704.0
69.90	1109.0
70.90	771.0
72.90	2807.0
74.00	18856.0
75.00	27336.0
76.10	10234.0
77.00	190080.0
78.00	12138.0
79.00	16223.0
80.00	11694.0
80.90	14131.0
82.10	3927.0
83.00	2675.0
84.00	1014.0
85.10	2964.0
86.00	6449.0
86.90	2863.0
88.10	1037.0
91.00	3764.0
91.90	4241.0
93.00	26704.0
94.10	1446.0
95.00	513.0
96.00	2057.0
97.20	1268.0
98.00	26104.0
99.00	16180.0
99.90	1295.0
101.00	10902.0
101.90	465.0
102.80	2562.0
104.10	4719.0
104.90	2883.0
105.20	3280.0
106.00	1174.0
107.00	69640.0
108.00	10187.0
108.80	1400.0
110.00	138944.0
111.00	17632.0
112.00	1504.0
112.70	1317.0
115.00	325.0
115.90	3269.0
117.00	61816.0
118.00	5487.0
119.80	994.0
121.10	532.0
121.90	4304.0
123.00	6642.0
123.90	3624.0
125.10	1745.0
127.00	166720.0
128.00	14351.0
129.00	74272.0
130.00	7756.0
130.80	1329.0
131.70	534.0
132.20	889.0
133.20	474.0
134.00	4177.0
135.00	8186.0
135.90	3649.0
137.10	3261.0
137.80	954.0
139.90	605.0
141.00	13769.0

Instrument :
BNA_N
ClientSampleId :
DFTPP042

141.90	4600.0
142.90	2336.0
144.00	441.0
144.80	608.0
145.90	1131.0
147.00	5422.0
148.00	13074.0
148.90	3324.0
149.90	1112.0
151.60	1010.0
152.90	3372.0
154.10	2688.0
155.00	5732.0
156.00	7693.0
157.00	2357.0
157.90	2311.0
158.90	2844.0
160.00	3146.0
160.90	4558.0
162.10	1551.0
163.00	670.0
164.10	1195.0
164.90	4415.0
166.20	3328.0
167.00	25920.0
168.00	11554.0
168.90	2142.0
169.90	803.0
171.00	1120.0
171.90	2430.0
172.90	2436.0
174.00	8938.0
175.10	7739.0
176.10	3557.0
177.00	5415.0
177.90	938.0
179.00	14906.0
179.90	12405.0
181.00	5504.0
182.00	966.0
182.60	712.0
183.80	1343.0
185.10	10015.0
186.00	78552.0
187.00	23264.0
188.00	3662.0
188.90	5525.0
189.80	1040.0
190.90	2912.0
191.90	4454.0
193.00	6749.0
193.80	1453.0
194.90	2009.0
195.90	15849.0
197.90	667328.0
198.90	52512.0
200.00	3059.0
200.80	1684.0
201.30	1555.0
202.30	1796.0
203.00	4293.0
204.00	18536.0
205.00	35736.0
206.00	139264.0
207.00	24008.0
208.00	6700.0
209.10	2421.0
210.20	2126.0
210.90	5799.0
211.80	798.0
212.80	397.0
214.90	2905.0
216.00	2776.0
216.90	32488.0
217.90	6242.0
219.00	619.0
219.90	1395.0
221.00	16464.0
221.80	4313.0
223.00	9188.0
224.00	78344.0
225.00	23360.0
226.20	2728.0
227.00	31464.0
228.00	4927.0
229.00	5584.0
229.80	1657.0
231.10	2787.0
232.00	589.0
233.00	1456.0
233.80	1575.0
234.90	3184.0
236.00	3221.0
237.10	4008.0
239.10	669.0
240.10	843.0

Instrument :
BNA_N
ClientSampleId :
DFTPP042

240.90	2021.0
241.90	5190.0
243.20	5486.0
244.00	62920.0
245.00	12891.0
246.00	12825.0
246.90	2130.0
248.00	746.0
249.00	1994.0
249.90	353.0
250.90	674.0
252.20	775.0
252.90	1098.0
253.40	2314.0
255.00	359168.0
256.00	56720.0
257.10	3093.0
258.00	23000.0
259.10	4194.0
259.80	562.0
261.10	1203.0
265.00	7463.0
265.80	1176.0
267.00	376.0
269.90	1077.0
271.30	699.0
272.90	11595.0
274.00	40096.0
275.00	201536.0
276.00	26728.0
277.00	16744.0
277.90	1914.0
282.90	1698.0
283.90	1346.0
284.30	671.0
285.10	4372.0
288.90	873.0
289.90	670.0
291.00	378.0
292.10	1325.0
293.00	3114.0
294.10	1252.0
294.90	945.0
296.00	47872.0
296.90	4988.0
297.90	807.0
301.00	792.0
302.20	1024.0
302.90	5671.0
304.00	2721.0
308.90	655.0
310.00	1606.0
312.90	735.0
314.00	3807.0
314.90	4435.0
315.90	2249.0
317.10	819.0
320.90	2777.0
322.00	1165.0
323.00	23216.0
324.00	3590.0
326.90	4367.0
328.10	1902.0
329.10	510.0
332.10	1602.0
333.10	1649.0
334.00	10504.0
334.90	3664.0
335.90	316.0
339.20	384.0
341.00	1995.0
345.90	2985.0
347.00	1005.0
351.10	612.0
352.00	5963.0
353.00	4522.0
354.00	4911.0
354.80	728.0
359.20	386.0
365.00	18112.0
366.00	2541.0
370.90	1962.0
372.00	10065.0
373.10	2101.0
383.00	2472.0
384.00	1331.0
390.00	1008.0
390.80	604.0
395.10	343.0
402.10	2768.0
403.10	7428.0
404.10	3408.0
405.20	335.0
414.90	400.0
420.90	4139.0
422.00	3735.0

Instrument :
BNA_N
ClientSampleId :
DFTPP042

423.00	29248.0
424.00	7887.0
424.90	1317.0
426.00	624.0
437.90	387.0
438.40	463.0
439.20	479.0
441.10	83864.0
442.00	512704.0
443.00	90488.0
444.10	7731.0

Instrument :
BNA_N
ClientSampleId :
DFTPP042