

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052623\
 Data File : BN025609.D
 Acq On : 26 May 2023 15:18
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: May 27 03:15:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 27 03:12:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.027	152	6132	0.400	ng	0.00
7) Naphthalene-d8	10.846	136	16562	0.400	ng	0.00
13) Acenaphthene-d10	14.662	164	9655	0.400	ng	0.01
19) Phenanthrene-d10	17.393	188	20055	0.400	ng	0.00
29) Chrysene-d12	21.578	240	14403	0.400	ng	0.00
35) Perylene-d12	24.075	264	11948	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.557	112	44849	2.855	ng	0.00
5) Phenol-d6	7.167	99	57816	3.069	ng	0.00
8) Nitrobenzene-d5	9.180	82	45595	3.288	ng	-0.01
11) 2-Methylnaphthalene-d10	12.423	152	76177	3.224	ng	0.00
14) 2,4,6-Tribromophenol	16.152	330	9314	4.772	ng	0.00
15) 2-Fluorobiphenyl	13.283	172	124383	3.121	ng	0.00
27) Fluoranthene-d10	19.426	212	146813	3.231	ng	0.00
31) Terphenyl-d14	20.015	244	96008	3.044	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.397	88	21063	2.966	ng	97
3) n-Nitrosodimethylamine	3.722	42	26094	3.412	ng	# 84
6) bis(2-Chloroethyl)ether	7.434	93	55575	2.751	ng	89
9) Naphthalene	10.889	128	143158	3.160	ng	98
10) Hexachlorobutadiene	11.177	225	24056	3.062	ng	# 96
12) 2-Methylnaphthalene	12.495	142	101125	3.245	ng	100
16) Acenaphthylene	14.373	152	153402	3.365	ng	99
17) Acenaphthene	14.715	154	96302	3.208	ng	98
18) Fluorene	15.693	166	127753	3.218	ng	99
20) 4,6-Dinitro-2-methylph...	15.767	198	13360	5.022	ng	# 45
21) 4-Bromophenyl-phenylether	16.586	248	37046	3.254	ng	93
22) Hexachlorobenzene	16.711	284	32591	3.132	ng	99
23) Atrazine	16.847	200	29817	3.282	ng	95
25) Phenanthrene	17.443	178	198320	3.249	ng	99
26) Anthracene	17.530	178	187582	3.398	ng	100
28) Fluoranthene	19.453	202	218859	3.288	ng	99
30) Pyrene	19.816	202	221807	3.164	ng	99
32) Benzo(a)anthracene	21.560	228	177685	3.312	ng	98
33) Chrysene	21.614	228	184121	3.163	ng	98
34) Bis(2-ethylhexyl)phtha...	21.480	149	95930	2.936	ng	# 97
36) Indeno(1,2,3-cd)pyrene	26.680	276	149365	3.259	ng	# 61
37) Benzo(b)fluoranthene	23.306	252	155591	3.416	ng	# 86
38) Benzo(k)fluoranthene	23.358	252	160317	3.452	ng	# 85
39) Benzo(a)pyrene	23.961	252	145831	3.355	ng	# 79
40) Dibenzo(a,h)anthracene	26.703	278	119576	3.289	ng	# 86
41) Benzo(g,h,i)perylene	27.475	276	129071	3.117	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052623\
Data File : BN025609.D
Acq On : 26 May 2023 15:18
Operator : CG/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Quant Time: May 27 03:15:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 27 03:12:34 2023
Response via : Initial Calibration







#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 8.027 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:152 Resp: 6132
Ion Ratio Lower Upper
152 100
150 151.7 121.0 181.6
115 58.9 48.5 72.7



#2
1,4-Dioxane
Concen: 2.966 ng
RT: 3.397 min Scan# 30
Delta R.T. 0.000 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18

Tgt Ion: 88 Resp: 21063
Ion Ratio Lower Upper
88 100
43 30.1 26.5 39.7
58 69.3 56.8 85.2









#3
n-Nitrosodimethylamine
Concen: 3.412 ng
RT: 3.722 min Scan# 71
Delta R.T. -0.014 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 26094
Ion Ratio Lower Upper
42 100
74 118.8 82.2 123.2
44 5.7 0.0 0.0#



#4
2-Fluorophenol
Concen: 2.855 ng
RT: 5.557 min Scan# 329
Delta R.T. -0.007 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18

Tgt Ion: 112 Resp: 44849
Ion Ratio Lower Upper
112 100
64 57.2 45.7 68.5
63 31.9 26.2 39.2









#5
Phenol-d6
Concen: 3.069 ng
RT: 7.167 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 57816
Ion Ratio Lower Upper
99 100
42 21.4 17.1 25.7
71 36.7 29.3 43.9



#6
bis(2-Chloroethyl)ether
Concen: 2.751 ng
RT: 7.434 min Scan# 589
Delta R.T. -0.008 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18

Tgt Ion: 93 Resp: 55575
Ion Ratio Lower Upper
93 100
63 71.9 49.1 73.7
95 31.6 22.6 33.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.846 min Scan# 948
Delta R.T. 0.000 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	16562
Ion Ratio	Lower	Upper	
136	100		
137	11.3	9.1	13.7
54	6.5	5.4	8.0
68	4.8	4.3	6.5



#8
Nitrobenzene-d5
Concen: 3.288 ng
RT: 9.180 min Scan# 792
Delta R.T. -0.011 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18

Tgt Ion:	82	Resp:	45595
Ion Ratio	Lower	Upper	
82	100		
128	33.3	32.5	48.7
54	54.4	40.6	60.8





#9

Naphthalene

Concen: 3.160 ng

RT: 10.889 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN025609.D

Acq: 26 May 2023 15:18

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:128 Resp: 143158

Ion Ratio Lower Upper

128 100

129 11.0 9.3 13.9

127 13.0 11.0 16.6



#10

Hexachlorobutadiene

Concen: 3.062 ng

RT: 11.177 min Scan# 979

Delta R.T. 0.000 min

Lab File: BN025609.D

Acq: 26 May 2023 15:18




Tgt Ion:225 Resp: 24056

Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.9 48.4 72.6





#11

2-Methylnaphthalene-d10

Concen: 3.224 ng

RT: 12.423 min Scan# 11

Delta R.T. -0.004 min

Lab File: BN025609.D

Acq: 26 May 2023 15:18

Instrument :

BNA_N

ClientSampleId :





SSTDICC3.2

Tgt Ion:152 Resp: 76177

Ion Ratio Lower Upper

152 100

151 20.5 16.1 24.1



#12

2-Methylnaphthalene

Concen: 3.245 ng

RT: 12.495 min Scan# 1271

Delta R.T. -0.004 min

Lab File: BN025609.D

Acq: 26 May 2023 15:18




Tgt Ion:142 Resp: 101125


Ion Ratio Lower Upper

142 100

141 88.7 70.7 106.1

115 38.9 31.5 47.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.662 min Scan# 1561
Delta R.T. 0.011 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 9655
Ion Ratio Lower Upper
164 100
162 102.6 86.4 129.6
160 48.7 41.9 62.9



#14
2,4,6-Tribromophenol
Concen: 4.772 ng
RT: 16.152 min Scan# 1692
Delta R.T. -0.000 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18

Tgt Ion:330 Resp: 9314
Ion Ratio Lower Upper
330 100
332 95.8 80.0 120.0
141 57.8 45.8 68.6





#15

2-Fluorobiphenyl

Concen: 3.121 ng

RT: 13.283 min Scan# 1432

Delta R.T. 0.000 min

Lab File: BN025609.D

Acq: 26 May 2023 15:18

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:172 Resp: 124383

Ion Ratio Lower Upper

172 100

171 35.0 28.6 42.8

170 24.1 19.7 29.5



#16

Acenaphthylene

Concen: 3.365 ng

RT: 14.373 min Scan# 1534

Delta R.T. -0.011 min

Lab File: BN025609.D

Acq: 26 May 2023 15:18




Tgt Ion:152 Resp: 153402


Ion Ratio Lower Upper

152 100

151 19.4 15.8 23.6

153 12.9 10.6 15.8









#17
Acenaphthene
Concen: 3.208 ng
RT: 14.715 min Scan# 1566
Delta R.T. 0.000 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 96302
Ion Ratio Lower Upper
154 100
153 113.0 91.9 137.9
152 55.5 45.4 68.2



#18
Fluorene
Concen: 3.218 ng
RT: 15.693 min Scan# 1655
Delta R.T. -0.000 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18

Tgt Ion:166 Resp: 127753
Ion Ratio Lower Upper
166 100
165 98.5 79.3 118.9
167 13.4 10.7 16.1





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.393 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN025609.D

Acq: 26 May 2023 15:18

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:188 Resp: 20055

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 14.3 12.2 18.2



#20

4,6-Dinitro-2-methylphenol

Concen: 5.022 ng

RT: 15.767 min Scan# 1661

Delta R.T. -0.025 min

Lab File: BN025609.D

Acq: 26 May 2023 15:18




Tgt Ion:198 Resp: 13360


Ion Ratio Lower Upper

198 100

51 54.2 110.8 166.2#

105 52.4 61.9 92.9#









#21
4-Bromophenyl-phenylether
Concen: 3.254 ng
RT: 16.586 min Scan# 1727
Delta R.T. 0.000 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 37046
Ion Ratio Lower Upper
248 100
250 96.5 77.5 116.3
141 66.2 64.0 96.0



#22
Hexachlorobenzene
Concen: 3.132 ng
RT: 16.711 min Scan# 1737
Delta R.T. 0.000 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18

Tgt Ion:284 Resp: 32591
Ion Ratio Lower Upper
284 100
142 51.3 42.0 63.0
249 35.3 28.6 43.0









#23
Atrazine
Concen: 3.282 ng
RT: 16.847 min Scan# 1748
Delta R.T. 0.000 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 29817
Ion Ratio Lower Upper
200 100
173 29.2 27.9 41.9
215 43.3 35.1 52.7



#25
Phenanthrene
Concen: 3.249 ng
RT: 17.443 min Scan# 1796
Delta R.T. 0.000 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18

Tgt Ion:178 Resp: 198320
Ion Ratio Lower Upper
178 100
176 19.3 15.4 23.2
179 15.1 12.6 19.0









#26
 Anthracene
 Concen: 3.398 ng
 RT: 17.530 min Scan# 1803
 Delta R.T. 0.000 min
 Lab File: BN025609.D
 Acq: 26 May 2023 15:18




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:178 Resp: 187582
 Ion Ratio Lower Upper
 178 100
 176 18.7 14.8 22.2
 179 15.2 12.2 18.2



#27
 Fluoranthene-d10
 Concen: 3.231 ng
 RT: 19.426 min Scan# 2079
 Delta R.T. 0.000 min
 Lab File: BN025609.D
 Acq: 26 May 2023 15:18

Tgt Ion:212 Resp: 146813
 Ion Ratio Lower Upper
 212 100
 106 15.6 12.3 18.5
 104 9.0 7.2 10.8









#28
Fluoranthene
Concen: 3.288 ng
RT: 19.453 min Scan# 2085
Delta R.T. 0.000 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:202 Resp: 218859
Ion Ratio Lower Upper
202 100
101 11.5 9.2 13.8
203 17.1 13.3 19.9



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.578 min Scan# 2483
Delta R.T. 0.000 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18

Tgt Ion:240 Resp: 14403
Ion Ratio Lower Upper
240 100
120 16.4 15.4 23.0
236 29.0 23.2 34.8





#30

Pyrene

Concen: 3.164 ng

RT: 19.816 min Scan# 2163

Delta R.T. 0.000 min

Lab File: BN025609.D

Acq: 26 May 2023 15:18

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:202 Resp: 221807

Ion Ratio Lower Upper

202 100

200 21.1 17.1 25.7

203 17.7 14.5 21.7



#31

Terphenyl-d14

Concen: 3.044 ng

RT: 20.015 min Scan# 2206

Delta R.T. 0.000 min

Lab File: BN025609.D

Acq: 26 May 2023 15:18




Tgt Ion:244 Resp: 96008


Ion Ratio Lower Upper

244 100

212 9.7 8.5 12.7

122 16.6 14.9 22.3





#32

Benzo(a)anthracene

Concen: 3.312 ng

RT: 21.560 min Scan# 2481

Delta R.T. 0.000 min

Lab File: BN025609.D

Acq: 26 May 2023 15:18

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:228 Resp: 177685

Ion Ratio Lower Upper

228 100

226 27.7 22.9 34.3

229 19.2 15.9 23.9



#33

Chrysene

Concen: 3.163 ng

RT: 21.614 min Scan# 2487

Delta R.T. -0.009 min

Lab File: BN025609.D

Acq: 26 May 2023 15:18




Tgt Ion:228 Resp: 184121


Ion Ratio Lower Upper

228 100

226 31.4 24.5 36.7

229 19.2 16.2 24.4









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 2.936 ng
 RT: 21.480 min Scan# 2472
 Delta R.T. 0.000 min
 Lab File: BN025609.D
 Acq: 26 May 2023 15:18




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2


Tgt Ion:149 Resp: 95930
 Ion Ratio Lower Upper
 149 100
 167 26.8 20.4 30.6
 279 2.8 3.3 4.9#



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 24.075 min Scan# 3145
 Delta R.T. 0.000 min
 Lab File: BN025609.D
 Acq: 26 May 2023 15:18

Tgt Ion:264 Resp: 11948
 Ion Ratio Lower Upper
 264 100
 260 26.5 22.3 33.5
 265 57.2 54.3 81.5









#36
Indeno(1,2,3-cd)pyrene
Concen: 3.259 ng
RT: 26.680 min Scan# 4036
Delta R.T. 0.000 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:276 Resp: 149365
Ion Ratio Lower Upper
276 100
138 30.5 10.5 15.7#
277 24.7 9.2 13.8#



#37
Benzo(b)fluoranthene
Concen: 3.416 ng
RT: 23.306 min Scan# 2882
Delta R.T. -0.006 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18

Tgt Ion:252 Resp: 155591
Ion Ratio Lower Upper
252 100
253 23.2 26.2 39.4#
125 12.3 12.9 19.3#









#38
Benzo(k)fluoranthene
Concen: 3.452 ng
RT: 23.358 min Scan# 2900
Delta R.T. -0.003 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 160317
Ion Ratio Lower Upper
252 100
253 22.6 26.1 39.1#
125 12.1 12.5 18.7#



#39
Benzo(a)pyrene
Concen: 3.355 ng
RT: 23.961 min Scan# 3106
Delta R.T. -0.003 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18

Tgt Ion:252 Resp: 145831
Ion Ratio Lower Upper
252 100
253 22.7 30.2 45.2#
125 13.2 14.8 22.2#









#40
Dibenzo(a,h)anthracene
Concen: 3.289 ng
RT: 26.703 min Scan# 4044
Delta R.T. -0.009 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:278 Resp: 119576
Ion Ratio Lower Upper
278 100
139 20.2 18.6 28.0
279 24.1 27.8 41.8#



#41
Benzo(g,h,i)perylene
Concen: 3.117 ng
RT: 27.475 min Scan# 4308
Delta R.T. 0.003 min
Lab File: BN025609.D
Acq: 26 May 2023 15:18

Tgt Ion:276 Resp: 129071
Ion Ratio Lower Upper
276 100
277 23.8 21.4 32.0
138 26.5 22.8 34.2

