

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052623\
 Data File : BN025627.D
 Acq On : 27 May 2023 03:18
 Operator : CG/JU
 Sample : 02852-15
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 CB-32-G3-SO-69.5-051823

Quant Time: May 27 04:38:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 27 03:51:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.026	152	7309	0.400	ng	0.00
7) Naphthalene-d8	10.846	136	20427	0.400	ng	0.00
13) Acenaphthene-d10	14.651	164	10972	0.400	ng	0.00
19) Phenanthrene-d10	17.393	188	20407	0.400	ng	0.00
29) Chrysene-d12	21.578	240	9403	0.400	ng	0.00
35) Perylene-d12	24.060	264	7893	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.564	112	7384	0.394	ng	0.00
5) Phenol-d6	7.174	99	8644	0.385	ng	0.00
8) Nitrobenzene-d5	9.191	82	6605	0.386	ng	0.00
11) 2-Methylnaphthalene-d10	12.423	152	11581	0.397	ng	0.00
14) 2,4,6-Tribromophenol	16.152	330	938	0.423	ng	0.00
15) 2-Fluorobiphenyl	13.283	172	19343	0.427	ng	0.00
27) Fluoranthene-d10	19.425	212	16052	0.347	ng	0.00
31) Terphenyl-d14	20.011	244	10716	0.521	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.471	149	10839	0.508	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052623\
Data File : BN025627.D
Acq On : 27 May 2023 03:18
Operator : CG/JU
Sample : 02852-15
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
CB-32-G3-SO-69.5-051823

Quant Time: May 27 04:38:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 27 03:51:06 2023
Response via : Initial Calibration

