

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052623\
 Data File : BN025630.D
 Acq On : 27 May 2023 05:06
 Operator : CG/JU
 Sample : 02880-02
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D5-20230517

Quant Time: May 27 05:45:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat May 27 03:51:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.026	152	6252	0.400	ng	0.00
7) Naphthalene-d8	10.846	136	18076	0.400	ng	0.00
13) Acenaphthene-d10	14.651	164	10214	0.400	ng	0.00
19) Phenanthrene-d10	17.393	188	20261	0.400	ng	0.00
29) Chrysene-d12	21.578	240	11379	0.400	ng	0.00
35) Perylene-d12	24.060	264	10392	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.564	112	1807	0.113	ng	0.00
5) Phenol-d6	7.181	99	1127	0.059	ng	0.00
8) Nitrobenzene-d5	9.191	82	3572	0.236	ng	0.00
11) 2-Methylnaphthalene-d10	12.423	152	7898	0.306	ng	0.00
14) 2,4,6-Tribromophenol	16.152	330	868	0.420	ng	0.00
15) 2-Fluorobiphenyl	13.283	172	11710	0.278	ng	0.00
27) Fluoranthene-d10	19.425	212	16017	0.349	ng	0.00
31) Terphenyl-d14	20.011	244	10005	0.402	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.397	88	42919	5.928	ng	Qvalue 98
6) bis(2-Chloroethyl)ether	7.441	93	606	0.031	ng	# 85
30) Pyrene	19.816	202	1178	0.021	ng	98
34) Bis(2-ethylhexyl)phtha...	21.471	149	3418	0.132	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052623\
Data File : BN025630.D
Acq On : 27 May 2023 05:06
Operator : CG/JU
Sample : 02880-02
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D5-20230517

Quant Time: May 27 05:45:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat May 27 03:51:06 2023
Response via : Initial Calibration

