

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052623\
 Data File : BN025672.D
 Acq On : 28 May 2023 08:30
 Operator : CG/JU
 Sample : 02911-18
 Misc :
 ALS Vial : 69 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DUP04-20230522

Quant Time: May 29 01:14:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 29 00:47:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.019	152	5565	0.400	ng	0.00
7) Naphthalene-d8	10.835	136	15116	0.400	ng	# 0.00
13) Acenaphthene-d10	14.651	164	8894	0.400	ng	0.00
19) Phenanthrene-d10	17.393	188	17851	0.400	ng	0.00
29) Chrysene-d12	21.578	240	10988	0.400	ng	0.00
35) Perylene-d12	24.057	264	9722	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.564	112	1535	0.108	ng	0.00
5) Phenol-d6	7.203	99	780	0.046	ng	0.03
8) Nitrobenzene-d5	9.202	82	3241	0.256	ng	0.01
11) 2-Methylnaphthalene-d10	12.427	152	6232	0.289	ng	0.00
14) 2,4,6-Tribromophenol	16.152	330	509	0.283	ng	0.00
15) 2-Fluorobiphenyl	13.283	172	10445	0.285	ng	0.00
27) Fluoranthene-d10	19.421	212	15123	0.374	ng	0.00
31) Terphenyl-d14	20.011	244	10912	0.454	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.397	88	2954	0.458	ng	96
34) Bis(2-ethylhexyl)phtha...	21.471	149	1739	0.070	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052623\
Data File : BN025672.D
Acq On : 28 May 2023 08:30
Operator : CG/JU
Sample : 02911-18
Misc :
ALS Vial : 69 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DUP04-20230522

Quant Time: May 29 01:14:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 29 00:47:45 2023
Response via : Initial Calibration

