




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052623\
Data File : BN025698.D
Acq On : 29 May 2023 01:46
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 97 Sample Multiplier: 1

Quant Time: May 29 03:19:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 29 03:15:08 2023
Response via : Initial Calibration



TIC: BN025698.D\data.ms

(70) Pentachlorophenol (C)




17.039min (-0.006) 12331.19 ng

response 196157

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	65.75
264.00	61.60	70.85
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052623\
Data File : BN025698.D
Acq On : 29 May 2023 01:46
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 97 Sample Multiplier: 1

Quant Time: May 29 03:19:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 29 03:15:08 2023
Response via : Initial Calibration



TIC: BN025698.D\data.ms

(77) Benzidine

19.627min (-0.006) 0.00 ng

response 1408458

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.36
183.00	13.20	11.10
0.00	0.00	0.00