

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052623\
 Data File : BN025699.D
 Acq On : 29 May 2023 02:57
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 98 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 29 03:33:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 29 00:47:45 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	115	0.00
2	1,4-Dioxane	0.400	0.433	-8.2	117	0.00
3	n-Nitrosodimethylamine	0.400	0.329	17.8	98	0.00
4 S	2-Fluorophenol	0.400	0.408	-2.0	115	0.00
5 S	Phenol-d6	0.400	0.413	-3.2	120	0.00
6	bis(2-Chloroethyl)ether	0.400	0.426	-6.5	120	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	118	0.00
8 S	Nitrobenzene-d5	0.400	0.398	0.5	122	-0.01
9	Naphthalene	0.400	0.418	-4.5	122	-0.01
10	Hexachlorobutadiene	0.400	0.408	-2.0	115	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.412	-3.0	123	0.00
12	2-Methylnaphthalene	0.400	0.411	-2.7	123	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	121	0.00
14 S	2,4,6-Tribromophenol	0.400	0.483	-20.7	191	-0.01
15 S	2-Fluorobiphenyl	0.400	0.404	-1.0	121	-0.01
16	Acenaphthylene	0.400	0.397	0.8	122	0.00
17	Acenaphthene	0.400	0.412	-3.0	123	-0.01
18	Fluorene	0.400	0.411	-2.7	125	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	123	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.491	-22.7	176	-0.01
21	4-Bromophenyl-phenylether	0.400	0.409	-2.2	123	0.00
22	Hexachlorobenzene	0.400	0.413	-3.2	123	0.00
23	Atrazine	0.400	0.391	2.3	122	0.00
24	Pentachlorophenol	0.400	0.792	-98.0#	1340	0.00
25	Phenanthrene	0.400	0.415	-3.7	126	0.00
26	Anthracene	0.400	0.402	-0.5	125	0.00
27 SURR	Fluoranthene-d10	0.400	0.392	2.0	123	0.00
28	Fluoranthene	0.400	0.394	1.5	126	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	115	0.00
30	Pyrene	0.400	0.478	-19.5	125	0.00
31 S	Terphenyl-d14	0.400	0.444	-11.0	122	0.00
32	Benzo(a)anthracene	0.400	0.400	0.0	115	0.00
33	Chrysene	0.400	0.435	-8.7	124	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.433	-8.2	131	0.00
35 I	Perylene-d12	0.400	0.400	0.0	117	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.451	-12.7	127	-0.02
37	Benzo(b)fluoranthene	0.400	0.420	-5.0	125	0.00
38	Benzo(k)fluoranthene	0.400	0.404	-1.0	114	-0.01
39 C	Benzo(a)pyrene	0.400	0.426	-6.5	121	0.00
40	Dibenzo(a,h)anthracene	0.400	0.444	-11.0	128	-0.03
41	Benzo(g,h,i)perylene	0.400	0.450	-12.5	124	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4