

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052623\
 Data File : BN025710.D
 Acq On : 29 May 2023 09:30
 Operator : CG/JU
 Sample : 02911-01
 Misc :
 ALS Vial : 109 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW10-MW01D-20230522

Quant Time: May 30 00:53:02 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 29 00:47:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.019	152	5684	0.400	ng	0.00
7) Naphthalene-d8	10.835	136	16724	0.400	ng	# 0.00
13) Acenaphthene-d10	14.651	164	9943	0.400	ng	0.00
19) Phenanthrene-d10	17.393	188	21268	0.400	ng	# 0.00
29) Chrysene-d12	21.569	240	14853	0.400	ng	0.00
35) Perylene-d12	24.054	264	13422	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.564	112	1568	0.108	ng	0.00
5) Phenol-d6	7.182	99	1127	0.065	ng	0.00
8) Nitrobenzene-d5	9.191	82	3349	0.239	ng	0.00
11) 2-Methylnaphthalene-d10	12.419	152	6813	0.286	ng	0.00
14) 2,4,6-Tribromophenol	16.140	330	752	0.374	ng	-0.01
15) 2-Fluorobiphenyl	13.283	172	11381	0.277	ng	0.00
27) Fluoranthene-d10	19.421	212	16074	0.334	ng	0.00
31) Terphenyl-d14	20.006	244	12806	0.394	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.471	149	5113	0.152	ng	# 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN052623\
Data File : BN025710.D
Acq On : 29 May 2023 09:30
Operator : CG/JU
Sample : 02911-01
Misc :
ALS Vial : 109 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW10-MW01D-20230522


Quant Time: May 30 00:53:02 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN052623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 29 00:47:45 2023
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 8.019 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN025710.D
Acq: 29 May 2023 09:30

Instrument :
BNA_N
ClientSampleId :
RW10-MW01D-20230522






Tgt Ion:152 Resp: 5684
Ion Ratio Lower Upper
152 100
150 151.0 121.0 181.6
115 59.8 48.5 72.7



#4
2-Fluorophenol
Concen: 0.108 ng
RT: 5.564 min Scan# 330
Delta R.T. 0.007 min
Lab File: BN025710.D
Acq: 29 May 2023 09:30

Tgt Ion:112 Resp: 1568
Ion Ratio Lower Upper
112 100
64 61.7 45.7 68.5
63 33.2 26.2 39.2








#5
Phenol-d6
Concen: 0.065 ng
RT: 7.182 min Scan# 511
Delta R.T. 0.007 min
Lab File: BN025710.D
Acq: 29 May 2023 09:30




Instrument :
BNA_N
ClientSampleId :
RW10-MW01D-20230522

Tgt Ion: 99 Resp: 1127
Ion Ratio Lower Upper
99 100
42 24.8 17.1 25.7
71 45.3 29.3 43.9#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.835 min Scan# 947
Delta R.T. 0.000 min
Lab File: BN025710.D
Acq: 29 May 2023 09:30

Tgt Ion: 136 Resp: 16724
Ion Ratio Lower Upper
136 100
137 11.4 9.1 13.7
54 8.3 5.4 8.0#
68 6.4 4.3 6.5









#8
 Nitrobenzene-d5
 Concen: 0.239 ng
 RT: 9.191 min Scan# 793
 Delta R.T. 0.000 min
 Lab File: BN025710.D
 Acq: 29 May 2023 09:30




Instrument :
 BNA_N
 ClientSampleId :
 RW10-MW01D-20230522


Tgt Ion:	82	Resp:	3349
Ion	Ratio	Lower	Upper
82	100		
128	38.3	32.5	48.7
54	55.9	40.6	60.8



#11
 2-Methylnaphthalene-d10
 Concen: 0.286 ng
 RT: 12.419 min Scan# 1251
 Delta R.T. -0.004 min
 Lab File: BN025710.D
 Acq: 29 May 2023 09:30

Tgt Ion:	152	Resp:	6813
Ion	Ratio	Lower	Upper
152	100		
151	20.7	16.1	24.1









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.651 min Scan# 1560
Delta R.T. 0.000 min
Lab File: BN025710.D
Acq: 29 May 2023 09:30




Instrument :
BNA_N
ClientSampleId :
RW10-MW01D-20230522


Tgt Ion:164 Resp: 9943
Ion Ratio Lower Upper
164 100
162 104.5 86.4 129.6
160 48.3 41.9 62.9



#14
2,4,6-Tribromophenol
Concen: 0.374 ng
RT: 16.140 min Scan# 1691
Delta R.T. -0.012 min
Lab File: BN025710.D
Acq: 29 May 2023 09:30

Tgt Ion:330 Resp: 752
Ion Ratio Lower Upper
330 100
332 96.5 80.0 120.0
141 58.6 45.8 68.6









#15
2-Fluorobiphenyl
Concen: 0.277 ng
RT: 13.283 min Scan# 1432
Delta R.T. 0.000 min
Lab File: BN025710.D
Acq: 29 May 2023 09:30




Instrument :
BNA_N
ClientSampleId :
RW10-MW01D-20230522


Tgt Ion:172	Resp:	11381
Ion	Ratio	Lower Upper
172	100	
171	34.1	28.6 42.8
170	23.3	19.7 29.5



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.393 min Scan# 1792
Delta R.T. 0.000 min
Lab File: BN025710.D
Acq: 29 May 2023 09:30

Tgt Ion:188	Resp:	21268
Ion	Ratio	Lower Upper
188	100	
94	0.0	0.0 0.0
80	10.4	12.2 18.2#





#27

Fluoranthene-d10

Concen: 0.334 ng

RT: 19.421 min Scan# 2079

Delta R.T. 0.000 min

Lab File: BN025710.D

Acq: 29 May 2023 09:30

Instrument :

BNA_N

ClientSampleId :

RW10-MW01D-20230522





Tgt Ion:212 Resp: 16074

Ion Ratio Lower Upper

212 100

106 16.2 12.3 18.5

104 10.2 7.2 10.8



#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.569 min Scan# 2482

Delta R.T. -0.009 min

Lab File: BN025710.D

Acq: 29 May 2023 09:30




Tgt Ion:240 Resp: 14853


Ion Ratio Lower Upper

240 100

120 19.3 15.4 23.0

236 30.0 23.2 34.8









#31
Terphenyl-d14
Concen: 0.394 ng
RT: 20.006 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN025710.D
Acq: 29 May 2023 09:30




Instrument :
BNA_N
ClientSampleId :
RW10-MW01D-20230522


Tgt Ion:244 Resp: 12806
Ion Ratio Lower Upper
244 100
212 10.3 8.5 12.7
122 18.8 14.9 22.3



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.152 ng
RT: 21.471 min Scan# 2471
Delta R.T. 0.000 min
Lab File: BN025710.D
Acq: 29 May 2023 09:30

Tgt Ion:149 Resp: 5113
Ion Ratio Lower Upper
149 100
167 25.2 20.4 30.6
279 3.1 3.3 4.9#





#35

Perylene-d12

Concen: 0.400 ng

RT: 24.054 min Scan# 31

Delta R.T. -0.015 min

Lab File: BN025710.D


Acq: 29 May 2023 09:30

Instrument :

BNA_N

ClientSampleId :

RW10-MW01D-20230522



Tgt Ion:264 Resp: 13422

Ion Ratio Lower Upper

264 100

260 26.9 22.3 33.5

265 52.2 54.3 81.5#

